

## CERTIFICADO PATRON DE CND


**TEND** 

NDT Reference Standard Certificate



2C112





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth			Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	[mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Longitudinal Externo longitudinal OD	1,00	5	±15	1,236	0,914	1,075	38,10	38,10	1,070	1,066	1,00	Sebastian Monteros	05/12/06
В	Transversal Externo Transverse OD	1,00	5	±15	1,236	0,914	1,075	38,10	38,10	1,085	1,090	1,00	Sebastian Monteros	05/12/06

## Instrumentos de Medicion Utilizados:

## Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

## Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO