

CERTIFICADO PATRON DE CND

TEND

TAND-001-05


17/04/2006

NDT Reference Standard Certificate



1c591

Datos del Producto Product Information							
Diametro [mm]	33,4	Cliente Client	Ciclo Product	22565			
Espesor [mm] Wall Thickness	3,38	Tratamiento Termico Heat Treatment	Colada _{Heat}	0			
Trazabilidad Traceability	0		Expediente Manufacturing Order				



Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam			Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A		Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Agujero Pasante Drilled hole	1,60	-	-	-	-	-	-	-	-	-	1,60	Sebastian Monteros	05/11/04
В	Longitudinal Externo longitudinal OD	1,00	7	±15	0,272	0,201	0,237	50,00	50,00	7,000	7,000	1,00	Sebastian Monteros	05/11/04

Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO