

CERTIFICADO PATRON DE CND


TEND

NDT Reference Standard Certificate



1C258





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth	Tolerancia [%] Tolerance	Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	[mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Ι Δ	Longitudinal Externo longitudinal OD	1,00	13	±15	2,738	2,024	2,381	38,00	38,00	2,380	2,382	0,99	Sebastian Monteros	13/11/06
В	Transversal Externo Transverse OD	1,00	5	±15	1,095	0,810	0,953	38,00	38,00	0,950	0,954	0,99	Sebastian Monteros	13/11/06
1 (-	Longitudinal Externo longitudinal OD	1,00	5	±15	1,095	0,810	0,953	38,00	38,00	0,949	0,957	0,99	Sebastian Monteros	13/11/06

Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

Notes:

1.-All dimensions in milimeters unless stated.

1/2

2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO