

CERTIFICADO PATRON DE CND


TEND

NDT Reference Standard Certificate

Tubo patrón número Reference standard #

1C417





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth	Tolerancia [%] Tolerance	Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Longitudinal Interno Longitudinal ID	1,00	5	±15	0,546	0,404	0,475	50,00	0,94	0,510	0,510	46,00	Sebastian Monteros	06/11/06
В	Longitudinal Interno Longitudinal ID	1,00	7	±15	0,765	0,565	0,665	50,00	46,20	0,560	0,580	0,95	Sebastian Monteros	06/11/06
С	Longitudinal Externo longitudinal OD	1,00	5	±15	0,546	0,404	0,475	50,00	46,00	0,410	0,450	0,93	Sebastian Monteros	06/11/06
D	Longitudinal Interno Longitudinal ID	1,00	7	±15	0,765	0,565	0,665	50,00	46,00	0,410	0,450	0,93	Sebastian Monteros	06/11/06
Е	Longitudinal Interno Longitudinal ID	1,00	5	±15	0,546	0,404	0,475	50,00	47,00	0,650	0,630	0,94	Sebastian Monteros	06/11/06

Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA SI

Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA YES