

CERTIFICADO PATRON DE CND


TEND

NDT Reference Standard Certificate



1C766





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

| _ | Posicion Position | Notch Type | Ancho / Diam Max [mm] Max Width / Diam | Profundidad Nominal [%] Depth | | | | Profundidad Nominal [mm] Nom Depth | Longitud Maxima [mm] Max Length | Longitud [mm] Length | Profundida d A [mm] Depth A | Profundida d B [mm] Depth B | Ancho / Diam [mm] Width / Diam | Responsable Responsible | Fecha Date |
|---|----------------------|--------------------------------------|----------------------------------------------|-------------------------------------|-----|-------|-------|------------------------------------------|------------------------------------------|----------------------------|-----------------------------------|-----------------------------------|--------------------------------------|----------------------------|---------------|
| | Δ | Longitudinal Externo longitudinal OD | 1,00 | 6 | ±15 | 1,150 | 0,850 | 1,000 | 30,00 | 28,00 | 1,016 | 1,034 | 0,98 | Sebastian Monteros | 03/12/04 |

Instrumentos de Medicion Utilizados:

Proyector de Perfiles Nikon V-16E Codigo: 90012410

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA SI

Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA YES

1/2