

## CERTIFICADO PATRON DE CND


**TEND** 

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

1C1984





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

<u>'</u>														
Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth	Tolerancia [%] Tolerance	Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Longitudinal Interno Longitudinal ID	1,00	10	±15	0,651	0,481	0,566	25,40	25,00	0,649	0,641	0,84	Default Default	21/03/11
В	Longitudinal Externo longitudinal OD	1,00	5	±15	0,345	0,255	0,300	25,40	25,00	0,340	0,341	0,85	Default Default	21/03/11
С	Transversal Externo Transverse OD	1,00	5	±15	0,345	0,255	0,300	25,40	23,00	0,339	0,342	0,84	Default Default	21/03/11
D	Transversal Interno Transverse ID	1,00	10	±15	0,651	0,481	0,566	25,40	24,50	0,650	0,650	0,88	Default Default	21/03/11

## Instrumentos de Medicion Utilizados:

## Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA SI

## Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA YES

1/2