

CERTIFICADO PATRON DE CND


TEND

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

2c173





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth	Tolerancia [%] Tolerance	Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Longitudinal Interno Longitudinal ID	1,00	5	±15	0,396	0,292	0,344	50,00	47,00	0,340	0,300	0,81	Sebastian Monteros	01/09/04
В	Longitudinal Externo longitudinal OD	1,00	5	±15	0,396	0,292	0,344	50,00	48,20	0,390	0,370	0,81	Sebastian Monteros	01/09/04
С	Transversal Externo Transverse OD	1,00	5	±15	0,396	0,292	0,344	50,00	50,00	0,360	0,360	0,60	Sebastian Monteros	01/09/04
D	Transversal Interno Transverse ID	1,00	5	±15	0,396	0,292	0,344	50,00	48,00	0,340	0,340	0,72	Sebastian Monteros	01/09/04

Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO