


## CERTIFICADO PATRON DE CND

**TEND** 

NDT Reference Standard Certificate



Positio		Max [mm] Max Width / Diam	Nominal [%] Depth	[%] Tolerance		Minima [mm] Min Depth	Nominal [mm] Nom Depth	Maxima [mm] Max Length	[mm] Length	d A [mm] Depth A	d B [mm] Depth B	[mm] Width / Diam	Responsible	Date
Α	Longitudinal Externo longitudinal OD	0,30	5	±15	0,572	0,423	0,498	10,00	10,00	0,525	0,530	0,30	Sebastian Monteros	23/07/20
В	Longitudinal Externo longitudinal OD	0,30	5	±15	0,572	0,423	0,498	10,00	10,00	0,530	0,525	0,30	Sebastian Monteros	23/07/20

Instrumentos de Medicion Utilizados:

Comparador a cuchilla Codigo:

## Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

## Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO