

CERTIFICADO PATRON DE CND


TEND

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

1C1987





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth	Tolerancia [%] Tolerance	Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Longitudinal Interno Longitudinal ID	0,85	5	±15	0,638	0,472	0,555	25,00	40,00	0,781	0,790	0,81	Default Default	10/05/10
В	Longitudinal Externo longitudinal OD	1,00	5	±15	0,638	0,472	0,555	50,80	46,70	0,584	0,582	0,82	Default Default	10/05/10
С	Transversal Externo Transverse OD	1,00	5	±15	0,638	0,472	0,555	50,80	41,60	0,574	0,578	0,84	Default Default	10/05/10
D	Transversal Interno Transverse ID	0,85	5	±15	0,638	0,472	0,555	23,00	36,30	1,080	1,086	0,82	Default Default	10/05/10
Е	Agujero Pasante Drilled hole	2,20	-	-	-	-	-	-	-	-	-	2,20	Default Default	01/07/10

Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA SI

Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA YES

1/2