



REV NO.	REVISION NOTE	DRN. BY	CHD. BY	DATE


SR. NO.	PART NAME	MATERIAL	QTY
01	LEFT SIDE END PLATE	G1 0.6 mm THK	01
02	RIGHT SIDE END PLATE	G1 0.6 mm THK	01
03	FINS	MS WIRE Ø1.2 mm x 236mm	2 x 45±1
04	TUBE	MS Ø4.76 x 0.7 THK	-
05	RUBBER PLUG	-	02




TOP VIEW





LEFT SIDE VIEW




FRONT VIEW



RIGHT SIDE VIEW





PATH



BOTTOM VIEW

NOTE :
GENERAL TOLERANCE : ±0.2".
ALL DIMENSIONS ARE IN 'INCH' (mm).
*** COIL TO BE WATER LEAK TESTED AT 550 PSI.**

CONDENSER COIL : 9"W x 9.5"H x 4ROWS.
MS TUBE : Ø4.76mm x 0.7 THK
FEEDING : SINGLE FEEDING.
PITCH : 0.8" (20.2mm)

APPROVED BY	TP RAJENDRA	DATE	18-JUN-2024		TITLE	DETAIL VIEWS OF CONDENSER			QTY : ----
DRAWN BY	NIMESH	DATE	18-JUN-2024		EQUIPMENT	SMALL CHAMBER (GNV)			
	NEWTRONIC LIFECARE EQUIPMENT PVT. LTD.				SCALE	SHEET. NO.	DWG. NO.	REV. NO.	
KANAKIA WALL STREET, A-WING-612, 6TH FLOOR, CHAKALA, ANDHERI-KURLA ROAD, NEAR CHAKALA METRO STATION, ANDHERI (E), MUMBAI-400093, INDIA.					NTS.	01 OF 01	RFCON - 994	02	