

Signed revision notice RN-8

The approved lot-specific blank for plate O-441 is 0.070 absorbance units.

Do not use the draft universal blank of 0.050 for this plate.

RN-8 supersedes worksheet header note OM-6-draft.

Well validity correction

Control-Q well B07 is invalid because of a confirmed pipette skip.

Control-Q well B08 remains valid and may be evaluated alone.

No Orchid-K well is invalidated by this notice.