

Dispenser Functionality Tester Data Analysis

Report Name: Ops samps_20181002_084133**UserID:** venturf2**User Email:** Franklin.Ventura@roche.com**Report Created:** 2019-02-01 08:42:09.063747**Analysis Version:** 0.1.0**Total Dispensers:** 3**Failing Dispensers:** 2

Dispenser Details

Label	Stress	Type	Reagent	Average (uL)	Std Dev (uL)	Pass/Fail	DFTM Complete	DFTM Incomplete	Failing Reason	Drop First Fail
S2L3S3	Purple	1695501 E90092	water	91.9	1.01	Fail	6	9	12 drop(s) greater than 100	301
S2SL7S4	25-OCT-2018TG6	1695501 E90092	Avidin Diluent with B5 Blocker for Multimer	93.6	8.33	Fail	3	3	1 drop(s) greater than 175	801
19-OCT-2018CTR3	CONTROL	W/O OIL ON BARREL TIP	Avidin Diluent with B5 Blocker for Multimer	91.9	0.97	Pass	2	3	N/A	N/A

Audits and Exclusions

Label	Reason	User	Date	Description
19-OCT-2018CTR3	Insufficient Prime	venturf2	2019-12-09 23:12:00.93 0807+00:00	Drops excluded from analysis: 100 drops Run: 09-Nov-2018 testgroup8 100_20181109_142408
19-OCT-2018CTR3	Update Dispenser	venturf2	2019-12-09 23:12:00.93 0807+00:00	Dispenser information updated: reagent: water, stress: Purple
S2SL7S4	Rename Dispenser	venturf2	2019-12-09 23:12:00.93 0807+00:00	Dispenser renamed from D1 to 19-OCT-2018CTR3

Operator Signature**Date (DD-MMM-YYYY)****Print Name****Reviewer Signature****Date (DD-MMM-YYYY)****Print Name**