

Clinical Validation Summary

Retrospective and prospective results to July 2026

Retrospective validation

24,000 held out studies from six sites not used in training. Sensitivity 94.1 percent, specificity 88.7 percent, area under the curve 0.961 for the combined critical finding endpoint.

Prospective arm

Two hospitals, March to July 2026, 11,400 consecutive studies. Sensitivity held at 92.8 percent in live conditions. Twelve false negatives were reviewed case by case, of which nine were subtle findings the original report also missed.

Clinical impact

Mean time from image acquisition to radiologist review for critical studies fell from 4 hours 12 minutes to 31 minutes at the two prospective sites.

Detailed results

Finding	Studies	Sensitivity	Specificity	AUC
Intracranial haemorrhage	6,410	95.8%	90.1%	0.974
Midline shift	2,880	93.2%	91.4%	0.968
Hydrocephalus	1,940	91.7%	89.6%	0.951
Pneumothorax	3,120	94.4%	88.2%	0.963
Pleural effusion	5,270	93.9%	86.8%	0.955
Consolidation	4,380	92.1%	85.4%	0.941
Combined critical endpoint	24,000	94.1%	88.7%	0.961

Publications

One abstract accepted at the Indian Radiological and Imaging Association annual conference 2026. A full manuscript is in preparation for submission in Q3 FY27.