

CDSCO Filing Acknowledgement

Central Drugs Standard Control Organisation

Application

Application for manufacturing licence in Form MD 3 for a Class B software as a medical device, filed through the ONDLS portal on 14 April 2026. Acknowledgement reference MD 15 2026 0417.

Device particulars

- Generic name: radiological image processing software for triage.
- Risk class: B, per the Medical Devices Rules 2017 classification for software intended to inform clinical management without immediate diagnostic claim.
- Manufacturing site: Kochi, Kerala, audited to ISO 13485 in February 2026.

Documents submitted

Device master file, plant master file, essential principles checklist, risk management file to ISO 14971, clinical evaluation report, software lifecycle documentation to IEC 62304, and the post market surveillance plan.

Status

Under technical review by the state licensing authority as at 31 July 2026. One query raised on 9 June regarding the labelling of the notification screen, responded to on 17 June. No further queries received.