OrthoCast AI — Cast Generation Report

PROTOTYPE REPORT — NOT APPROVED FOR PATIENT USE

Cast ID: cast_4df817091e3a
Region: wrist
Generated (UTC): 2026-10-12 08:35:53
Status: prototype_generated

Geometry Parameters

Parameter	Value
Station diameters (mm)	36.9, 45.2, 51.8
Cast length (mm)	152.0
Wall thickness (mm)	4.5
Ventilation enabled	False
Vent hole (mm)	6.0
Vent spacing (mm)	18.0

Mesh Validation Checks

File	Watertight	Faces	Volume (mm³)
cast_half_A.stl	True	3196	43151.45
cast_half_B.stl	True	3196	43151.45

Warnings and Limitations

• Prototype geometry only; not approved for patient use.

• Mesh checks do not establish structural strength, fit, biocompatibility, or clinical safety.

• Existing CAD closure tabs are prototype geometry; the requested tab count is not yet configurable in the CAD engine.

• Material and print-time figures from the legacy generator are estimates, not slicer-verified results.

A watertight mesh check is a geometric check only. This report does not certify fit, strength, material safety, printability, or clinical suitability.