OrthoCast AI — Cast Generation Report

PROTOTYPE REPORT — NOT APPROVED FOR PATIENT USE

Cast ID: cast_a843307eea2a
Region: wrist
Generated (UTC): 2026-10-12 10:39:14
Status: prototype_generated

Geometry Parameters

Parameter	Value
Station diameters (mm)	74.4, 91.7, 105.4
Cast length (mm)	286.0
Wall thickness (mm)	4.4
Ventilation enabled	True
Vent hole (mm)	6.0
Vent spacing (mm)	18.0

Mesh Validation Checks

File	Watertight	Faces	Volume (mm³)
cast_half_A.stl	False	20240	None
cast_half_B.stl	False	20231	None

Warnings and Limitations

• Prototype geometry only; not approved for patient use.

• Mesh checks do not establish structural strength, fit, biocompatibility, or clinical safety.

• Existing CAD closure tabs are prototype geometry; the requested tab count is not yet configurable in the CAD engine.

• Material and print-time figures from the legacy generator are estimates, not slicer-verified results.

A watertight mesh check is a geometric check only. This report does not certify fit, strength, material safety, printability, or clinical suitability.