



CERTIFICATE OF CALIBRATION

Issued By: Vericas Integrated Services Limited

2 Peace Valley Estates Woji Port Harcourt

Email: vericasintegrated@yahoo.com, vericasintegrated@gmail.com

Phone: +234-7036091657, +234-8120023534

Client:

ELSHCON NIGERIA LIMITED
Trans-Woji Road PH Rivers State.

Certificate Number:

VISL-CAS-20-047

Page 1 of 1

Date Calibrated:

16 June, 2020.

Due Date:

15 December, 2020

Calibration Validity:

Six (6) Months

Environmental Conditions:

Temperature: 25°C ±2°C

Relative Humidity: 76% ±20%

Instrument:

Description: Breath Analyzer
Manufacturer: Kangfu Medical Equipment
Model: KF-100AT
Serial Number: N/A
Range: 0.00 – 1.9%BAC (0.0 – 1.9g/l)

Calibration Method & Standard:

Direct Comparison with Dry Gas Standard Mixture (LIFELOC) Traceable to NIST.

Calibration Data

SET NO.	DRY GAS CONCENTRATION	UUT VALUE (%BAC)	UUT VALUE (g/l)	ALARM
1	0.00%BAC	0.00	0.0	No Alarm
2	0.5%BAC	0.49	4.9	Audible Alarm
DISPLAY		OK		
RANGE		OK		
ALARM		OK		

Calibrated By:

Izevbekhai O. Gabriel

Remark: Equipment is fit for Purpose

