



CERTIFICATE OF CALIBRATION

Issued By: Vericas Integrated Services Limited

2 Peace Valley Estates Woji Port Harcourt

Email: vericasintegrated@yahoo.com

Phone: +234-7036091657, +234-8120023534

Client: ELSHCON NIGERIA LIMITED
GRA, PH Rivers State.

Certificate Number:
VISL-CAS-19-006

Page 1 of 1

Date Calibrated: 17 January, 2019.
Due Date: 16 July, 2019
Calibration Validity: Six (6) Months

Environmental Conditions:
Temperature: 25°C \pm 2°C
Relative Humidity: 76% \pm 20%

Instrument:
Description: Breath Analyzer
Manufacturer: Kangfu Medical Equipment
Model: KF-100AT
Serial Number: N/A
Range: 0.00 – 1.9%BAC (0.0 – 1.9g/l)

Calibration Method & Standard:

Direct Comparison with Dry Gas Standard Mixture (LIFELOC) Traceable to NIST.

Calibration Data

SET NO.	DRY GAS CONCENTRATION	UUT VALUE (%BAC)	UUT VALUE (g/l)	ALARM
1	0.00%BAC	0.00	0.0	No Alarm
2	0.5%BAC	0.49	4.9	Audible Alarm
DISPLAY		OK		
RANGE		OK		
ALARM		OK		

Calibrated By:

Izevbekhai O. Gabriel
Calibration Expert

Remark: Equipment is fit for Purpose



This certificate provides traceability of measurement to recognized NIST.
Copyright of this certificate is owned by Vericas Integrated Services Limited (VISL) and may not be reproduced except with the prior written approval of VISL.