



CERTIFICATE OF CALIBRATION

Issued By: Vericas Integrated Services Limited
2 Peace Valley Estates Woji Port Harcourt
Email: vericasintegrated@yahoo.com
Phone: +234-7036091657, +234-8120023534

Client: ELSHCON NIGERIA LIMITED
Trans-Woji Road PH Rivers State.

Certificate Number:
VISL-CAS-19-005

Page 1 of 1

Date Calibrated: 17 January, 2019.
Due Date: 16 July, 2019
Calibration Validity: Six (6) Months

Environmental Conditions:
Temperature: 25°C ±2°C
Relative Humidity: 76% ±20%

Instrument:
Description: Portable Gas Detector
Manufacturer: Drager Safety
Model: X-am 2500
Serial Number: ARHB-4617

Calibration Method & Standard:

Direct Comparison Adjustment with Standard Drager Calibration Gas Mixture Traceable to NIST.

Calibration Data

CHANNEL NO.	GAS NAME	UNITS	MEASUREMENT RANGE	STANDARD CALIBRATION GAS	RESOLUTION	ALARM 1	ALARM 2	STEL	TWA
1	LEL (CH ₄)	%LEL	0-100%LEL	50	1%LEL	10	20	-	-
2	OXYGEN (O ₂)	%VOL	0-25%Vol.	18.0	0.1%	19.5	23.0	-	-
3	HYDROGEN SULPHIDE (H ₂ S)	PPM	0-100PPM	15	1PPM	10	15	15	10
4	CARBON MONOXIDE (CO)	PPM	0-2000PPM	50	1PPM	35	50	50	35

Pre Calibration: Pass
Calibration Result: Pass
Alarms: OK

DISPLAY	RANGES
OK	OK

Calibrated By:

Izevbekhai O. Gabriel
Calibration Engineer.



Remark: Equipment is fit for Purpose

This certificate provides traceability of measurement to recognized NIST.
Copyright of this certificate is owned by Vericas Integrated Services Limited (VISL) and may not be reproduced except with the prior written approval of VISL.