

CERTIFICATE OF CONFORMANCE



S.E. International, Inc.

P.O. Box 39, 436 Farm Rd. Summertown, TN 38483

1.800.293.5759 | Fax: 1.931.964.3564

www.seintl.com | radiationinfo@seintl.com

This is to certify that this instrument was manufactured using the standards of MIL-45208-A, and that the instrument was tested using standards whose accuracies are traceable to the National Institute of Standards and Technology. The accuracy is typically $\pm 15\%$ of reading relative to Cs137. This Electronic Calibration complies with ANSI-Z-540.

This certificate warrants that the factory calibration is valid for 12 months from the date placed in service. If you would like to receive a notice for annual calibration service, please fill out the calibration database form in the back of the operation manual and send it to the address above.

NOTICE: * The "Due Date of Recalibration" may be established (by the customer) by adding the "Cal Interval" to the "Date (the instrument) is Placed in Service" (Battery Installed).

* Date Placed in Service: 08-03-2022

* Due Date of Recalibration: 07-03-2023

Respectfully
Submitted by:
Robbin Cramer
Quality Assurance

R Cramer

(Place sticker on instrument)

