



GE Healthcare

Technical Publications

2302323-100

Rev.9



E8C Type Probe/E8C-RS Type Probe User Manual

Operating Documentation

Copyright© 2001-2004, 2008-2009 By General Electric Co.

Regulatory Requirement

The E8C Type probe complies with regulatory requirements of the following European Directive 93/42/EEC concerning medical devices for the ultrasound system.



The E8C-RS Type Probe complies with regulatory requirements of the following European Directive 93/42/EEC concerning medical devices for the ultrasound system.



GE Healthcare Japan Corporation
7-127, Asahigaoka 4-chome, Hino-shi, Tokyo
191-8503, JAPAN



GE Medical Systems Information Technologies GmbH
(GEMS IT GmbH)
Munzinger Strasse 3, D-79111 Freiburg, GERMANY
Tel: +49 761 45 43 -0; Fax: +49 761 45 43 -233

Revision History

Reason for Change

REV	DATE	REASON FOR CHANGE
Rev. 0	June 1st, 2001	Initial Release
Rev. 1	Aug. 10, 2001	Add the applicable system
Rev. 2	May 14, 2002	Add E8C-RS Probe
Rev. 3	Dec. 25, 2002	Add the applicable system
Rev. 4	May 27, 2003	Add the manufacturer and the applicable system
Rev. 5	Feb. 26, 2004	Add the description of a removal tool and Cidex Opa
Rev. 6	April 17, 2008	Remove the applicable system description from Page 1.
Rev. 7	Nov. 6, 2008	Revise Eu Rep and List 3 of Storing the probe.
Rev. 8	July. 17, 2009	Add EURep symbol and Alter the manufacturer name.
Rev. 9	Dec. 25, 2009	Remove system name from Regulatory requirement description.

List of Effective Pages

PAGE NUMBER	REVISION NUMBER	PAGE NUMBER	REVISION NUMBER
Title Page	Rev.9	Table of Contents	Rev.9
Revision History	Rev.9	1 thru 16	Rev.9

Please verify that you are using the latest revision of this document. Information pertaining to this document is maintained on MyWorkshop/ePDM (GE Healthcare Electronic Product Data Management). If you need to know the latest revision, contact your distributor, local GE Sales Representative or in the USA call the GE Ultrasound Clinical Answer Center at 1 800 682 5327 or 1 262 524 5698.

This page intentionally left blank.

Table of Contents

Instructions

General - - - - -	1
Composition - - - - -	1
Specifications - - - - -	1
Outside Dimensions and Construction - - - - -	2
Operation - - - - -	3
Cleaning and Disinfecting the Probe - - - - -	6
Coupling gels - - - - -	9
Storing the Probe - - - - -	10
Contact Information - - - - -	11
Manufacturer - - - - -	15

Instructions

General

The probe is 6.5MHz micro-convex device for conducting transvaginal scan.

Composition

1. Probe body, including:
 - Head-Assy
 - Grip
 - Carrying Case
2. Operation manual
3. Filling remover tool

Specifications

Scanning Method	Micro-convex, 11R
Frequency	6.5MHz
Field of View	133 degrees
Number of elements	128
Weight	Approximately 140g (not including the cable and connector)

Outside Dimensions and Construction




Figure -1. Probe dimensions

Operation

Refer to the section entitled Connecting the probe in the ultrasound system user manual and connect the probe to the probe connector provided on the control console.



Ultrasound transducers can easily be damaged by improper handling and by contact with certain chemicals. Failure to follow these precautions can result in serious injury and equipment damage.

- Do not immerse the probe into any liquid beyond the level specified for that probe. Never immerse the transducer connector or probe adapters into any liquid.
- Avoid mechanical shock or impact to the transducer and do not apply excessive bending or pulling force to the cable.
- Transducer damage can result from contact with inappropriate coupling or cleaning agents:
 - Do not soak or saturate transducers with solutions containing alcohol, bleach, ammonium chloride compounds or hydrogen peroxide
 - Avoid contact with solutions or coupling gels containing mineral oil or lanolin
 - Avoid temperatures above 60°C.
- Inspect the probe prior to use for damage or degeneration to the housing, strain relief, lens and seal. Do not use a damaged or defective probe.

How to install

Before installing the attachment, the filling must be removed from the probe head where the biopsy guide attachment will be installed. This is only for first time installation. See Figure 2.




Figure -2. Remove the filling

1. Fill the probe sheath with acoustic coupling gel as shown in Figure 3.




Figure -3. Ultrasound gel placement

2. Place the gel-filled probe sheath on the transducer face and bring it up the probe body so that the adjustment knob of the probe is completely covered with the probe sheath as shown in Figure 4.



Figure -4. Probe sheath application



Do not use pre-lubricated condoms as a sheath. In some cases, they may damage the probe. Lubricants in these condoms may not be compatible with probe construction.



DO NOT use an expired probe sheath. Before using probe sheaths, verify whether the term of validity has expired.

3. Apply gel to outer surface of the probe sheath at the transducer location for scanning.

4. Figure 5 shows the relationship between the displayed image direction during an axial scan compared to the probe and probe position.




Figure -5. Scan plane orientation

Cleaning and Disinfecting the Probe

The probe is driven with electrical energy that can injure the patient or user if live internal parts are contacted by conductive solution.



DO NOT immerse the probe into any liquid beyond the level indicated by the immersion level in Figure 6 on page-9. Never immerse the probe connector or probe adaptors in any liquid.

Probe Cleaning Process

Avoid cross contamination. Follow all infection control policies established by your office, department or hospital as they apply to personnel and equipment.

1. Disconnect the probe from the ultrasound console and remove all coupling gel from the probe by wiping with a soft cloth and rinsing with flowing water.
2. Wash the probe with mild soap in lukewarm water. Scrub the probe as needed using a soft sponge, gauze, or cloth to remove all visible residue from the probe surface. Prolonged soaking or scrubbing with a soft bristle brush (such as a toothbrush) may be necessary if material has dried onto the probe surface.
3. Rinse the probe with enough clean potable water to remove all visible soap residue.
4. Air dry or dry with a soft cloth.



DO NOT clean any portion of the probe with methanol, ethanol, isopropanol, or any other alcohol based cleaner. Such substances can cause irreparable damage to the probe.

Disinfecting probes



Ultrasound probes can be disinfected using liquid chemical germicides. The level of disinfection is directly related to the duration of contact with the germicide. Increased contact time procedures a higher level of disinfection. 2% Glutaraldehyde-based solutions have been shown to be very effective for this purpose. Cidex and Cidex Opa are the germicide that have been evaluated for compatibility with the material used to construct the probes.

Refer to the Probe Care Card or http://www.gehealthcare.com/user/ultrasound/products/probe_care.html.



In order for liquid chemical germicides to be effective, all visible residue must be removed during the cleaning process. Thoroughly clean the probe as described earlier before attempting disinfection.

1. Prepare the germicide solution according to the manufacturer's instructions. Be sure to follow all precautions for storage, use and disposal.
2. Place the cleaned and dried probe in contact with the germicide for the time specified by the germicide manufacturer. High-level disinfection is recommended for surface probes and is required for endocavitary and intraoperative probes (follow the germicide manufacturer's recommended time).


Probes for neuro surgical intra-operative use must not be sterilized with liquid chemical sterilants because of the possibility of neuro toxic residues remaining on the probe. Neurological procedures must be done with the legally marketed, sterile, pyrogenfree probe sheaths.

3. After removing from the germicide, rinse the probe following the germicide manufacturer's rinsing instructions. Flush all visible germicide residue from the probe and allow to air dry.



CREUTZFELD-JACOB DISEASE

Neurological use on patients with this disease must be avoided. If a probe becomes contaminated, there is no adequate disinfecting means.



High Level Probe Disinfection




Figure -6. Probe immersion level



- DO NOT attempt to disinfect or sterilize the probe with an autoclave or steam sterilization.
- DO NOT clean the probe by immersing in an alcohol solution.
- DO NOT soak, disinfect, or sterilize the connector and relief.
- DO NOT immerse the grip part of the probe.

Coupling gels



Do not use unrecommended gels (lubricants). They may damage the probe and void the warranty.

Applying

In order to assure optimal transmission of energy between the patient and probe, a conductive gel or couplant must be applied liberally to the patient where scanning will be performed.



Do not apply gel to the eyes. If there is gel contact to the eye, flush eye thoroughly with water.

Precautions

Coupling gels should not contain the following ingredients as they are known to cause probe damage:

- Methanol, ethanol, isopropanol, or any other alcohol-based product
- Mineral oil
- Iodine
- Lotions
- Lanolin
- Aloe Vera
- Olive Oil
- Methyl or Ethyl Parabens (para hydroxybenzoic acid)
- Dimethylsilicone
- Polyether glycol based

Storing the Probe

1. After cleaning and disinfection, store the probe in it's carrying case for protection.
2. Store the probe even when traveling for short distances in its carrying case or in the system probe holder.
3. The probe should be stored or transported within the parameters outlined below.

Table 1: 2D Probe Environmental Requirements

	Operational	Storage	Transport
Temperature	10° - 40° C 50° - 104° F	-10° - 60° C 14° - 140° F	-40° - 60° C -40° - 140° F
Humidity	30 - 85% non-condensing	30 - 90% non-condensing	30 - 90% non-condensing
Pressure	700 - 1060hPa	700 - 1060hPa	700 - 1060hPa

The following maintenance schedule is suggested for the system and probes to ensure optimum operation and safety..

Table 2: Planned Maintenance Program

Do the Following	Daily	After Each Use	As Necessary
Inspect the probes	X		
Clean the probes		X	
Disinfect the probes		X	
Sterilize probes			X

Contact Information

Contacting GE Healthcare Ultrasound

For additional information or assistance, please contact your local distributor or the appropriate support resource listed on the following pages:

INTERNET

<http://www.gehealthcare.com>

http://www.gehealthcare.com/usen/ultrasound/products/probe_care.html

USA

GE Healthcare TEL: (1) 800-437-1171
Ultrasound Service Engineering FAX: (1) 414-721-3865
9900 Innovation Drive
Wauwatosa, WI 53226

Clinical Questions

For information in the United States, Canada, Mexico and parts of the Caribbean, call the Customer Answer Center
TEL: (1) 800-682-5327 or (1) 262-524-5698

In other locations, contact your local Applications, Sales or Service Representative.

Service Questions

For service in the United States, call GE CARES

TEL: (1) 800-437-1171

In other locations, contact your local Service Representative.

Accessories Catalog Requests

To request the latest GE Accessories catalog or equipment brochures in the United States, call the Response Center

TEL: (1) 800-643-6439

In other locations, contact your local Applications, Sales or Service Representative.

Contacting GE Healthcare Ultrasound (continued)

Placing an Order To place an order, order supplies or ask an accessory-related question in the United States, call the GE Access Center

TEL: (1) 800-472-3666

In other locations, contact your local Applications, Sales or Service Representative.

CANADA GE Healthcare TEL: (1) 800-664-0732
Ultrasound Service Engineering
9900 Innovation Drive
Wauwatosa, WI 53226
Customer Answer Center TEL: (1) 262-524-5698

LATIN & SOUTH AMERICA GE Healthcare TEL: (1) 262-524-5300
Ultrasound Service Engineering
9900 Innovation Drive
Wauwatosa, WI 53226
Customer Answer Center TEL: (1) 262-524-5698

EUROPE GE Ultraschall TEL: 0130 81 6370 toll free
Deutschland GmbH & Co. KG TEL: (33) 130.831.300
Beethovenstrasse 239 FAX: (49) 212.28.02.431
Postfach 11 05 60
D-42655 Solingen

ASIA GE Ultrasound Asia (Singapore) TEL: 65-291 8528
Service Department - Ultrasound FAX: 65-272-3997
298 Tiong Bahru Road #15-01/06
Central Plaza
Singapore 169730

JAPAN GE Healthcare Japan Corporation TEL: (81) 42-648-2910
Customer Service Center FAX: (81) 42-648-2905

Contacting GE Healthcare Ultrasound (continued)

ARGENTINA	GEME S.A. TEL: (1) 639-1619 Miranda 5237 FAX: (1) 567-2678 Buenos Aires - 1407
AUSTRIA	GE GesmbH Medical Systems Austria TEL: 0660 8459 toll free Prinz Eugen Strasse 8/8 FAX: +43 1 505 38 74 A-1040 WIEN TLX: 136314
BELGIUM	GE Healthcare BVBA TEL: +32 0 2 719 72 04 Kouterveldstraat 20 FAX: +32 0 2 719 72 05 B-1831 Diegem
BRAZIL	GE Sistemas Medicos TEL: 0800-122345 Av Nove de Julho 5229 FAX: (011) 3067-8298 01407-907 Sao Paulo SP
CHINA	GE Healthcare - Asia TEL: (8610) 5806 9403 No. 1, Yongchang North Road FAX: (8610) 6787 1162 Beijing Economic & Technology Development Area Beijing 100176, China
DENMARK	GE Medical Systems TEL: +45 4348 5400 Fabriksparken 20 FAX: +45 4348 5399 DK-2600 GLOSTRUP
FRANCE	GE Medical Systems TEL: 05 49 33 71 toll free 738 rue Yves Carmen FAX: +33 1 46 10 01 20 F-92658 BOULOGNE CEDEX
GERMANY	GE Ultraschall TEL: 0130 81 6370 toll free Deutschland GmbH & Co. KG TEL: (49) 212.28.02.207 Beethovenstrasse 239 FAX: (49) 212.28.02.431 Postfach 11 05 60 D-42655 Solingen
GREECE	GE Medical Systems Hellas TEL: +30 1 93 24 582 41, Nikolaou Plastira Street FAX: +30 1 93 58 414 G-171 21 NEA SMYRNI
ITALY	GE Medical Systems Italia TEL: 1678 744 73 toll free Via Monte Albenza 9 FAX: +39 39 73 37 86 I-20052 MONZA TLX: 3333 28
LUXEMBOURG	TEL: 0800 2603 toll free

Contacting GE Healthcare Ultrasound (continued)

MEXICO	GE Sistemas Medicos de Mexico S.A. de C.V. Rio Lerma #302, 1° y 2° Pisos TEL: (5) 228-9600 Colonia Cuauhtemoc FAX: (5) 211-4631 06500-Mexico, D.F.
NETHERLANDS	GE Medical Systems Nederland B.V. TEL: 06 022 3797 toll free Atoomweg 512 FAX: +31 304 11702 NL-3542 AB UTRECHT
POLAND	GE Medical Systems Polska TEL: +48 2 625 59 62 Krzywickiego 34 FAX: +48 2 615 59 66 P-02-078 WARSZAWA
PORTUGAL	GE Medical Systems Portuguesa S.A. TEL: 05 05 33 7313 toll free Rua Sa da Bandeira, 585 FAX: +351 2 2084494 Apartado 4094 TLX: 22804 P-4002 PORTO CODEX
RUSSIA	GE VNIEM TEL: +7 495 739 6931 18C, Krasnopresnenskaya nab. FAX: +7 495 739 6932 MOSCOW 123317 TLX: 613020 GEMED SU
SPAIN	GE Healthcare TEL: +34 (91) 663 25 00 Avda. Europa, 22 FAX: +34 (91) 663 25 01 E-28108 Alcobendas, Madrid
SWEDEN	GE Medical Systems TEL: 020 795 433 toll free PO-BOX 1243 FAX: +46 87 51 30 90 S-16428 KISTA TLX: 12228 CGRSWES
SWITZERLAND	GE Medical Systems (Schweiz) AG TEL: 155 5306 toll free Sternmattweg 1 FAX: +41 41 421859 CH-6010 KRIENS
TURKEY	GE Healthcare Turkiye TEL: +90 212 366 29 00 Sun Plaza FAX: +90 212 366 29 99 Dereboyu Sok. No 24/7 34398 Maslak ISTANBUL

Contacting GE Healthcare Ultrasound (continued)

UNITED KINGDOM	GE Medical Systems TEL: 0800 89 7905 toll free Coolidge House FAX: +44 753 696067 352 Buckingham Avenue SLOUGH Berkshire SL1 4ER
OTHER COUNTRIES	NO TOLL FREE TEL: international code + 33 1 39 20 0007

Manufacturer

GE Medical System (China) Co., Ltd.
No. 19, Changjiang Road
WuXi National Hi-Tech Development Zone
Jiangsu, P.R. China 214028
TEL: +86 510 5225888; FAX: +86 510 5226688

GE Healthcare Japan Corporation
7-127, Asahigaoka 4-Chome
Hino-Shi
Tokyo, 191-8503
JAPAN

