

GeneLab Standard Operating Procedure: Tissue homogenization using Bullet Blender Gold Bead Beater

May 2020

Version 1.0



## **Document Revisions**

| Document<br>Number | Revision<br>Number | Date     | Description of Changes |
|--------------------|--------------------|----------|------------------------|
| GL-SOP-2.1         | 1.0                | May 2020 | Original document      |
|                    |                    |          |                        |
|                    |                    |          |                        |
|                    |                    |          |                        |

## Scope and Purpose

The procedure below describes the steps required to lyse and homogenize biological material using a Bullet Blender 24 Gold bead beater. This procedure was validated for downstream RNA/DNA extraction using the Qiagen AllPrep kits (SOP #3.1), it is also possible to use the procedure with downstream RNA extraction using Trizol (SOP #3.2).

## **Equipment and Consumables**

- 1. Eppendorf Centrifuge 5424/5424 R
- 2. Styrofoam box and dissection tools (forceps, scalpel, scissors)
- 3. Bullet Blender 24 Gold (NextAdvanced Cat#BB24-AU)
- 4. Sterile Microcentrifuge Tube 1.5 mL RINO® (NextAdvanced Cat# TUBE1R5-S or Navy, Red, Green kits, refer to Table 1 to select appropriate tube/bead combination)

## Procedure

- 1. Obtain dry ice, it will be used to maintain a cold environment inside the bead beater.
- 2. Select the homogenization bead type, homogenization duration and homogenization speed setting based on the material processed, using table 1 below:
  - a. If using beads that are not part of the kit, use RINO tubes (Next Advance #TUBE1R5-S).
  - b. Kits (Navy, Red, Green) have tubes pre-filled with beads.



Table 1: Validated homogenization settings (preferred condition in bold)

| Tissue/Bead Type     | NAVY kit (Next<br>Advanced<br>#NAVYR5)/Speed<br>setting | Zirconium Oxide 2.0mm (Next Advance #ZROB20- RNA)/Speed setting | Red kit (Next<br>Advanced<br>#REDR5-<br>RNA)/Speed<br>setting | SSB (Next<br>Advanced<br>#SSB14B-<br>RNA)/Speed<br>setting |
|----------------------|---|---|---|--|
| Liver                | 3min/12   | 3min/12   |   |  |
| Thymus               | 3min/12   |   |   |  |
| Feces                | 3min/12   |   |   |  |
| Colon                | 3min/12   | 5min/12   |   |  |
| Lung                 |   | 3min/12   |   |  |
| Skin                 | TBD   | TBD   |   |  |
| Spleen               |   | 5min/12   |   |  |
| Muscle               |   |   |   | 5min/12  |
| Kidney               | 3min/12   |   |   |  |
| Brain                | 3min/10   |   | 3min/10   |  |
| Heart                |   |   |   | 5min/12  |
| Lymph Nodes          |   |   | 3min/10   |  |
| White Adipose        |   |   | 3min/10   |  |
| <b>Brown Adipose</b> |   |   | 3min/10   |  |
| Reproductive Tract   | 3min/12   |   |   |  |
| Adrenal Gland        |   | 3min/12   |   |  |

- 3. Select a lysis buffer depending on the downstream extraction protocol used:
  - a. If following SOP #3.1, on the day of the extraction prepare 600-800uL of QIAGEN buffer RLT + 1% b-ME per sample.
  - b. If following SOP #3.2, on the day of the extraction prepare 600-800uL of TRIzol solution per sample.
  - c. If following SOP #3.3, on the day of the extraction prepare 1000uL CTAB buffer per sample.
- 4. Pre-chill a centrifuge that can fit RINO tubes to 4°C (Suggested Eppendorf Centrifuge 5424/5424 R, this equipment will be used in the consecutive extraction SOP's).
- 5. Fill the Bullet Blender Gold (BBG) side compartment with dry ice pellets or crushed dry ice.



Figure 1: Ice compartment on left side.

- 6. Pre-chill the bullet blender by closing the tube compartment and turning the BBG on for 5min at speed setting 12.
- 7. When the BBG is cold, a blue "4°C" indicator light will show on the front screen.
- 8. Follow SOP#1.2 for tissue cutting procedures. When completed, you should have lysis tubes with a tissue ready.
- 9. Close the RINO tubes **tightly** and place them inside the pre-chilled BBG. Make sure to equally distribute the weight of the RINO tubes throughout the rotor.




Figure 2: Example of a well-balanced rotor with six sample tubes (in blue).

- 10. Homogenize the tissues following the speed and time setting indicated in Table 1.
- 11. Move RINO tubes with the homogenate into the centrifuge and spin for 5min at full speed in 4°C. (Full speed using Eppendorf 5424 is 21130 x g/15000 rpm.)
- 12. Carefully transfer lysate in to clean, labeled 1.5ml Eppendorf tube and proceed with SOP #3.1/3.2/3.3.
- 13. Take out the dry ice pellets and return to storage.
- 14. Wipe the inside of the dry ice pocket and the tube rotor.