

LEGAL METROLOGY DETAILED TEST REPORT

NON-AUTOMATIC WEIGHING INSTRUMENT (NAWI) - TECHNICAL AUDIT RECORD

Regulatory Standard: OIML R 76-1:2006 | Evaluation Mode: SUBSEQUENT_VERIFICATION

Generated: 2026-09-26T17:05:35.944Z | Tamper-Evident ID: R76-2026-B6B57DE7



REPORT NO: R76-2026-000072

QR VERIFICATION ID: R76-2026-B6B57DE7

1. LABORATORY INFORMATION

Facility: Central Legal Metrology Testing Laboratory

Address: Central Metrology Directorate

Assigned Technician: Pramod Patil

2. ENVIRONMENTAL CONDITIONS

Ambient Temperature: 23 °C

Relative Humidity: 50 %

Atmospheric Pressure: 1013.25 hPa

Compliance: Within standard metrological tolerance limits

3. INSTRUMENT TECHNICAL SPECIFICATION

Manufacturer: Essae-Teraoka

Model: DS-252

Serial No: ES-84210

Type: Non-Automatic Weighing
Instrument (Bench/Platform Scale)
Interval (d): 0.005 kg

Accuracy Class: Class III

Max: 30 kg

Min: 0.2 kg

Verification (e): 0.01 kg

Resolution (n): 3,000 intervals

Unit: kg

Indication: Electronic Digital

4. DETAILED METROLOGICAL TEST AUDITS & CALCULATION TRACES

TEST: WEIGHING PERFORMANCE & INTRINSIC ERROR (WEIGHING PERFORMANCE)

Reference: OIML R 76-1:2006 Clause 3.5.1 & A.4.4 STATUS: PASS

Formulae: $P = I + 0.5e - 9d \hat{A} \hat{A}$ & w Error $E = P - L$ | Corrected Error $E_c = E - E_0$

No.	Phase	Load L (kg)	Ind. I (kg)	dL	E0	Raw E	Corr. Ec	MPE (+/-)	Decision
1	LOADING	0.000	0.000	0.005	0.000	0.000	0.000	+/-	PASS
2	LOADING	7.500	7.500	0.005	0.000	0.000	0.000	+/-	PASS
3	LOADING	15.000	15.005	0.005	0.000	0.005	0.005	+/-	PASS
4	LOADING	30.000	30.005	0.005	0.000	0.005	0.005	+/-	PASS

TEST: REPEATABILITY TEST (REPEATABILITY)

Reference: OIML R 76-1:2006 Clause 3.6.1 & A.4.4 STATUS: PASS

Evaluation Formula: $\Delta I = I_{max} - I_{min}$ | Applicable Limit: MPE at Test Load

RECORDED TEST READINGS (RUNS):

R1: 15.000 kg | R2: 15.000 kg | R3: 15.002 kg | R4: 15.000 kg | R5: 15.000 kg | R6: 15.002 kg | R7: 15.000 kg | R8: 15.000 kg | R9: 15.002 kg | R10: 15.000 kg

Test Load: 15.000 kg

Max Indication (Imax): 15.002 kg

Min Indication (Imin): 15.000 kg

Difference (Delta I): 0.003 kg

Applicable MPE: 0.010 kg

Decision: PASS

TEST: ECCENTRIC LOADING TEST (ECCENTRIC_LOADING)

Reference: OIML R 76-1:2006 Clause 3.6.2 & A.4.4 STATUS: PASS

Evaluation: Test Load = 1/3 Max (10.000 kg) | Error at each position <= MPE

Position	Applied Load (kg)	Indication (kg)	dL	Raw Error E	Corrected Ec	MPE	Result
CENTER	10.000	10.000	0.005	0.000	0.000	±0.010	PASS
FRONT_LEFT	10.000	10.005	0.005	0.005	0.005	±0.010	PASS
BACK_LEFT	10.000	10.005	0.005	0.005	0.005	±0.010	PASS
BACK_RIGHT	10.000	10.005	0.005	0.005	0.005	±0.010	PASS
FRONT_RIGHT	10.000	10.005	0.005	0.005	0.005	±0.010	PASS

TEST: ZERO-SETTING & ZERO-TRACKING EVALUATION (ZERO SETTING)

Reference: OIML R 76-1:2006 Clause 4.5 & A.4.2

STATUS: PASS

Initial Indication: 0.000 kgChangeover 9D 0.005 kgCalculated Zero Error E0: 0.000 kg

Applicable Limit: ±0.25 e (±0.003 kg)Result: PASS

TEST: DISCRIMINATION TEST (DISCRIMINATION)

Reference: OIML R 76-1:2006 Clause 3.8 & A.4.8

STATUS: PASS

Requirement: Extra load of 1.4 d smoothly placed must step indication to I + d (Clause 3.8 / A.4.8)

Base Load (L): 18.000 kgInitial Indication: 18.000 kgAdditional Load: 0.0070 kg (1.4 d)

Resulting Indication: 18.005 kgObserved Shift (9D) 0.0050 kgMandated Shift: 0.0050 kg (1.0 d)

Decision: PASSNote: Addition of 1.4d extra load (0.0070 kg) produced the mandated indication st

TEST: MULTIPLE INDICATING DEVICES COMPARISON (MULTIPLE INDICATING DEVICES)

Reference: OIML R 76-1:2006 Clause 3.6.3

STATUS: NOT_APPLICABLE

This test module is not applicable to the current instrument configuration.

TEST: DIFFERENT POSITIONS OF EQUILIBRIUM (DIFFERENT POSITIONS OF EQUILIBRIUM)

Reference: OIML R 76-1:2006 Clause 3.6.4

STATUS: NOT_APPLICABLE

This test module is not applicable to the current instrument configuration.

OVERALL VERIFICATION RESULT: PASSED (COMPLIANT)

All 5 applicable OIML R 76-1:2006 metrological tests have been verified compliant with Maximum Permissible Errors.

PUBLIC VERIFICATION RECORD & TRACEABILITY

Permanent Verification ID: R76-2026-B6B57DE7
Authentic URL to inspect full measurement calculations:
<http://localhost:5173/public/report/R76-2026-B6B57DE7>

OFFICIAL ATTESTATION & SIGN-OFF

Officer: Pramod Patil
Facility: Central Legal Metrology Testing Laboratory
[DIGITALLY RECORDED & ARCHIVED]

