

LEGAL METROLOGY VERIFICATION CERTIFICATE

NON-AUTOMATIC WEIGHING INSTRUMENT (NAWI)

Standard: OIML R 76-1:2006 | Evaluation Mode: TYPE_EVALUATION

Issued: 2026-09-27 | Certificate No: R76-2026-000075



REPORT NUMBER: R76-2026-000075

QR VERIFICATION ID: R76-2026-1BCF9F12

1. VERIFICATION LABORATORY

Facility: Central Legal Metrology Testing Laboratory

Location: Directorate of Metrology & Standards

Authorized Officer: Pramod Patil

2. ENVIRONMENTAL TEST CONDITIONS

Ambient Temperature: 22.5 deg C

Relative Humidity: 50 %

Atmospheric Pressure: 1013.25 hPa

OIML R-76 Condition: Compliant with standard ambient tolerances

3. NON-AUTOMATIC WEIGHING INSTRUMENT (NAWI) SPECIFICATION

Manufacturer: Avery Weigh-Tronix

Model: ZM301

Serial No: AW-90214

Type: Non-Automatic Weighing
Instrument (Standard Bench Scale)
Scale Interval (d): 0.005 kg

Accuracy Class: Class III

Max: 30 kg

Min: 0.2 kg

Verification Interval (e): 0.01 kg

Resolution (n): 3,000 intervals

Unit: kg

Indication: Electronic Digital

4. METROLOGICAL EVIDENCE & COMPLIANCE SUMMARY (OIML R 76-1:2006)

CODE	TEST MODULE & CLAUSE	KEY EVIDENCE & CALCULATIONS	APPLIED LIMIT	DECISION
WEIGHING	Weighing Performance & OIML R 76-1:2006 Clause 3.5.1 & A.4.4.3	Max corr. error: 0.006 kg	MPE: +/-0.015 kg	PASS
REPEATABILITY	Repeatability Test OIML R 76-1:2006 Clause 3.6.1 & A.4.10	Imax: 15.009, Imin: 14.999, Delta I: 0.010 kg	MPE: +/-0.010 kg	PASS
ECCENTRIC	Eccentric Loading Test OIML R 76-1:2006 Clause 3.6.2 & A.4.7	5 points tested, max corr: 0.004 kg	MPE at 1/3 Max	PASS
ZERO_SETTING	Zero-Setting & Zero-Tr OIML R 76-1:2006 Clause 4.5 & A.4.2	Zero error E0: 0.001 kg	+/-0.25 e	PASS
DISCRIMINATION	Discrimination Test OIML R 76-1:2006 Clause 3.8 & A.4.8	Mandated I+d step response confirmed	1.4 d trigger	PASS
MULTIPLE_INDICATING_DEVICES	Multiple Indicating De OIML R 76-1:2006 Clause 3.6.3	Not applicable to device specification	N/A	NOT_APPLICABLE
DIFFERENT_POSITIONS_OF_EQUIBRIBUM	Different Positions of OIML R 76-1:2006 Clause 3.6.4	Not applicable to device specification	N/A	NOT_APPLICABLE

OVERALL VERIFICATION RESULT: PASSED (COMPLIANT)

All 5 applicable OIML R 76-1:2006 metrological tests have been verified compliant with Maximum Permissible Errors.

TAMPER-EVIDENT QR VERIFICATION

Verification ID: R76-2026-1BCF9F12

Public verification portal URL:

<http://localhost:5173/public/report/R76-2026-1BCF9F12>

METROLOGICAL AUTHORITY ATTESTATION

Authorized Inspector: Pramod Patil

Laboratory: Central Legal Metrology Testing Laboratory

[DIGITALLY CERTIFIED & VERIFIED]