

NITRI HAND PRODUCT SPECIFICATIONS

| | |
|--|---|
| Name of Product | - 4 Mil Powder Free Nitrile Exam Gloves. |
| Name of Manufacturer | - Innovative Gloves Co. Ltd. Hatyai, Thailand. |
| Base Material | - Nitrile. |
| Characteristics | - Non-Sterile, Ambidextrous Beaded Cuff. |
| Available Sizes | - X-Small, Small, Medium, Large, X-Large, XX-Large. |
| Surface | - Textured. |
| Surface Treatment | - Polymer Coating. |
| Color | - Blue. |
| Total Extractable protein content | - < 2.0 milligram per glove. |



Physical Specifications:

| Size | Length(mm/inch) | Thickness(mm) Single Wall | | | Width(mm) |
|----------|-----------------|------------------------------|-----------|-----------|-----------|
| | | Tip | Palm | Cuff | |
| X-Small | 240±5 | 0.11±0.01 | 0.08±0.01 | 0.06±0.01 | 75±5 |
| Small | 240±5 | 0.11±0.01 | 0.08±0.01 | 0.06±0.01 | 85±5 |
| Medium | 240±5 | 0.11±0.01 | 0.08±0.01 | 0.06±0.01 | 95±5 |
| Large | 240±5 | 0.11±0.01 | 0.08±0.01 | 0.06±0.01 | 105±5 |
| X-Large | 240±5 | 0.11±0.01 | 0.08±0.01 | 0.06±0.01 | 115±5 |
| XX-Large | 240±5 | 0.11±0.01 | 0.08±0.01 | 0.06±0.01 | 120±3 |

Tensile (ASTM D412 & EN 455-2)

Before Aging

| Tensile(MPA) | Elongation(%) | Force At Break(N) |
|--------------|---------------|-------------------|
| 22 Mpa Min. | 500 Min. | ≥ 6 N |

After Accelerated Aging

| Tensile(MPA) | Elongation(%) | Force At Break(N) |
|--------------|---------------|-------------------|
| 22 Mpa Min. | 500 Min. | ≥ 6 N |

| | |
|--------------------------------|--|
| Total Residue Content | < 2.0 Milligrams per glove. |
| Packaging (Normal) | 100±2 pieces in one Dispenser Box, 10 Dispenser Box in one Carton. |
| Labeling on Dispenser (Normal) | Product Name, Size, Lot no. Quantity, Manufacturer's address. |
| Labeling on Carton (Normal) | Product Name, Size, Lot no. Quantity, Country of Origin. |

Container Loadability (No MOQ in manufacturer owned brand in standard Packaging)

| | |
|----------------------|----------------------|
| 20 FCL=1360 Cartons. | 40 FCL=2740 Cartons. |
|----------------------|----------------------|

Certification

