

Kyamwinula HC III - facility report

1. Identification

Facility	Kyamwinula HC III
Named on the programme list	Kyamwinula HCII
Type	Health centre
Local government	Tororo Mc District
Region / sub-region	Eastern / Bukedi
Date of visit	2026-08-28
Team members present	Kayongo Tonny, Lydia Nakiganda
Informant interviewed	None recorded
Supervisor	Kassim Luminsa
Supervisor's contact	+256 702 806116
Date of submission	

2. Entry

The team reached this facility on 28 August in Tororo Municipal Council. A hand-filled register naming Kyamwinula HC III is on file together with 110 photographs. Nobody is named as having received the team.

3. Condition and status

The four things the exercise exists to establish. Only what was established on site appears here; the rest is marked as not reported, not guessed.

Dimension	What was established on site
Ownership	Not reported.
Functionality	Established on the hand-filled register. Blood pressure machines carry KYAM-BPA tags and an autoclave carries KYAM-OCE-01-UGIFT.
Usage / performance	Partly established. The informant says maternity service improved but power is unreliable, the baby warmer voltage is low, the water pump is not functional and staffing is low.
Location	Recorded on the register and shown in photographs of the maternity ward, OPD, store and water tanks.

4. Supplied against found

The hand-filled register is the found side and nothing was counted on the day.

5. Engraving and identification

Several items carry KYAM- series tags on the register, the first numbered marking seen in this team's work.

6. Obsolete, unserviceable or for disposal

Nothing on the register was named for disposal.

7. Operation and maintenance

The discussion guide was answered. The facility holds a record, marks breakdowns on its inventory, and asks for solar power, staff houses and transport.

8. Queries

- Nothing here was counted on the day.
- Ownership was not established.
- The programme list carries Kyamwinula HCII and the register names level III.