# Converting Raw LC-HRMS/MS Files into mzML files

Benjamin J. Place\*

2023-03-03

# Converting To The mzML Format

To use DIMSpec tools, raw data files produced by vendor software must be converted into \*.mzML files. The easiest way to convert data files is to download and install the most recent version of ProteoWizard¹ from https://proteowizard.sourceforge.io/ to use the MSConvert tool (Adusumilli, Ravali and Mallick, Parag 2017). Once installed, follow the next steps to convert the raw file(s) to \*.mzML format.

### 1) Start MSConvert

From the Proteowizard main page, start the MSConvert program with Menu > Proteowizard > MSConvert<sup>2</sup>




Figure 1: Initial screen when running MSConvert

<sup>\*</sup>NIST | MML | CSD | benjamin.place@nist.gov

<sup>&</sup>lt;sup>1</sup>Any mention of commercial products is for information only; it does not imply recommendation or endorsement by NIST.

<sup>&</sup>lt;sup>2</sup>MSConvert includes a command line interface for advanced use cases; settings must be the same as described in this guide.

#### 2) Select the files to be converted

The MSConvert software can convert to mzML data files from the the following vendor formats at the time this guide was written: Thermo Fisher Scientific (\*.raw); Waters (\*.raw); SCIEX (\*.wiff2); Agilent (\*.D); Shimadzu (\*.LCD); and Bruker (\*.D)

- 1. Select the files to be converted using the **Browse** button at the top left (Fig 2A).

  Note: If selecting only one file at a time, instead click the Add button to include the file in the list.
- 2. Select the *Output Directory* to save files by clicking the **Browse** button at the center left (Fig 2B). **Note**: It will default to the same directory as the original file.

After loading the file, the program should look similar to Figure 2.




Figure 2: Interface for MSConvert after a data file has been loaded

# 3) Adjust the mzML file parameters in the Options box (Fig 2C).

- 1. Select mzML from the Output format input and make sure the Extension input is blank.
- 2. Select **64-bit** under *Binary encoding precision*.
- 3. Ensure boxes are checked next to Write Index, TPP Compatibility, and Use zlib compression; leave all other boxes unchecked.

The proper selections have been made in Figure 2.

### 4) Add conversion filters (Fig 2D)

Apply the following data conversion filters; to adhere to current NIST practice the following settings must be used. Select the filter to be applied from the drop down menu at the top of the **Filters** box. Do this sequentially as described here. After setting each parameter click the **Add** button. To remove a parameter (in case of an entry or selection error), select that filter in the table below Fig 2D and click the **Remove** button.

- 1. Select the *Peak Picking* filter and apply the following parameters:
  - Algorithm: Vendor or CWT
     Note: Vendor filter does not work for Waters files, you must use CWT
  - MS Levels: 1 \_ (leave the second box blank)
  - If using CWT:
    - Min. SNR: 0.1
    - Min. peak spacing: 0.1

Parameters should match those in Figure 3.

Click the Add Button



Figure 3: Peak picking filter parameters

- 2. Select the Threshold Peak Filter and include the following parameters:
  - Threshold type: Absolute intensity
  - Orientation: Most intense
  - Value: 1

Parameters should match those in Figure 4 (next page)

Click the Add Button




Figure 4: Threshold peak filter parameters

- If the raw data is from a Waters Corporation mass spectrometer, apply the Lockmass Refiner filter.
  - Reference m/z: use the mass-to-charge ratio of the lockmass, for Leucine-Enkephalin (most common), the negative ionization m/z is 554.2615 and the positive ionization m/z is 556.2771.
  - m/z Tolerance: 0.1

Parameters should match those in Figure 5 Click the  $\mathbf{Add}$  Button




Figure 5: Lockmass refiner parameters

Selected filter parameters should be very similar to those in Figure 2 prior to conversion to \*.mzML.

**Note**: The TitleMaker filter is automatically selected upon opening MSConvert; this can be included and will not affect the data analysis. However, peakPicking must still be the first line in the filters.

Note: Waters Corporation instruments should have the lockmass refiner filter in the as well.

### 5) Convert data files to the \*.mzML format.

Once all settings are as required, click the **Start** button at the bottom right (Fig 2E). This will convert all added files to \*.mzML format and put them in the assigned Output directory. Wait for the conversions to finish.

## 6) (Optional) Save the settings as a preset

Use the **Save Preset** drop down menu to use the same settings in the future (Fig 6) by selecting  $Save\ Preset$  As... To apply a preset profile to future files, select it from this same drop down menu.




Figure 6: MSConvert presets menu

Once all files have been converted to the \*.mzML format they are ready for use in other data tools.

This concludes the Quick Guide to Converting Raw LC-HRMS/MS Files into mzML files.

# References

Adusumilli, Ravali and Mallick, Parag. 2017. "Data Conversion with ProteoWizard msConvert." *Methods in Molecular Biology (Clifton, N.J.)* 1550: 339–68. https://doi.org/10.1007/978-1-4939-6747-6\_23.