SIMCO electronics

6295 FERRIS SQ, SUITE A
SAN DIEGO, CA 92121

Certificate No. **8750741**

# CERTIFICATE OF CALIBRATION
# FOR
# WOODS HOLE OCEANOGRAPHIC INST.

**297 ROSECRANS STE 3505**
**SAN DIEGO, CA 92106**

Description: **AEMC, 1060, Megohmmeter**

Serial No: **154856 DC DV** Asset No: SIMCO ID: **14769-67**

Dept: **NONE** PO No: **CC/3906/WOODS**

| Calibration Date: **10/31/2019** | Calibration Interval: **12 Months** | Next Calibration Date: **10/31/2020** |
|---|---|---|
| Arrival Condition: **MEETS MANUFACTURER'S SPEC'S.** | Service: **CALIBRATED TO MFR SPEC,& CLEAN** | |

Procedure: **33K1-4-3066-1 05/13**
Temperature: **70°F** Relative Humidity: **30%**

Standards Used:

| Manufacturer, Model | Description | SIMCO ID | Due Date | Certificate |
|---|---|---|---|---|
| **ONSET COMPUTER CORP, MX1101** | **Temperature/Humidity Logger** | **1016-555** | **09/07/2020** | **8588272** |
| **IET LABS, RS-201W-2W** | **Resistance Substituter** | **1016-313** | **12/05/2019** | **8388249** |
| **MID-EASTERN, 10 Mohm to 1 Tohm** | **Fixed Resistor Set, 6 pcs** | **1016-535** | **10/17/2020** | **8605747** |
| **GENERAL RADIO, 1409-Y** | **Standard Capacitor, 1.0 uF** | **1016-350** | **11/15/2020** | **8652923** |
| **WAVETEK, 9100** | **Calibration System** | **1016-621** | **01/21/2020** | **8395129** |
| **SIMCO, MSR100KM** | **Standard Resistor, 100 Gohm** | **1016-384** | **06/26/2021** | **8617388** |

Work performed by: **Alexander Combs**

Reviewed by:

SIMCO Electronics' quality management system conforms to ISO 9001:2015, ISO/IEC 17025:2017, and ANSI/NCSL Z540-1-1994. All calibrations are performed using internationally recognized standards traceable to the International System of Units (SI Units). Traceability is achieved through calibrations by the National Institute of Standards and Technology (NIST), other National Measurement Institutes (NMIs'), or by using natural physical constants, intrinsic standards or ratio calibration techniques. Instruments are calibrated with a test uncertainty ratio of 4:1 or greater, otherwise measurement uncertainty analysis and/or guard bands are applied during the measurement process. The information shown on this certificate applies only to the instrument identified above and may not be reproduced, except in full, without prior written consent from SIMCO Electronics. There is no implied warranty that the instrument will maintain its specified tolerances during the calibration interval due to possible drift, environment, or other factors beyond our control.

Dated: **10/31/2019**