

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
.sea-birdscientific.com

WETStar Calibration and Repairs

Date June 6, 2017 **Customer** Woods Hole Oceanographic Institution

S/N# WS3S-582P **Repair Order** 34381

Standard Service

- Performed noise test: 1 sample/sec for 60 sec
- Performed stability test: 1 sample/min for 12 hrs
- Performed temperature test: 25–2 °C
- Performed saturation test
- Shake-tested unit
- Pressure-tested unit
- Updated unit's calibration sheet

Diagnosis

Evaluated instrument and found no problems.

Repairs

Standard service.

Comments