



Biospherical Instruments Inc.

Calibration Date: 05/07/26
Model Number: QSP2200
Serial Number: 20465
Operator: JA
Standard Lamp: V-037 (11/7/25)
Operating Voltage Range: 6 to 15VDC

Job No.: R-50993

To calculate irradiance, divide the net sensor reading in Volts by the appropriate Calibration Factor

Dry Calibration Factor: 1.64E-17 Volts / (quanta/(cm²·s)) 9.89E+00 Volts / (μEinsteins/(cm²·s))
Wet Calibration Factor: 9.30E-18 Volts / (quanta/(cm²·s)) 5.60E+00 Volts / (μEinsteins/(cm²·s))

Sensor Test Data and Results:

Sensor Supply Current (Dark): 3.9 mA
Sensor Supply Voltage: 6 Volts

Sensor Calibration Signal Voltage: 153.0 mV
Sensor Dark Voltage: 0.8 mV

Lamp Integrated PAR Irradiance: 9.27E+15 quanta/(cm²·s) 0.01539 μEinsteins/(cm²·s)
Immersion Coefficient: 0.5664

Note:

1. Annual calibration is recommended.
2. Calibration is performed using a Standard of Spectral Irradiance traceable to the National Institute of Standards and Technology (NIST).