



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2725
CALIBRATION DATE: 17-Mar-26

SBE 43F OXYGEN CALIBRATION DATA

COEFFICIENTS:

Soc = 3.0753e-04 (adj)

Foffset = -856.93

Tau20 = 0.98

A = -2.7680e-03

B = 8.5269e-05

C = -1.1812e-06

E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS

D1 = 1.92634e-4

D2 = -4.64803e-2

H1 = -3.300000e-2

H2 = 5.00000e+3

H3 = 1.45000e+3

BATH OX (ml/l)	BATH TEMP (° C)	BATH SAL (PSU)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT OXYGEN (ml/l)	RESIDUAL (ml/l)
1.17	12.00	0.00	1396.519	1.17	-0.00
1.18	20.00	0.00	1507.364	1.18	-0.00
1.18	6.00	0.00	1324.803	1.18	0.00
1.19	26.00	0.00	1596.737	1.19	0.00
1.19	30.00	0.00	1655.026	1.19	0.00
1.21	2.00	0.00	1284.021	1.21	0.00
3.97	12.00	0.00	2692.312	3.97	-0.00
3.97	2.00	0.00	2262.448	3.97	-0.00
3.97	6.00	0.00	2434.361	3.97	0.00
3.98	20.00	0.00	3059.377	3.98	-0.00
4.00	26.00	0.00	3349.951	4.00	0.00
4.01	30.00	0.00	3547.158	4.01	-0.00
6.76	2.00	0.00	3250.654	6.76	-0.00
6.81	6.00	0.00	3562.051	6.81	0.00
6.88	12.00	0.00	4041.344	6.88	-0.00
6.97	20.00	0.00	4709.089	6.97	-0.00
6.97	30.00	0.00	5530.187	6.97	-0.00
7.01	26.00	0.00	5222.605	7.01	0.00

F = instrument output (Hz); T = temperature (°C); S = salinity (PSU); K = temperature (°K)

Oxsol(T,S) = oxygen saturation (ml/l); P = pressure (dbar)

Oxygen (ml/l) = Soc * (F + Foffset) * (1.0 + A * T + B * T² + C * T³) * Oxsol(T,S) * exp(E * P / K)

Residual (ml/l) = instrument oxygen - bath oxygen

