



Calibration Date: 05/07/26

Job No.: R-50994

Model Number: QSP2200

Serial Number: 20582

Operator: JA

Standard Lamp: V-037 (11/7/25)

Operating Voltage Range: 6 to 15VDC

To calculate irradiance, divide the net sensor reading in Volts by the appropriate Calibration Factor

Dry Calibration Factor:	<u>1.55E-17</u>	Volts / (quanta/(cm ² ·s))	<u>9.33E+00</u>	Volts / (μEinsteins/(cm ² ·s))
Wet Calibration Factor:	<u>8.77E-18</u>	Volts / (quanta/(cm ² ·s))	<u>5.28E+00</u>	Volts / (μEinsteins/(cm ² ·s))

Sensor Test Data and Results:

Sensor Supply Current (Dark): 3.9 mA
Sensor Supply Voltage: 6 Volts

Sensor Calibration Signal Voltage: 144.0 mV
Sensor Dark Voltage: 0.4 mV

Lamp Integrated PAR Irradiance: 9.27E+15 quanta/(cm²·s) 0.01539 μEinsteins/(cm²·s)
Immersion Coefficient: 0.5664

Note:

1. Annual calibration is recommended.
2. Calibration is performed using a Standard of Spectral Irradiance traceable to the National Institute of Standards and Technology (NIST).