Radboud University Nijmegen



PV Calibration Facility Nijmegen

Device Code:

GQN2736

Date:

25-Mar-19

Reference: Area:

GQN1964

4 cm2

Irradiance:

1000 W/m²

Spectrum:

AM1.5G


Temperature:

24.7 °C

Mismatch Corrected

Yes





I _{sc}	101.7 mA
V_{oc}	1020 mV
FF	82.1 %
Fta	21.3 %

 J_{sc} $I_{\rm mpp}$

 V_{mpp}

25.4 mA/cm2 96.1 mA 887 mV

Source: Load:

IEC 60904-9 compliant

Temperature:

Keithley 2400

Pt100

Software:

ReRa Tracer3

Contact calibration lab for uncertainty analysis


averkamp

RERA SOLUTIONS



Reference Cell (RR-1001/1002/ 1003/1004):

Wiring instructions




Open Cell Connections:


- 1: Positive Voltage Sense: Red Plug
- 2: Positive Current: Red Plug
- 3: Negative Voltage Sense: Black Plug
- 4: Negative Current: Black Plug

Shunted Cell Connections:

1+2: Positive Voltage: Red Plug

3+4: Negative Voltage: Black Plug





Use a sourcemeter to measure the Cell Output for open cells, use a voltmeter for shunted cells.

Use an RTD measurement device to measure the Pt100 Temperature.