



# Patient De-Identification Manual for GE Scanners

**Download PDF** 

MRA0252 August, 2021

#### Copyright © 2021 Perspectum Ltd

All rights reserved. No part of this publication may be reproduced, transmitted, transcribed, stored in a retrieval system or translated into any language or computer language, in any form or by any means, electronic, mechanical, magnetic, optical, chemical, manual or otherwise, without the prior written permission of Perspectum Ltd. All logos or products mentioned in this manual are the registered trademarks or copyrights of Perspectum Ltd, and are used for informational purposes only.

#### **Use and Limitation**

This Manual is intended to be used as a reference guide for the acquisition of images using at a GE 1.5T MRI system. Perspectum Ltd. shall not be held liable for improper installation, operation, or maintenance of the system in circumstances where procedures and specifications set out in this Manual have not been followed correctly. It is essential therefore, that you follow the specifications, procedures, and recommendations outlined in this Manual.


#### **Disclaimer**

Except as required by law, Perspectum Ltd. makes no representations or warranties, either expressed or implied, with respect to the contents hereof. Every effort has been made to ensure the reliability and accuracy of the information contained in this Manual at the time of going to print. However, specifications and procedures are subject to change due to our constant endeavours to meet customer requirements, and to maintain a research and development process of continuous improvement for all Perspectum Ltd. products. Consequently, Perspectum Ltd. reserves the right to make changes to the Manual at any time and without notice. You should always check with Perspectum Ltd. first to make sure that the version of the Manual you are referring to is the latest version. Unauthorised use means that the user assumes all liability for its use. To the maximum extent permitted by law, in no event shall Perspectum Ltd. be liable for damages or for, without limitation, loss of profits, business, revenue, loss of use, time, data, reputation, opportunity, goodwill, or savings, inconvenience, incidental, indirect or consequential damages, or damages arising from the use or failure to use the Manual strictly as prescribed herein.



## **Table of Contents**

1.	De-Identification Introduction	1
2.	Compatibility	2
3.	De-Identification - On Console	3
4.	De-Identification - Offline Mac	4
5.	De-Identification - Offline Windows	5
6.	Approvals	7





## 1. De-Identification Introduction

This document describes the correct procedure to de-identify patient identifiable information(PII) on GE MRI scanners on the console and offline.

Issue 4.1 Restricted Page 1 of 8



## 2. Compatibility

The procedures described in this document are compatible with MRI-scanners support by LiverMultiscan as described in System Requirements – MRA001.

Page 2 of 8 Issue 4.1 Restricted



## 3. De-Identification - On Console

Follow the steps described below to edit patient identifiable information.

- 1. Close the patient in all task and processes
- 2. Select the patient that is to be corrected from the Local DB section
- 3. At Data Apps Tab, click Edit Patient.
- 4. Correct the patient information to the appropriate identifiers as described in the Patient Registration section for the project as shown in




Figure 1. Patient de-identification on GE systems.

Issue 4.1 Restricted Page 3 of 8



## 4. De-Identification - Offline Mac

We recommend the use of open-source programs such as *Horos* or *OsiriX* to de-identify data.

- 1. Select the DICOM dataset you want to de-identify into Horos or OsiriX after loading into Document Data Base.
- 2. Click Anonymize as shown in Figure 2 (page 4) left
- 3. Correct the patient information to the appropriate identifiers as described in the Patient Registration section for the project as shwon in Figure 2 (page 4) right.
- 4. If needed to customise additional DICOM tags, click the **plus sign** on the **Select a DICOM tag...** option.





Figure 2. Select anonymize to correct the patient information (left) and modify additional DICOM tags if needed (right).

5. Ensure all the following DICOM tags are selected before exporting the dataset.



Page 4 of 8 Issue 4.1 Restricted



## 5. De-Identification - Offline Windows

We recommend the use of open source program such as DicomCleaner™ to de-identify data.

Follow the steps described below, to edit patient information.

1. **Import** the dataset to de-identify as shown in Figure 3 (page 5)




Figure 3. DicomCleaner de-identification for Windows system. Import data to de-identify

### 2. Check Replace box

In Patient's Name, Patient's ID and Accession Number fields, fill in the naming convention as described in the Patient Registration section of Image Acquisition Manuals

Check Remove all unreplaced identities, and Remove patient characteristics boxes

Press Clean button as shown in Figure 4 (page 6)

Issue 4.1 Restricted Page 5 of 8






Figure 4. Fill in information according to the agreed naming convention

3. Create a folder with the patient information as described in the Patient Registration section in your local hard drive, select the proper patient dataset to export, press **Export** button, select the folder created in your local hard drive.

Press Choose and data will be exported for the correct folder as shown in Figure 5 (page 6)




Figure 5. Export to a specific folder with de-identified data

Page 6 of 8 Issue 4.1 Restricted



# 6. Approvals

**Head of Imaging Solutions** 

**Head of Imaging Applications** 

Dr. Alex Smith

Andrea Borghetto

Issue 4.1 Restricted Page 7 of 8

