	§III	§VI-C	§IV	§V	§VI	§VII
Stage	Sensing & device wearable, CGM, home	Link & gateway BLE, Wi-Fi, cellular	Edge gateway or clinic	Cloud & platform multisite records	Synchronized twin one patient	Clinician decision human in the loop
Data leaving	raw streams, 1 Hz–2 kHz; events, video frames	windows, summaries; gaps and loss	features, windowed tensors	harmonized records, grids, imaging	versioned latent patient state	risk, forecast; rollouts (proposed)
Model families	int8 CNN, TCN, time-aware RNN	none required; on-device summarization	CNN–RNN, ConvLSTM, state space (Mamba)	STGNN, Transformer, 3D-CNN, GAN, DDPM	longitudinal Transformer, disease-progression model	world model and planner (proposed)
Resource limits	kB–MB memory, mW, battery	kb/s–Mb/s, duty cycle, interval	latency, memory, energy per inference	governance, harmonization	update cadence, privacy budget	safety filter, review time
Validation gate	subject-disjoint test; on-device latency and memory	loss visible to the model; bytes per patient-day	p95 latency on target; non-inferior to server model	external site held out unadapted; calibration	synchronization defined; per-patient calibration	identified effects; prospective study
Evidence in the 22 studies	11 device-captured; 3 device-class cost figures	0 report data rate or bandwidth	1 latency, desktop CPU; 0 on an edge board	8 second cohorts or sites, all adapted; 0 untouched	3 visit-cadence re-predictions; 0 sensor-synchronized	0 prospective; world models conceptual