

## Final review — D0.2

DRAFT FOR REVIEW | 2026-10-04

### Completed content

30 SOP PDFs and 68 editable WI documents have been expanded with applicable decision execution, exceptions and record requirements. Four physical workflow annexes reflect the official receiving, inspection, leakage and rejection forms. All 28 decision topics have substantive written proposals. Twenty editable blank continuation records support their execution.

### Verification

PASS: all 98 delivered SOP/WI bodies match their archived Markdown; all 98 source hashes match the master register. PASS: 84 original image hashes and original archive integrity. Prior D0.1 authoring sources are preserved inside the current source archive. The three-folder delivery structure is retained.

### Source reconciliation and limits

DEC-028 physical source availability is resolved for authoring: the owner confirms official physical use and all 84 original images have readable derivatives. Individual unreadable form-number suffixes, conflicting limits/units, equipment identity/manual, licensed clauses, applicability and real owner adoption remain identified in the decision book. The new policy text is complete as a review proposal; it is not a fabricated adopted decision.

The pressure-decay calculation example identifies an approximately tenfold difference between a simple dimensional calculation and one printer output. This requires instrument-method investigation; it does not establish that the instrument is defective. No real execution record, approval, effectivity or human UAT has been generated.