

Saudi Mais Co., for Medical Products
Work Instruction

Production & Test Procedure for Ureteric Catheter

Doc. #	:	WI-8-2-2-16
Rev. #	:	8
Date.#	:	19.11.2018
Page.#	:	2 / 4

2	Appearance	Radio opaque tube / Good	Visual
3	Free from Irregularities	No Damage / No Defected	Visual
4	Tip Forming of tube	Closed tip with smooth edge	Visual
5	Lateral Eyes status	Eyes should smooth edge	Visual / Physical
6	Connector status	No flashing, Black spot, short shot	Visual
7	Fitting connector + catheter	Good fitting	Visual / Physical
8	Connector Dimension	According to size	By Caliper
9	Dimension of tube (mm)	OD & ID as per specification	By Caliper
10	Dimension of Wire (mm)	OD as per specification	By caliper
11	Tube length	75 cm ~75 cm	By Scale
12	Wire length	75 cm ~75 cm	By Scale
13	Graduation scale	Printed clearly as Required	Visual
14	Logo	Without Logo	Visual

5.3 Visual Inspection.

- Ensure that the received material conform to as written in the purchase order.
- Inspect the packaging for any deterioration during transportation.
- There must be no visible contamination on the product.
- The tube surface must be smooth, no bumps, and no flashing at all.
- The length of tubing must be accurately in order to customer required.
- The graduation scale must be clear printed, accurately marking and not in washout.
- Red mark and Black mark graduation printed, not vividly.
- Length must be accurately 70 – 75 cm.
- Dimension must be accurately in order to quality specification.

5.4. Dimension Control and color requirement for the different sizes:

Size	O.D. ± 0.5 mm	Length (cm)	Graduation Mark
CH 03	0.37 mm	70 – 75 cm	Red color
CH 04	0.34 mm	70 – 75 cm	Red color

Prepared by Q.S:
Mr. Esmayaden

Approved by Q.C.M:
Mr. Javeed Gafoor

Controlled Copy

audi Mais Co., for Medical Products
Work Instruction

Production & Test Procedure for Ureteric Catheter

Doc. #	:	WI-8-2-2-16
Rev. #	:	8
Date.#	:	19.11.2018
Page.#	:	3 / 4

CH 05	0.34 mm	70 – 75 cm	Red color
CH 06	0.35 mm	70 – 75 cm	Red color
CH 07	0.36 mm	70 – 75 cm	Red color
CH 03	0.37 mm	70 – 75 cm	Black color
CH 04	0.34 mm	70 – 75 cm	Black color
CH 05	0.34 mm	70 – 75 cm	Black color
CH 06	0.35 mm	70 – 75 cm	Black color
CH 07	0.36 mm	70 – 75 cm	Black color

6. Packaging and Labeling:

- After issue the product from the store to production area, the QC and production controller must be check before go to the machine section.
- Each product package in blister pack, the label contains the following. Product name and code no., reference no, size, lot no., type, temperature, logo, sterilization logos, production date, expiry date, etc.
- After packaging on machine, the product should be checked by the repacking personnel, packaging should be intact & in good condition.
- Then to pack in specified small boxes, and arrange in big carton.
- The big carton must be with label and detail information.
- During packaging process, quality controller to be check the package sealing, peeling, sealing width, forming, cutting and printing.

7. Sterilization

The finished products are sterilized as per work instruction (WI-7-5-1-1)

8. Final Inspection and Product Release

- The Quality controller must be check, the unloaded products after sterilization.
- The quality assurance manager or quality control manager have to review the completed cycle parameters and inspection reports both documents have to be sign.
- After confirmation all the test from the Lab, to release the sterilized products, the document once again has to review the quality control manager then sign.
- The production department responsible to record the quantity of products before transferring to the finished goods ware-house and then it could be released to the customer.

Prepared by Q.S:
Mr. Esmayaden

Approved by Q.C.M:
Mr. Javeed Gafoor

Controlled Copy

audi Mais Co., for Medical Products
Work Instruction

Production & Test Procedure for Ureteric Catheter

Doc. #	:	WI-8-2-2-16
Rev. #	:	8
Date.#	:	19.11.2018
Page.#	:	4 / 4

9. Sampling:

- a. Inspection & Test report for Ureteric Catheter : F-823-T69
- b. In-process control for packaging machine & final checking : F-823-i4
- c. Final inspection and Release Record : F-8-2-4


10.Reference:

- a. BS EN-1615
- b. BS EN-1618


11.Modification / Documentation change History

Date	Changes
19.11.2018	Item 7, 8, new added

Prepared by Q.S:
Mr. Esmayaden



Approved by Q.C.M:
Mr. Javeed Gafoor



Controlled Copy