

Calibration Record

Training artifact — created for skills demonstration

Service Provider: Metro Metrology Services
Instrument: Digital Caliper 0–6 in
Asset ID: CAL-DC-006-1842
Calibration Date: 2026-02-01
Next Due Date: 2027-02-01
Procedure: MET-CAL-12 Rev B

Results

| Point (in) | Tolerance (in) | As-Found | As-Left | Status |
|------------|----------------|----------|---------|--------|
| 0.500 | ±0.0010 | 0.5004 | 0.5000 | PASS |
| 1.000 | ±0.0010 | 1.0008 | 1.0001 | PASS |
| 2.000 | ±0.0015 | 2.0009 | 2.0001 | PASS |
| 4.000 | ±0.0020 | 4.0011 | 4.0002 | PASS |
| 6.000 | ±0.0025 | 6.0018 | 6.0003 | PASS |

Standards traceability: NIST-traceable standards used during service.

Technician: A. Patel
Signature/Date: Signed electronically • 2026-02-01