

Characterising and correcting the underestimation of vessel diameter in optical coherence tomography angiography

YASH PATEL,¹ PRAJAKTA BELEKAR,¹ GEORG LADURNER,¹ SYBREN WORM,¹ LUCAS MAY,¹ MARIA VARAKA,¹ THERESA ROHRMOSER,¹ FERNANDO GARCÍA-RAMÍREZ,¹ RENÉ WERKMEISTER,¹ BERNHARD BAUMANN,^{1,2} AND CONRAD W. MERKLE^{1,*}

¹ Center for Medical Physics and Biomedical Engineering, Medical University of Vienna, Währinger Gürtel 18-20/4L, 1090 Vienna, Austria

² Institute of Biomedical Physics, Medical University of Innsbruck, Müllerstraße 44, 6020 Innsbruck, Austria

*conrad.merkle@meduniwien.ac.at

Abstract:

Vessel diameter is a widely used biomarker of vascular health, commonly assessed in the retina using optical coherence tomography angiography (OCTA). However, shear-induced alignment of red blood cells at vessel edges reduces backscattering and suppresses the OCTA signal. This leads to systematic underestimation of vessel diameter. In this study, we established ground-truth diameters in the in vivo mouse retina using Intralipid contrast enhancement, compared binary mask-based calculation against model-based profile fitting on maximum and mean intensity projections, and used flicker stimulation to evaluate the performance of each method. Before contrast, a generalised Gaussian fit to the maximum intensity projection retained the strongest linear relationship with ground truth of all methods tested, although it underestimated diameter by approximately 35%, making it correctable by calibration. Post-contrast, the same fit applied to the mean intensity projection agreed with ground truth ($m = 1.001$, $R^2 = 0.97$). Under flicker stimulation, model-based fitting detected the expected vasodilation, whereas binary mask-based calculation instead reported vasoconstriction. These findings characterise OCTA diameter underestimation and establish model-based fitting as a practical means of correcting it.

1. Introduction

Vessel diameter is an important biomarker that can provide valuable information regarding local vascular health. Thickening of the arteriolar wall has been shown to narrow the vessel lumen in proliferative diabetic retinopathy [1]. Retinal arteriolar narrowing has been shown to precede and predict the onset of systemic hypertension [2]. In glaucoma suspects, reduced vessel diameter has been shown to track both retinal ganglion cell dysfunction and visual field loss [3]. Diameter changes are also informative beyond the eye, where a meta-analysis of observational studies found smaller retinal arteriolar caliber to be associated with coronary heart disease [4]. Different diseases have been shown to preferentially affect different plexuses, with glaucoma altering the superficial vascular complex [5] while diabetic retinopathy [6] and retinal vein occlusion [7] affect the deeper plexuses more severely. Combined with measurements of blood flow, diameter can also be used to investigate local blood pressure and vessel elasticity [8–10]. For these reasons, it is imperative that accurate measurements of vessel diameter be used for these extrapolations or when modeling behaviours of flow in digital twins [11, 12]. Optical imaging techniques are often used to measure vessel diameter due to their high resolution, but accurate measurement of vessel size is more difficult than it first appears.

Optical coherence tomography (OCT) is a non-invasive optical imaging method, which uses interferometry to measure the transit time of light that is backscattered from a sample, thereby producing depth-resolved images of tissue [13, 14]. Optical coherence tomography angiography

maybe put in the context of intensity

when making...
sub-extrapolations

← not really true
which can lead to additional max + mean proj.
although this could be...
readers won't know what this means. Would mention the it is the most reliable measure in our evaluation.

and systemic
ank
is this for vessel diameter
8 - does not say this
Elasticity not mentioned

(OCTA) is an extension of OCT that visualizes vascular perfusion using the motion contrast from red blood cells (RBCs) [15–18]. Over the years, OCTA has demonstrated widespread acceptance and use for clinical evaluations of vascular features in relation to disease [19]. OCT angiograms can be used to evaluate characteristics on the scale of vascular layers by measuring vascular density or non-perfusion zones [5–7, 20–22], or it can be used to gauge vessel size [23] or tortuosity [24].

Analysis of OCTA datasets is typically achieved by performing a maximum intensity projection of a 3D vascular network to compress it into 2D before binarising the resulting angiogram using a global or local binarisation method. From this binarised angiogram, various vascular metrics can then be quantified. Importantly, there are many methods for performing this binarisation step and the choice of method can significantly affect the reported vascular metrics, with no clear consensus for a one size fits all approach [25–29]. Similarly, there are vascular enhancement algorithms to highlight structures that resemble vessels before quantification, but these methods also risk affecting the apparent size of the vessels in the network [28, 30, 31].

Furthermore, OCT is prone to certain artifacts which specifically affect visualization of the vasculature [32–34]. When viewed in cross-section, large vessels often demonstrate an hourglass appearance with stronger scattering in the center than at the sides. This appearance is due to orientation effects of RBCs, which align with the blood flow, causing a reduced scattering cross-section at the sides of the vessels [32]. We hypothesise that these artifacts may lead to underestimations of measurements of vessel diameter. To compensate for these artifacts, OCT contrast agents such as Intralipid have previously been used to increase intravascular intensity signals, particularly in regions where RBC scattering is diminished [33–37].

In this study we examine different methods of measuring vessel diameter from in vivo OCT scans of mouse retinal vasculature, and we use a contrast agent to obtain ground truth vessel diameter measurements for comparison. In addition to traditional binarisation-based methods, we also employ model-based methods for quantifying vessel diameter. Traditional Gaussian models have been employed for vessel diameter measurements in techniques such as fluorescein angiography [38] and fundus photography [39] (based on other optical imaging methods) and OCT [40], but we expand this by also investigating the generalized Gaussian model. As part of this study, we measure how much diameter is underestimated from enface projections and evaluate how reliably it can be compensated through calibration. Flicker stimulation is well established to evoke vasodilation in the retina [41–47]. After identifying the best methods for quantifying diameter and compensating for underestimations, we finally evaluated a test scenario using flicker stimulation to induce a measurable change in diameter to evaluate the performance of different methods.

2. Methods and Materials

2.1. Imaging system

All imaging was performed using a custom-built polarisation-sensitive spectral-domain OCT ophthalmoscope, described in detail previously [48]. The system was centred at 840 nm with a spectral bandwidth of 100 nm at full width half maximum (FWHM), providing an axial resolution of approximately 5.1 μm in air and 3.8 μm in tissue (assuming a refractive index of 1.35). A sensitivity of 96 dB was achieved at an A-scan rate of 80 kHz and a beam diameter of 0.5 mm at the pupil plane (power at cornea = 2.85 mW) [49]. This system was employed to acquire volumetric OCT data from mouse retinas before, during, and after an intravascular contrast agent administration, as well as before and after functional stimulation of the mouse retina using a custom 502 nm light-emitting diode (LED) ring flashing at 10 Hz (power at cornea = 26.6 μW). This wavelength was selected to target the peak absorption of mouse rhodopsin (~500 nm), enabling predominantly rod-driven photoreceptor stimulation [50].

2.2. Experimental protocol

Three mouse models were used in this study: VLDLR knockout mouse (B6;129S7-Vldlr^{tm1Her/J}, The Jackson Laboratory, Bar Harbor, USA), SOD1 non-transgenic mouse, and C57BL/6J mice (C57BL/6J, The Jackson Laboratory, Bar Harbor, USA) with $n = 1, 1, 2$ mice per VLDLR, SOD1 and C57BL/6J group respectively. All imaging sessions were conducted under general anaesthesia using either isoflurane, medetomidine + midazolam + fentanyl + ketamine (MMFK) combination, or a ketamine + xylazine combination. For the VLDLR knockout and the SOD1 non-transgenic mice, isoflurane was induced at 4% in oxygen for 4 minutes and subsequently maintained at 2% in oxygen at a flow rate of 1.5–2 L/min for the duration of the imaging session. In cases where isoflurane alone did not provide adequate sedation, an intraperitoneal injection of ketamine + xylazine cocktail (100 mg/kg ketamine, 6 mg/kg xylazine; 10 mL/kg body weight) was administered. For the C57BL/6J mice, isoflurane was induced at 4% in oxygen for 4 minutes and then MMFK (10 mL/kg body weight) was administered. To maintain core body temperature of the mice throughout imaging, they were positioned on a heating pad. Mouse pupils were dilated with one drop of tropicamide (5 mg/mL, Agepha Pharma s.r.o., Senec, Slovakia) per eye, and artificial tear drops (Oculotect, ALCON Pharma GmbH, Freiburg im Breisgau, Germany) were applied at regular intervals to prevent the cornea from drying out.

Intralipid 20% (3 mL/kg body weight) was used as a contrast agent and was administered as a bolus injection via the tail vein. Volumetric angiography data were acquired over a 1 mm × 1 mm field of view centred on the optic nerve head (ONH), with a scan density of 500 A-scans per B-scan and five repeated B-scans at each of 400 slow-axis positions (acquisition time approximately 16 seconds). For this study, pre- and post-injection volumetric datasets were acquired. This volumetric data was used solely for visualisation purposes. Additionally, a dynamic-contrast OCT (DyC-OCT) protocol [37] was employed during the contrast injection or during stimulation, consisting of 2000 repeated B-scans over approximately 16 seconds (interscan time 8.2 ms) at the same cross-sectional location over a 1 mm lateral field of view. DyC-OCT scans allow us to track the same vessels over time to precisely measure the effects of the contrast agent or the stimulation with a high degree of spatial coregistration. All quantitative vessel diameter measurements reported in this study were derived from the DyC-OCT data.

Functional flicker stimulation was performed in a C57BL/6 mouse to assess whether the choice of diameter measurement method affects the detection of physiological vascular responses. For this experiment, ketamine/xylazine cocktail was chosen over isoflurane to minimise anaesthesia-induced vasodilation [51, 52]. During the DyC-OCT acquisition, the retina was unstimulated for approximately 7 seconds, followed by 10 Hz flicker stimulation from the LED ring for the rest of the scan duration. To compare results before and after stimulation, the data was divided into baseline (pre-stim) and post-stim. Considering the time required for the OCT signal to stabilise after stimulation, four-second time windows at the beginning and end of each scan were selected for the pre- and post-stimulation analyses respectively.

All animal procedures were approved by the ethics committee of the Medical University of Vienna and the Austrian Federal Ministry of Education, Science, and Research (BMBWF/66.009/0272-V/3b/2019 and GZ 2024-0.802.524).

2.3. Post-processing and analysis

Cross-polarised OCT intensity data from the retinal pigment epithelium (RPE) was used for inter-frame alignment and B-scan flattening [49, 53]. Angiograms were derived by applying the complex subtraction OCT angiography (OCTA) algorithm [16]. We evaluated multiple combinations of methods for quantifying vessel diameter from the OCTA data to determine which combinations of steps work best under different conditions. Specifically, we looked at two types of data projections (maximum intensity and mean intensity) to generate vessel profiles and several different quantifications of those profiles including a traditional global threshold, a

reference a bit → combination of existing and new
→ follows 3 R's of animal research
→ mouse genetic model or anaesthesia
type should not impact this study

stim was k/x?

swap order

(approx...)

were

143 Gaussian model approach, and a Generalised-Gaussian (GG) model approach. The projection
 144 and global threshold steps follow established OCTA quantification practice [54–56], whereas the
 145 model-based approaches build on Gaussian profile fitting previously applied to retinal vessel
 146 diameter measurement in other optical imaging modalities and in OCT [38–40], which we extend
 147 to the GG model. The different methods were evaluated against the ground truth as described in
 148 this section. All data processing and analysis was performed in MATLAB (R2019b, MathWorks,
 149 Natick, MA, USA).

150 2.3.1. Ground truth calculation

151 Ground truth vessel diameters were obtained from contrast-enhanced DyC-OCT B-scan an-
 152 giograms. The lateral (d_x) and axial (d_z) diameters of the vessel were manually measured from
 153 the vessel's cross-section after arrival of contrast. The ground truth diameter was defined as
 154 the minimum of the two measurements, $d_{GT} = \min(d_x, d_z)$, to account for potential off-axis
 155 cross-sectioning of the vessel, which produces a more elliptical appearance in the B-scan. For a
 156 vessel perpendicular to the B-scan, the lateral and axial diameters are in agreement.

157 2.3.2. Vessel projection calculation

158 For each large vessel in the field of view, a rectangular region of interest (ROI) encompassing
 159 the vessel cross-section, along the lateral (x) and axial (z) dimensions, was selected from the
 160 DyC-OCT B-scan angiograms. En face angiogram profiles were then generated by projecting
 161 the angiogram signal within this ROI along the axial (z) direction for all time points. Two
 162 projection methods were compared: maximum intensity projection (MIP), which selects the
 163 highest intensity value along the depth axis of each ROI, and mean intensity projection (MeIP),
 164 which averages the values along each ROI. Each projection resulted in a time-stack of 1-D lateral
 165 intensity profiles for each vessel, which were then used for subsequent diameter quantification.

166 2.3.3. Skew correction

167 Off-axis cross-sectioning elongates the vessel along the lateral direction, so the lateral diameter
 168 measured from the ROI projection can overestimate the true vessel size. Because the lateral
 169 diameter is the quantity studied here, a correction factor of d_{GT}/d_x was defined from the ground
 170 truth diameter. ~~For an obliquely sectioned vessel, where the lateral diameter exceeds the axial~~
 171 ~~diameter, this evaluates to d_z/d_x , while for a vessel perpendicular to the B-scan it reduces to~~
 172 ~~unity.~~ By multiplying this correction factor against the diameter measured from each method, we
 173 ensured that the measured vessel diameter reflected the true size of the vessel. *without skew effects.*

174 2.3.4. Model-based vessel diameter calculation

175 For the model-based approaches, the vessel projection profiles were normalised to a range of 0–1
 176 to ensure a good fit. Two parametric models, the standard Gaussian and Generalised-Gaussian
 177 (GG) models, were fitted (Fig. 1a) to the one-dimensional intensity profiles using nonlinear
 178 least-squares regression in MATLAB.

179 Standard Gaussian function:

$$f(x) = A \exp\left(-\frac{(x - \mu)^2}{2\sigma^2}\right) + C \quad (1)$$

180 where A is the peak amplitude, μ is the centre position, σ is the standard deviation, and C is a
 181 constant baseline offset.

182 Generalised-Gaussian (GG) function:

$$f(x) = A \exp\left(-\left(\frac{|x - \mu|}{\alpha}\right)^\beta\right) + C \quad (2)$$

→ make consistent throughout

*and surrounding
tissue*

where A is the peak amplitude, μ is the centre position, α is the scale parameter, β is the shape parameter, and C is a constant baseline offset. Because the measured profile is the vessel's reflectivity profile convolved with the system point spread function, it cannot be sharper than the point spread function itself, and so profiles sharper than Gaussian profile ($\beta < 2$) were not considered here. The shape parameter β was hence constrained to $\beta \geq 2$ during fitting. With this constraint, the GG accommodated to profiles ranging from Gaussian ($\beta = 2$, where $\sigma = \alpha/\sqrt{2}$) to increasingly flat-topped ($\beta > 2$) (Fig. 1a).




Fig. 1. (a) GG function for shape parameters $\beta = 1, 2, 3$, and 4. (b) Example vessel intensity profile (black) taken from a MIP, with standard Gaussian (blue) and GG (red) fits. The MIP profile is flat-topped ($\beta = 9.9$), a shape the standard Gaussian cannot represent.

Two standard width metrics were calculated from the fitted models: the full width at half maximum (FWHM) and the $1/e^2$ width. For the standard Gaussian fit, the $1/e^2$ width was calculated as:

$$w_{1/e^2, G} = 4\sigma \quad (3)$$

The FWHM for a Gaussian is related to the $1/e^2$ width by a fixed proportionality:

$$w_{fwhm, G} = 2\sigma\sqrt{2\ln 2} = \frac{\sqrt{2\ln 2}}{2} w_{1/e^2, G} \quad (4)$$

and evaluated here

and therefore only the $1/e^2$ width was calculated directly from the fit, with the FWHM obtainable via this known relationship.

For the GG fit, no such linear relationship exists between the FWHM and $1/e^2$ width, since both depend on the shape parameter β . Both metrics were therefore calculated independently as:

$$w_{fwhm, GG} = 2\alpha(\ln 2)^{1/\beta} \quad (5)$$

$$w_{1/e^2, GG} = 2\alpha \cdot 2^{1/\beta} \quad (6)$$

2.3.5. Binary mask based vessel diameter calculation

For comparison with OCTA processing pipelines that rely on image binarisation, vessel diameters were also measured from binary masks generated by global thresholding. To determine the threshold level, the maximum angiogram intensity was measured after applying a 3D averaging filter with a kernel size of $50 \times 50 \times 10$ voxels (lateral $x \times$ axial $z \times$ temporal t) to the DyC-OCT angiogram volumes to reduce noise and minimize the effects of outliers. To produce a binarised black and white (BW) mask, the maximum angiogram intensity within these smoothed volumes was calculated, and five different threshold values were defined as fixed percentages of these maxima for MIP and MeIP. For MIP, the thresholds were defined as BW-1 = 40%, BW-2 = 50%,

208 BW-3 = 60%, BW-4 = 70%, and BW-5 = 80%. While for MeIP, the thresholds were defined as
209 BW-1 = 67%, BW-2 = 68%, BW-3 = 69%, BW-4 = 10%, and BW-5 = 11%. These thresholds were
210 selected empirically to cover the usable range of each projection. The upper and lower bounds
211 were selected based on the point at which vessels began to be lost, and at which background
212 noise and tissue signal were increasingly included, respectively. Each threshold was applied to
213 the unsmoothed en face projection (MIP and MeIP) to generate a binary mask, and the vessel
214 diameter was measured as the number of pixels above the threshold across the lateral vessel
215 cross-section.

216 2.3.6. Statistical analysis

217 To evaluate the accuracy and precision of all diameter measurement approaches, the calculated
218 diameter values were plotted against the ground truth for each vessel across three conditions:
219 pre-contrast, peak contrast, and post-contrast. A linear regression was fitted to each condition,
220 yielding the slope (m) and coefficient of determination (R^2). A slope of $m = 1$ indicated perfect
221 agreement with the ground truth, and R^2 quantified the goodness of fit of the linear model.
222 Additionally, the coefficient of variation (CoV) was computed for each vessel across the time
223 samples within each contrast condition and then averaged across vessels, to assess measurement
224 precision (repeatability). For the model-based methods, six combinations were evaluated: two
225 projection methods (MIP and MeIP) by three width metrics ($w_{1/e^2, G}$, $w_{fwhm, GG}$, and $w_{1/e^2, GG}$).
226 For the binary mask based method, diameter was measured at five threshold levels (BW-1 through
227 BW-5) for both MIP and MeIP. For the functional flicker stimulation experiment, a paired
228 t-test was used to evaluate whether vessel diameters differed significantly between the pre- and
229 post-stimulation windows, and the Pearson correlation coefficient (ρ) between measured diameter
230 values and angiogram intensity was computed for each method and vessel to assess whether the
231 diameter measurements were affected by changes in angiogram intensity. Measurements with
232 $p < 0.05$ were considered statistically significant (significance levels: * $p < 0.05$, ** $p < 0.01$,
233 *** $p < 0.001$).

234 3. Results

235 3.1. Qualitative assessment of vessel underestimation

236 Figure 2 shows OCTA MIP enface images of the superficial vascular plexus (SVP), intermediate
237 capillary plexus (ICP), and deep capillary plexus (DCP) before and after contrast injection. Across
238 all three layers, vessels appeared visibly wider and brighter post-contrast injection, consistent
239 with previous literature [49]. This effect was most pronounced in the larger vessels, especially in
240 the SVP, where the major retinal arteries and veins appeared noticeably thicker after contrast
241 administration. In the ICP and DCP, the capillary networks also showed improved visibility,
242 though the effect was less dramatic.

243 DyC OCTA projections helped visualise the temporal changes in vessel diameter pre- and
244 post-contrast injection (Fig. 3). Pre-contrast injection, the vessel cross-sections appeared narrower
245 than in the contrast-enhanced phase (Fig. 3a), which could be due to lateral edge signal suppression
246 by the RBC orientation artefact [34]. At peak contrast, the Intralipid filled the vascular lumen
247 completely. The vessels appeared slightly wider at peak than post-contrast, and we think this
248 could be due to the dynamic range of the image rather than a true diameter change (Fig. 3b). In
249 the post-contrast phase, the signal stabilised as the initial bolus cleared, and the vessel boundaries
250 remained more visible than before injection (Fig. 3c). The MIP and the corresponding binary
251 mask over time (Figs. 3d–e) showed the observed temporal changes in vessel diameter across the
252 three phases. The mean OCTA intensity (Fig. 3f) showed a sharp increase post-contrast injection,
253 followed by a decrease to a new steady state that remained elevated above the pre-contrast
254 baseline.

R^2 really linearity

is measured by the linear slope (m) and ...

give a brief intro to what m , R^2 , and CoV represent

3.2. Accuracy and precision of diameter measurements

The accuracy (m , R^2) and precision (CoV) of all diameter measurement methods are summarised in Table 1. All model-based methods underestimated vessel diameter in pre-contrast data, with slopes ranging from 0.42 to 0.73 (Table 1). Contrast enhancement improved accuracy across all metrics, though the degree of improvement varied.

include binary observations

The $w_{fwhm, GG}$ metric remained well below unity even post-contrast. The $w_{1/e^2, G}$ with MIP slightly overestimated at both peak and post-contrast ($m = 1.15$ and $m = 1.12$), while MeIP yielded values closer to unity ($m = 1.04$ in both conditions). The $w_{1/e^2, GG}$ with MeIP provided the best overall agreement, achieving a slope closest to unity with $m = 1.001$ ($R^2 = 0.97$) post-contrast, which was the closest to the identity line of all tested model-based methods. MeIP generally gave slopes closer to unity than MIP, but the linearity showed the opposite trend before contrast. Pre-contrast R^2 was higher for MIP across all three metrics, and MeIP only surpassed MIP post-contrast (Table 1, Fig. 4).

MeIP had smaller standard deviation

For binary mask based methods, the slope decreased with increasing threshold across all contrast conditions and projections (Fig. 5, Fig. 6) (Table 1). MeIP slopes were consistently higher than MIP at every threshold and contrast condition, though at lower thresholds and at peak and post-contrast, MeIP tended to slightly overestimate vessel diameter. The R^2 values showed opposite trends for the two projections, which are most directly compared in the summary bar plots of Fig. 7c,d. For MIP, R^2 decreased with increasing threshold across all contrast conditions, most strongly post-contrast (Fig. 7c). For MeIP, R^2 instead increased with threshold, most evident at peak contrast (Fig. 7d). Overall, MeIP provided more accurate binary mask based diameter measurements than MIP at most thresholds, and was more linear at the higher thresholds. The exception was the lowest threshold, where MIP stayed closer to the identity line at peak and post-contrast ($m = 0.96$ and $m = 0.92$, against $m = 1.14$ and $m = 1.11$ for MeIP) because MeIP overshoot unity, and was also more linear (Fig. 5a, Fig. 6a). MIP likewise provided higher R^2 at the two lowest thresholds (BW-1 and BW-2).

confusing, split up
sounds like MeIP? need to interpret a little

The precision of each method was assessed via the CoV (Table 1, Fig. 7). For MIP, the binary mask-based CoV increased substantially with threshold, whereas the model-based CoV remained stable across all contrast conditions. For MeIP, the binary mask based CoV was notably more stable across thresholds than for MIP, and the model-based CoV values were lower overall, though they increased slightly at peak contrast before decreasing post-contrast. Across both methods, MeIP yielded lower CoV values than MIP pre-contrast and more stable values post-contrast. The model-based metrics also maintained high linearity once contrast was present, with R^2 between 0.90 and 0.98 at peak and post-contrast for both projections, compared with 0.69–0.90 and 0.48–0.82 for the binary mask based methods over the same conditions (Fig. 7c,d).

make consistent

Among the model-based metrics, $w_{fwhm, GG}$ showed the largest underestimation in all three contrast conditions, with slopes between 0.42 and 0.66 across both projections. Its linearity was not correspondingly poor, however, post-contrast it reached $R^2 = 0.96$ for MIP and $R^2 = 0.98$ for MeIP, the highest and joint-highest of the model-based metrics respectively, and its precision remained comparable to the other model metrics. Overall, model-based methods outperformed binary mask based methods in both accuracy and precision, with two exceptions. At the two lowest MIP thresholds the binary mask based CoV was lower than that of any model-based metric (0.10 at BW-1 and BW-2 post-contrast, against 0.16–0.18), so MIP was more precise there but less linear ($R^2 = 0.81$ against 0.92–0.96). At the higher MeIP thresholds precision was comparable to the model-based methods, though accuracy was lower. MeIP yielded superior performance compared to MIP for most methods and thresholds, the main exception being the lowest binarisation thresholds.




Fig. 2. OCTA MIP en face of the SVP, ICP, and DCP before (left) and after (right) the administration of Intralipid 20%. Vessels appear visibly wider and brighter post-contrast across all layers, most prominently in the larger vessels. All scale bars are 100 μm .




Fig. 3. DyC-OCT B-scan angiograms at (a) pre-contrast, (b) peak contrast, and (c) post-contrast. (d) MIP and (e) binarisation of the vessel cross-sections over time. Coloured boxes indicate pre-contrast (blue), peak-contrast (green), and post-contrast (red) windows. (f) Mean OCTA intensity over time. All scale bars are 100 μm .




Fig. 4. Modelled versus ground truth vessel diameter using MIP (top row, a-c) and MeIP (bottom row, d-f). Columns correspond to $w_{1/e^2, G}$ (a, d), $w_{fwhm, GG}$ (b, e), and $w_{1/e^2, GG}$ (c, f). Linear regression lines and statistics are shown for pre-contrast (blue), peak contrast (green), and post-contrast (red) conditions. The black line indicates the identity line ($m = 1$).




Fig. 5. Vessel diameter measured from binarised masks versus ground truth for MIP at five threshold levels: (a) BW-1 through (e) BW-5. Linear regression lines and statistics are shown for pre-contrast (blue), peak contrast (green), and post-contrast (red) conditions.

Change BW
to BT
throughout?




Fig. 6. Vessel diameter measured from binarised masks versus ground truth for MeIP at five threshold levels: (a) BW-1 through (e) BW-5. Linear regression lines and statistics are shown for pre-contrast (blue), peak contrast (green), and post-contrast (red) conditions.

Table 1. Summary of slope (m), coefficient of determination (R^2), and coefficient of variation (CoV) for all diameter measurement methods across pre-contrast, peak contrast, and post-contrast conditions. Methods are grouped as binary mask based and model-based. Bold values indicate slope agreement within 10% of unity ($0.9 \leq m \leq 1.1$), strong linearity ($R^2 \geq 0.90$), or high precision (CoV ≤ 0.10).

	BW-1		BW-2		BW-3		BW-4		BW-5		$w_{1/e^2, G}$		$w_{fwhm, GG}$		$w_{1/e^2, GG}$	
	MIP	MeIP	MIP	MeIP	MIP	MeIP	MIP	MeIP	MIP	MeIP	MIP	MeIP	MIP	MeIP	MIP	MeIP
<i>Slope (m)</i>																
Pre	0.61	0.83	0.48	0.74	0.38	0.66	0.30	0.60	0.24	0.54	0.70	0.73	0.42	0.43	0.64	0.71
Peak	0.96	1.14	0.86	1.06	0.78	0.99	0.71	0.93	0.65	0.89	1.15	1.04	0.66	0.61	0.98	0.98
Post	0.92	1.11	0.82	1.03	0.73	0.96	0.63	0.91	0.54	0.86	1.12	1.04	0.66	0.61	1.02	1.00
<i>Coefficient of determination (R^2)</i>																
Pre	0.62	0.51	0.60	0.54	0.55	0.58	0.48	0.61	0.40	0.62	0.73	0.63	0.67	0.55	0.81	0.76
Peak	0.88	0.69	0.90	0.76	0.88	0.82	0.84	0.85	0.78	0.88	0.90	0.95	0.95	0.94	0.98	0.97
Post	0.81	0.75	0.76	0.79	0.68	0.81	0.58	0.82	0.48	0.82	0.92	0.98	0.96	0.98	0.95	0.97
<i>Coefficient of variation (CoV)</i>																
Pre	0.20	0.20	0.24	0.19	0.30	0.19	0.37	0.19	0.47	0.19	0.26	0.14	0.27	0.15	0.25	0.13
Peak	0.17	0.16	0.20	0.17	0.22	0.17	0.27	0.17	0.32	0.17	0.24	0.15	0.23	0.16	0.21	0.14
Post	0.10	0.13	0.10	0.12	0.12	0.11	0.15	0.10	0.18	0.09	0.18	0.10	0.18	0.11	0.16	0.09




Fig. 7. Summary of accuracy and precision across all diameter measurement methods. (a, b) Coefficient of variation (CoV) and (c, d) coefficient of determination (R^2) for MIP (a, c) and MeIP (b, d), respectively. Bars are grouped by method (BW-1 through BW-5, G-e², GG-HM, GG-e²) and colour-coded by contrast conditions.

3.3. Functional flicker stimulation

To assess whether angiogram intensity itself biases the diameter measurements, we examined the intensity time series alongside the diameter time series. The model-based diameter time series (Figs. 8a,b) showed a significant increase in diameter post-stimulation for most of the vessels. The binary mask based diameter time series (BW-5; Fig. 8c) showed a mixed trend, where most vessels showed a decrease in diameter post-stimulation. Bar plots of the mean diameter pre- and post-stimulation (Figs. 8d-f) confirmed this pattern. For the Gaussian model, four of five vessels showed significant diameter increases ($p < 0.01$), and the GG model showed significant increases in four of five vessels (Table 2). In contrast, the binary mask based diameter showed significant decreases in four of five vessels.

The binary mask based method showed uniformly positive per-vessel correlations, with an overall correlation of $\rho = 0.526$ ($p < 0.0005$; Fig. 9c). The Gaussian and GG models yielded overall correlations of $\rho = -0.101$ and $\rho = -0.098$, respectively (Fig. 9a,b). The angiogram intensity time series (Fig. 8g) showed a visible decline in angiogram intensity throughout the duration of the DyC-OCT scan across all five vessels. The mean intensity comparison pre- and post-stimulation (Fig. 8h) confirmed that the decrease was significant for all vessels, ranging from -13.6% to -19.8% (Table 2).

Based on the previous analysis we chose these variants...

Should we use MIP for BW-1?

and looked at intensity correlations

need to explain which 3 which

We want to see how the models perform in a controlled induction of diameter change

GG agreed closest to ground truth




Fig. 8. Functional flicker stimulation results. (a-c) Vessel diameter time series pre- and post-stimulation across five vessels. (d-f) Mean diameter per vessel before (dark bars) and after (light bars) stimulation for each method. Veins are shown in blue/cyan and arteries in dark/light red. (g) Angiogram intensity time series. (h) mean angiogram intensity per vessel pre- and post-stimulation. All measurements used MeIP.




Fig. 9. Measured vessel diameter as a function of angiogram signal intensity during functional flicker stimulation for (a) Gaussian model, (b) Generalised Gaussian model, and (c) binary mask based method. Each colour represents one of five vessels. ρ denotes the Pearson correlation coefficient.

Table 2. Flicker stimulation results. Δd : diameter change (%); ρ : Correlation between diameter and angiogram intensity; ΔI : intensity change (%). Statistically significant values are bold.

	Gaussian		GG		Binarised		Intensity
	Δd (%)	ρ	Δd (%)	ρ	Δd (%)	ρ	ΔI (%)
Vessel 1 (Artery)	+8.2***	-0.07	+3.5**	0.03	-12.5***	0.59***	-13.6***
Vessel 2	+9.9**	0.23*	+10.1***	0.07	-10.2***	0.80***	-18.2***
Vessel 3	+1.2	-0.28**	-0.4	-0.28**	-4.1**	0.27**	-14.3***
Vessel 4	+28.5***	-0.15	+28.4***	-0.14	+1.0	0.46***	-19.8***
Vessel 5	+4.9***	-0.24*	+3.1**	-0.18	-9.3***	0.52***	-16.3***
Overall ρ		-0.101		-0.098		0.526	
Overall arterial change	+15.5		+14.0		-7.2		-17.2
Overall venous change	+3.1		+1.4		-6.7		-15.3

Ground Truth
↓

4. Discussion

4.1. Model-based vs binary mask based diameter measurement

The model-based and binary mask based diameter methods define vessel boundary in fundamentally different ways. Binarisation of the en face maps discards the spatial intensity distribution for a simple binarisation and reduces each pixel to a binary decision. The measured diameter directly depends on the chosen threshold value and on angiogram intensity, both of which may vary across imaging sessions, contrast conditions, and even between vessels within the same scan. Model-based methods fit a parametric model function to the maximum or mean intensity profiles, and calculate the diameter from the fitted parameters. As the fit captures the shape of the profile rather than the absolute intensity, it is inherently less sensitive to changes in angiogram signal and does not require a threshold at all. $\frac{1}{2}$ same can be used for all datasets.

A practical consequence of this difference is how correctable each method is. Because the model-based fit requires no threshold, it returns a single consistent value for a given profile, whereas the binarisation slope shifts with the chosen threshold, so any calibration must also encode which threshold was used. Underestimation can be corrected by a fixed proportion using the slope, provided the relationship with the ground truth is strongly linear, but such a rescaling removes only the systematic offset and leaves the scatter about the regression unchanged. Linearity, rather than slope alone, therefore determines how reliably a method can be corrected. Post-contrast every model-based metric reached $R^2 \geq 0.92$, whereas binary mask based MIP fell from 0.81 at BW-1 to 0.48 at BW-5, so more than half the variance would remain after any correction. This is what makes the higher-linearity model-based metrics usable even pre-contrast, where $w_{1/e^2, GG}$ retained $R^2 = 0.81$ against 0.62 for the best binary mask based method.

This difference explains the trends observed in Fig. 7 and Table 1. Post-contrast, every model-based metric was more linear than every binary mask based method, and the GG $1/e^2$ metric reached near-unity slopes at peak and post-contrast without a threshold parameter to tune, whereas for the binary mask based methods linearity was overall lower and depended on threshold. Because the peak angiogram intensity varies between vessels, a fixed threshold removes a different fraction of the cross-section from each one. The measured diameters therefore vary from the ground truth by different amounts, which weakens the linearity. Pre-contrast, the hourglass artefact caused by RBC orientation within vessel lumen suppresses edge signals, causing all methods to underestimate diameter [34]. However, model-based methods still retained more of that linearity, producing systematic, predictable underestimation. Binary mask based

MeIP?

encapsulate

methods on the other hand showed weaker linearity in most conditions, potentially because the suppressed edge signals did not consistently exceed the binarisation threshold, and because different vessels have different intensity.

The same mechanism accounts for how each method responded to the change in angiogram intensity from peak to post-contrast. Angiogram intensity is highest as the bolus arrives and then settles to a lower steady state (Fig. 3f), so a fixed threshold encloses a smaller fraction of each cross-section post-contrast than at peak, an effect visible in the binarised cross-sections over time (Fig. 3d,e). Model-based metrics were largely insensitive to this, changing by at most 0.04 in slope and 0.04 in R^2 between the two conditions. Binary mask based measurements from MIP degraded progressively with threshold, the slope falling by 0.04 at BW-1 but 0.11 at BW-5, and R^2 by 0.07 and 0.30 respectively, while MeIP remained comparatively stable (slope change 0.02–0.03 at every threshold). Model-based methods therefore showed comparable agreement with the ground truth at both peak and post-contrast, whereas the agreement of binary mask based method depended on the combination of threshold and contrast condition.

The generalised Gaussian (GG) model will always outperform the standard Gaussian because the intensity profiles of the vessels are usually not purely Gaussian (Fig. 1b). The standard Gaussian cannot capture relatively flat-topped angiogram intensity profiles and tends to fit a narrower curve, whereas the GG model accommodates through the shape parameter β (Fig. 1b). The flat-topped shapes arise because the measured profile is effectively the vessel's reflectivity profile convolved with the system point spread function. Because the GG reduces to a standard Gaussian when $\beta = 2$ and departs from it only when the data support a flatter profile, a correctly applied GG fit should perform at least as well as the Gaussian, at the cost of additional computation.

Among the width metrics, $w_{fwhm, GG}$ consistently showed the weakest accuracy, although its linearity remained high post-contrast ($R^2 = 0.96$ for MIP and 0.98 for MeIP), indicating that the underestimation is systematic and therefore correctable. Near the half-maximum the profile shape changes rapidly and varies between vessels, from Gaussian to flat-topped, so the width measured at that level is more variable. The $1/e^2$ width captures more of the transition region and so may have provided a more accurate measurement of the vessel diameter.

4.2. Maximum vs mean intensity projection

Across nearly all methods and conditions, MeIP outperformed MIP in both accuracy and precision (Table 1). Linearity was the exception. Pre-contrast MIP produced higher R^2 than MeIP for all model-based metrics, and MIP also produced higher R^2 at the two lowest binarisation thresholds, with MeIP outperforming MIP only post-contrast. This finding is noteworthy because MIP has been shown to produce higher signal-to-noise ratio (SNR) and contrast in en face OCTA images [55], and is the default in many commercial OCTA systems and analysis pipelines [54,56]. However, we observed that higher SNR did not necessarily translate to more accurate diameter measurements. The MIP selects the single highest intensity value along each A-scan within the vessel ROI, which makes it inherently sensitive to outliers. MeIP averages all intensity values along each A-scan, which smooths out intensity fluctuations and provides a consistent profile. This averaging could be beneficial for vessel diameter measurement, as the goal is to capture the typical vessel boundary rather than extreme signal events. This trade-off may also explain why linearity behaved differently before and after contrast. Pre-contrast the vessel signal is weakest, and edge suppression by the RBC orientation artefact leaves little margin above the background, so the higher contrast of the MIP helps keep the profile distinguishable across vessels of different size. Once the lumen is filled with contrast agent, signal is no longer limiting and the outlier sensitivity of a maximum statistic becomes the dominant source of error, so the averaging of the MeIP yields the tighter relationship with the ground truth. Consistent with this, MeIP produced higher R^2 than MIP for all three model-based metrics post-contrast, whereas at peak contrast the

two projections remained comparable, with MIP higher for two of the three model-based metrics. The difference between MIP and MeIP was especially apparent in the binary mask based results (Table 1). For MeIP, the slope declined more gradually and remained at 0.86 even at BW-5 post-contrast. Whereas for MIP, the slope decreased steeply with increasing threshold, dropping from 0.92 at BW-1 to 0.54 at BW-5 post-contrast. These findings suggest that MeIP produces a more uniform intensity distribution across the vessel profile, making the binarisation result less sensitive to the exact threshold value. The CoV data (Table 1) reinforced this as the MIP-based binary mask CoV increased substantially with threshold, whereas the MeIP-based CoV remained relatively stable. This behaviour follows from how each projection responds to variance in the angiogram signal. Because the MIP keeps only the brightest pixel along each A-scan, it tracks whichever value happens to be largest, so the projected profile fluctuates from one time sample to the next. A threshold applied to a fluctuating profile returns a fluctuating width. How much the width fluctuates depends on where the threshold cuts the profile. A low threshold cuts the steep flanks of the profile, where a change in intensity moves the measured edge only slightly. A high threshold cuts near the flat top, where the same change moves it much further. The CoV therefore grows as the threshold is raised. MeIP averages along each A-scan instead, which suppresses the fluctuation and gives a broader, smoother profile, so its CoV changes little as the threshold increases. For model-based methods, MeIP also showed better performance. The $w_{1/e^2, G}$ using MIP slightly overestimated vessel diameter post-contrast with $m = 1.12$, while MeIP reduced this to $m = 1.04$, a much closer agreement with the ground truth. These results suggest that while MIP may be preferred for visualisation purposes where contrast and SNR are important [55], MeIP should be the preferred projection method for quantitative diameter measurements.

4.3. Influence of angiogram intensity on diameter measurements

The functional flicker stimulation experiment provided a direct demonstration of how changes in angiogram signal intensity can affect diameter measurements based on the method used. During the DyC-OCT acquisition, the angiogram intensity declined significantly across all five vessels (Fig. 8g,h), with decreases ranging from -13.6% to -19.8% (Table 2). The cause of this decline was not established with certainty, although inspection of the volumetric data ruled out the most important confounds. There was no out-of-plane motion or breathing artefact, and the image quality was consistent throughout the scan. We suspect a combination of rapid corneal drying and/or cataract formation over the course of the scan could have caused the decline in intensity. More importantly, the cause of the decline is not essential to our argument for this study, what matters is that the intensity decline biases the binary mask based diameter measurement while the model-based diameter remains unaffected.

The model-based methods correctly detected the expected vasodilation, with the Gaussian model showing significant diameter increases in four of five vessels (+15.5% overall in the arteries against +3.1% in the veins) and the GG model showing significant increases in four of five vessels (+14.0% and +1.4% respectively; Table 2). The binary mask based method, however, reported significant diameter decreases in four of five vessels, suggesting vasoconstriction which is a paradoxical result. The reason behind this behaviour could be understood by observing how each method responds to declining signal intensity (Fig. 9). For the binary mask based method, the threshold is fixed and diameter is determined by the number of pixels that exceed this threshold. When the overall angiogram intensity drops, fewer pixels at the vessel edges exceed the threshold, and the vessel appears to shrink even if the physical vessel diameter has increased. This is confirmed by the strong positive correlation between measured diameter and angiogram intensity for the binary mask based method ($\rho = 0.526$, $p < 0.0005$). In contrast, the model-based methods showed weak overall correlations with intensity ($\rho = -0.101$ for Gaussian, $\rho = -0.098$ for GG), indicating that the fitted diameter was effectively independent of the signal intensity. Among the methods, the GG model also produced the lowest standard deviations across

probably not
does not affect
our observation

the five vessels in the flicker experiment (Fig. 8d–f), consistent with its low CoV post-contrast (Table 1), indicating that it is the most repeatable of the metrics tested. Together these results demonstrated that the model-based diameter calculation provided more accurate and repeatable measurements, and recovered physiologically meaningful vascular responses that binary mask based methods misrepresented. *identified*

and closest to ground truth

These findings have important implications for longitudinal studies and any experiment where angiogram intensity may vary between or within imaging sessions. In longitudinal mouse eye imaging, the angiogram signal can vary between and within sessions as anaesthesia alters retinal perfusion and ocular physiology [52, 57], with additional factors such as cataract formation, and corneal drying if tear drops are not applied, further reducing OCT intensity over time [58, 59]. If a binary mask based method is used under such conditions, observed changes in vessel diameter may reflect intensity fluctuations rather than true vascular changes. In contrast, model-based methods may be less sensitive to these intensity variations, as they characterise the shape of the intensity profile rather than relying primarily on absolute intensity values to define vessel boundaries.

4.4. Threshold selection

The results presented here used a basic, global threshold for the binarisation of the angiogram data. There are numerous other approaches for binarising the angiogram, which will most likely each behave differently. For example, angiogram signals can be enhanced with vesselness filters [60] and dynamic thresholds can be used to binarise images with varying SNR. When applying a vesselness filter, it is possible to enhance the shape of the angiogram and omit noise, but it is also possible for such enhancement to overcompensate the signal, leading to larger than expected vessel appearance. Measurements obtained with different enhancement filters have been shown to differ substantially and are not interchangeable [31]. Dynamic thresholding can also yield a more uniform angiogram appearance, but similarly risks over- or under-compensating vessels. The threshold chosen is known to alter OCTA quantitative metrics systematically, with automatic thresholds tending to underestimate relative to a fixed threshold [25, 29]. Further analysis of these other methods would be required to determine their efficacy in returning accurate and precise diameter values.

include ref for intro

The key observation from the present work is that any thresholding approach, regardless of how the threshold is determined, shares the same fundamental limitation that the measured diameter depends on the absolute signal intensity at the vessel edges. As demonstrated by the flicker stimulation experiment, this intensity dependence can introduce artefactual diameter changes when the angiogram signal varies over time or between sessions. A more adaptive thresholding strategy may reduce variability across different regions of a single image, but it would not eliminate the sensitivity to temporal intensity fluctuations that we observed. The model-based approach avoids this limitation entirely by deriving diameter from the shape of the intensity profile rather than from an intensity cut-off, making it a fundamentally different and more robust alternative for quantitative diameter analysis.

maybe cut

4.5. Best choices under different conditions

Under the conditions where a contrast agent is available and accuracy is the primary goal, the $w_{1/e^2, GG}$ metric using MeIP provides the best results post-contrast. This combination achieved a slope of $m = 1.001$ with $R^2 = 0.97$ and a CoV of 0.09, yielding the best accuracy, linearity, and precision of all tested methods. If only pre-contrast data is available, the $w_{1/e^2, GG}$ metric still provides the best linearity among all tested methods ($R^2 = 0.81$ for MIP, $R^2 = 0.76$ for MeIP), though the MIP and MeIP slope of $m = 0.64$ and $m = 0.71$ indicates that diameter could be underestimated by approximately 35% and 30% respectively. In this case, a correction factor could be applied if absolute diameter values are needed, though relative comparisons between

497 vessels or timepoints would remain valid without correction.

498 When only MIP data is available, which is the case for many commercial OCTA machines
499 and existing datasets [54, 56], the $w_{1/e^2, GG}$ still provides good post-contrast accuracy ($m = 1.02$,
500 $R^2 = 0.95$). For binary mask based analysis of MIP data, only the lowest threshold (BW-1)
501 post-contrast achieved a slope near unity ($m = 0.92$), and the slope degraded rapidly with
502 increasing threshold values. This highlights that if binarisation must be used with MIP data, a
503 lower threshold value is essential, though this may come at the cost of increased noise sensitivity.

504 The improved linearity and repeatability of the model-based approaches come at the cost
505 of computation time, as fitting a model is considerably more expensive than binarisation. In
506 our testing, fitting the generalised Gaussian was approximately 40% slower than the standard
507 Gaussian. This difference is hardware dependent, and because the fits at each cross-section are
508 independent, the computation could be parallelised and accelerated with appropriate hardware.
509 If computation time is a limiting factor and model fitting is not practical, MeIP with a moderate
510 threshold (e.g. BW-3) provides a reasonable alternative. At post-contrast, BW-3 using MeIP
511 achieved $m = 0.96$ and $R^2 = 0.81$, which is acceptable for many applications. MeIP-based
512 binary mask based analysis showed more stable CoV values across all thresholds compared to
513 MIP, making the results less sensitive to the exact threshold choice. This stability makes MeIP
514 the preferred projection for binary mask based OCTA.

515 4.6. Literature comparison

516 Hormel et al. [55] showed that MIP produces en face OCTA images with higher SNR and contrast
517 than MeIP, and concluded that MIP should be preferred for OCTA visualisation. Our results
518 reveal a different picture for quantitative vessel diameter measurement. We found that MIP
519 provided better linearity before contrast injection, whereas MeIP performed better after contrast
520 injection (Table 1). These findings suggest that the projection method best suited for visualisation
521 is not necessarily optimal for quantitative vessel diameter measurement, and that the choice of
522 projection should depend on both the contrast condition and the intended application.

523 The challenge of standardising OCTA image analysis has been highlighted by Sampson et al. [56]
524 and Untracht et al. [54], both of whom emphasised the need for transparent, reproducible analysis
525 pipelines. Our findings add to this discussion by showing that the choice of diameter measurement
526 method introduces a source of variability that is at least as significant as the choice of binarisation
527 threshold. The model-based approach presented here offers a path toward more standardised
528 vessel diameter quantification, as it reduces the dependence on selected thresholds and angiogram
529 intensity, and provides measurements that are more robust across imaging conditions.

530 The RBC orientation artefact and its effect on OCTA vessel appearance has been described by
531 Cimalla et al. [32] and Bernucci et al. [34], who used Intralipid contrast to demonstrate that the
532 hourglass pattern in vessel cross-sections is caused by the directional scattering properties of
533 RBCs rather than by true variations in flow. Our work extends this observation by quantifying
534 the degree to which this artefact affects diameter measurements and by proposing model-based
535 fitting as a practical correction strategy. While contrast enhancement with Intralipid is effective
536 for revealing the true vessel diameter, it requires a contrast injection and is not yet applicable in
537 clinical settings. Using a calibrated model-based approach provides a way to partially recover
538 the underestimated diameter from native OCTA data without the need for a contrast agent.

539 In our study, the magnitude of vasodilation detected by the generalised Gaussian model ranged
540 from +3.1% to +28.4% across the vessels with a significant response (Table 2). The arteries
541 (Vessels 1, 2, and 4) showed increases of +3.5%, +10.1%, and +28.4%, while the veins (Vessels
542 3 and 5) showed a smaller response, with +3.1% in Vessel 5 and no significant change in Vessel
543 3. The standard Gaussian gave a broadly similar picture, with arterial increases of +8.2%,
544 +9.9%, and +28.5% and a venous increase of +4.9% in Vessel 5. Table 3 summarises the
545 stimulation protocols and vascular responses reported by previous flicker stimulation studies

at the cost of
repeatability
(worse CoV)
MeIP arguably still
better pre-contrast
using GS $\frac{1}{e^2}$ method

546 alongside our own. The listed human studies span several measurement techniques, from
547 fundus photography [61] through retinal vessel analysers [41, 62, 63] to Doppler OCT [44] and
548 OCTA [64]. Despite these methodological differences, the reported human arterial responses
549 cluster between +3% and +8%, and venous responses between +2% and +10%. Notably, these
550 studies report arterial and venous responses of broadly comparable magnitude, and in some cases
551 the venous response is the larger of the two, for example +9.8% venous versus +8.0% arterial in
552 Aschinger et al. [44]. This contrasts with our measurements, in which the arterial response was
553 on average larger than the venous response. Whether this asymmetry reflects a genuine feature
554 of the murine flicker response, the small number of vessels sampled, or a greater sensitivity of
555 the model-based fit to the smooth-muscle-driven dilation of arteries remains to be established.
556 Polak et al. [62] additionally showed that the diameter response is relatively stable across flicker
557 frequencies from 4 to 40 Hz, suggesting that frequency differences between studies are unlikely
558 to explain the variation.

559 Among the studies listed, Son et al. [65] used the most similar setup to ours, with the same
560 mouse strain (C57BL/6J), the same anaesthetic (ketamine/xylazine), and a comparable flicker
561 stimulus (10 Hz, 504 nm), though a different stimulation period of 5 s. Using functional OCT,
562 they reported that the response is anisotropic, with larger percent changes in the axial direction
563 than in the lateral direction. Because the diameters reported here are lateral, their lateral values
564 offer the closest comparison: +10.8% in the arteries and +8.9% in the veins, against +14.0%
565 and +1.4% respectively for the GG model in our study. ~~The arterial responses are therefore of~~
566 ~~comparable magnitude, whereas our venous response is substantially smaller.~~ The authors did
567 not describe how vessel diameters were quantified from their B-scans, so it is worth considering
568 whether their lateral measurements may have been affected by the intensity-dependent artefacts
569 characterised here. In the lateral direction, the RBC orientation artefact suppresses signal at
570 vessel edges, and depending on how the vessel boundary was defined, this could lead to an
571 underestimation of the lateral diameter. If this were the case, the apparent difference between
572 axial and lateral vasodilation could be influenced by the measurement methodology in addition
573 to the biomechanical anisotropy discussed by the authors. It would be interesting to apply the
574 model-based approach presented here to such data to determine whether their reported anisotropy
575 changes when the lateral diameter is quantified differently. An earlier OCT-based study by
576 Radhakrishnan and Srinivasan [66] measured arterial dilation in the rat inner retina during a 10 s,
577 12 Hz light stimulus, reporting a 95% confidence interval of +3.3% to +11.0% ($p < 0.005$) for
578 diameters quantified manually along the arterial tree. ~~Our arterial response (+14.0% for the GG~~
579 ~~model) sits just above the upper limit of that interval, which is a reasonable correspondence given~~
580 ~~the difference in species.~~

581 The magnitudes we observed are generally larger than those reported in the human studies,
582 though direct comparison is difficult given the differences in experimental conditions. Species,
583 anaesthesia, stimulation frequency and wavelength, stimulus power and delivery method, stimula-
584 tion duration, dark adaptation period, and imaging modality all vary across the reported studies,
585 and each of these factors could influence the measured vascular response. Previous studies
586 in mice have measured flicker-evoked retinal vessel diameter changes mostly using dynamic
587 vessel analysis (DVA), which is a fundus camera-based technique adapted from human retinal
588 vessel analysers. Albanna et al. [67] demonstrated the feasibility of DVA in the mouse eye using
589 12.5 Hz flicker at 530 nm but found that quantitative arterial recordings were possible only in a
590 minority of animals, while venous dilation was mostly below 1.5% post-stimulation. Another
591 study by Neumaier et al. [68] used the same DVA setup in wildtype and Cav2.3-deficient mice
592 with overnight dark adaptation and reported median arterial and venous diameter changes of only
593 +1.3% and +1.2%, respectively, in the wildtype group. These values are substantially smaller
594 than those we observed, which could be due to the limited spatial resolution and signal-to-noise
595 ratio of DVA when applied to the smaller mouse eye rather than being a genuine physiological

Speculation
move up?
OCT?

cont

response. The much larger responses we observed may therefore partly result from the higher lateral resolution of OCT combined with the model-based fitting approach, which captures vessel boundary shifts that could fall below the detection limit of DVA. In addition to these diameter-based studies, Rai et al. [69] recently reported that 12 Hz flicker stimulation significantly increased both arterial and venous retinal blood flow in light-adapted C57BL/6J mice.

Importantly, regardless of the absolute magnitude, the fact that we were able to detect vasodilation at all depended on the measurement method, because binary mask based analysis not only failed to detect vasodilation but reported vasoconstriction, which would have led to fundamentally incorrect physiological conclusions.

Table 3. Comparison of flicker stimulation protocols and reported diameter changes across studies. H = human, M = mouse, R = rat. DA = dark adaptation. Lat. = lateral, ax. = axial.

Study	Stimulation Freq. (Hz)	Stimulation Light Source	Stimulation Duration (s)	DA (h)	Diameter Change (%)	
					Artery	Vein
Formaz [61] (H)	10	Xenon arc, 5.4 log Td [‡]	60	—	+4.2	+2.7
Polak [62] (H)	8	< 550 nm, ~ 150 μ W, cm ^{-2†}	60	—	+3.5	+1.9
Garhofer [41] (H)	8	< 550 nm, 300 μ W, cm ^{-2†}	60	—	+3.8	+2.8
Nagel [63] (H)	12.5	530 – 600 nm, ~ 196 μ W, cm ^{-2†}	20	—	+6.9	+6.5
Aschinger [44] (H)	12.5	530 – 600 nm, ~ 196 μ W, cm ^{-2†}	120	—	+8.0	+9.8
Kallab [64] (H)	7	White, ~ 450 lux	*	—	+3.7	+4.5
Albanna [67] (M)	12.5	530 nm, 30 lux	20	12	— [§]	< 1.5 [§]
Neumaier [68] (M)	12.5	530 nm, 30 lux	20	12	+1.3	+1.2
Radhakrishnan [66] (R)	12	Hg:Xe, 2 mW, cm ^{-2†}	10	—	+3.3 to +11.0	—
Son [65] (M)	10	504 nm, ~ 460 lux	5	1 – 2	lat. +10.8, ax. +18.2	lat. +8.9, ax. +14.3
Present study (M)	10	502 nm, 26.6 μ W [‡]	Cont.	2	+3.5 to +28.4 [†]	+3.1 [†]

[‡]Diameter measured from fundus photographs taken post-stimulation; retinal luminance in trolands.

^{*}DVA feasibility study; arterial recordings were not reliably obtainable in the mouse eye.

[†]Values from individual vessels with a statistically significant response; other studies report group means.

[§]95% confidence interval, not a range of per-vessel means. Diameter was quantified along the arterial tree only, so no venous value is reported.

[‡]Flicker applied during OCTA scan acquisition. [†]Retinal irradiance. [‡]Power at cornea.

how long is it? average all arteries and veins

4.7. Future direction

A future direction of this work is to move toward fully automated segmentation and calculation of vessel diameters from en face angiogram maps. While ROIs were manually selected in this work, it is also possible to automatically segment angiographic data using traditional skeletonisation methods, which are commonly used for automated OCTA analysis. From a skeletonised image, vessel angle can be computed for further extraction of vascular cross-sections perpendicular to the vessel axis. These cross-sections could then be processed using the model-based methods described here to enable fully automated analysis of 3D OCTA data sets, rather than the semi-automated approach shown here for 2D OCTA data. This type of model-based approach is particularly important for comparing 3D volumes acquired at different time points in the mouse eye, as there are often intensity losses caused by drying of the eye, cataract formation, etc. As we have shown in this work, even low levels of intensity loss occurring over a short time window can significantly affect the measured vessel diameter when using a binary mask based approach, but not when using the model based approach presented here.

maybe cite

Additionally, this work focused on the larger vessels of the superficial vascular plexus. Extending the model-based approach to the intermediate and deep capillary plexuses, where vessels are smaller and signal levels are lower, would be valuable for studying microvascular changes in diseases. For very small vessels approaching the lateral resolution limit of the imaging

system, the measured intensity profile is a convolution of the true vessel profile with the point spread function (PSF) of the incident beam on the retina. Since the lateral resolution of OCT is typically worse than the axial resolution and is determined by the beam optics rather than the source bandwidth, this convolution becomes increasingly significant for smaller vessels. If the PSF were known, a deconvolution approach could be used to recover the true vessel profile before model fitting, potentially improving accuracy for capillary-scale vessels. However, characterising the retinal PSF in vivo remains challenging, as it depends on the ocular optics of each individual eye and may vary with retinal depth and eccentricity.

5. Conclusion

This study characterised the underestimation of vessel diameter in OCTA and evaluated how it can be corrected. We established ground truth in the in vivo mouse retina using Intralipid contrast enhancement. Binary mask based diameter calculation was then compared against model-based profile fitting, on both maximum and mean intensity projections. Every method underestimated vessel diameter before contrast. The generalised Gaussian $1/e^2$ metric retained the strongest linear relationship with the ground truth under these conditions, so the residual underestimation is systematic and can be removed by calibration. Post-contrast, the same metric applied to the mean intensity projection provided results closest to the ground truth of all combinations tested. Binary mask based measurements depended on the threshold and the contrast condition, with the post-contrast slope falling steeply as the threshold increased. The practical consequence of that dependence was demonstrated by the flicker stimulation results. As the angiogram signal declined over the course of the acquisition, binary mask based diameters followed it and reported vasoconstriction. The model-based fits were effectively independent of intensity and recovered the expected vasodilation. We therefore recommend a generalised Gaussian fit to the mean intensity projection for quantitative OCTA vessel diameter measurement, with a calibrated correction applied where contrast enhancement is not available.

References

1. Y. Ueno, T. Iwase, K. Goto, *et al.*, "Association of changes of retinal vessels diameter with ocular blood flow in eyes with diabetic retinopathy," *Sci. Reports* **11**, 4653 (2021).
2. M. K. Ikram, J. C. M. Witteman, J. R. Vingerling, *et al.*, "Retinal vessel diameters and risk of hypertension: The Rotterdam study," *Hypertension* **47**, 189–194 (2006).
3. A. Tirsi, N. Leung, R. Gupta, *et al.*, "Retinal vessel diameter reductions are associated with retinal ganglion cell dysfunction, thinning of the ganglion cell and inner plexiform layers, and decreased visual field global indices in glaucoma suspects," *Diagnostics* **15**, 1700 (2025).
4. C. Liang, C. Gu, and N. Wang, "Retinal vascular caliber in coronary heart disease and its risk factors," *Ophthalmic Res.* **66**, 151–163 (2022).
5. H. L. Takusagawa, L. Liu, K. N. Ma, *et al.*, "Projection-resolved optical coherence tomography angiography of macular retinal circulation in glaucoma," *Ophthalmology* **124**, 1589–1599 (2017).
6. T. S. Hwang, M. Zhang, K. Bhavsar, *et al.*, "Visualization of 3 distinct retinal plexuses by projection-resolved optical coherence tomography angiography in diabetic retinopathy," *JAMA Ophthalmol.* **134**, 1411–1419 (2016).
7. F. Coscas, A. Glacet-Bernard, A. Miere, *et al.*, "Optical coherence tomography angiography in retinal vein occlusion: Evaluation of superficial and deep capillary plexa," *Am. J. Ophthalmol.* **161**, 160–171 (2016).
8. D. Kannenkeril, J. M. Harazny, A. Bosch, *et al.*, "Retinal vascular resistance in arterial hypertension," *Blood Press.* **27**, 82–87 (2018).
9. J. P. Mynard, A. Kondiboyina, R. Kowalski, *et al.*, "Measurement, analysis and interpretation of pressure/flow waves in blood vessels," *Front. Physiol.* **11**, 1085 (2020).
10. A. C. Gonçalves Seabra, A. F. Silva, T. Stieglitz, and A. B. Amado-Rey, "In silico blood pressure models comparison," *IEEE Sensors J.* **22**, 23486–23493 (2022).
11. R. J. Hernandez, S. Madhusudhan, Y. Zheng, and W. K. El-Bouri, "Linking vascular structure and function: Image-based virtual populations of the retina," *Investig. Ophthalmol. & Vis. Sci.* **65**, 40 (2024).
12. R. J. Hernandez and W. K. El-Bouri, "Microvascular retinal digital twins from non-invasive clinical images," in *Digital Twin for Healthcare (DT4H 2025)*, vol. 16193 of *Lecture Notes in Computer Science* (Springer, Cham, 2026), pp. 12–22.
13. D. Huang, E. A. Swanson, C. P. Lin, *et al.*, "Optical coherence tomography," *Sci. (New York, N.Y.)* **254**, 1178–1181 (1991).

side

underestimation resulting from OCTA artifacts

what conditions means making the residual underest correctable by and highest precision highly intensity depends

needs funding and acknowledgment

14. A. F. Fercher, C. K. Hitzenberger, W. Drexler, *et al.*, "In vivo optical coherence tomography," *Am. J. Ophthalmol.* **116**, 113–114 (1993).
15. S. Makita, Y. Hong, M. Yamanari, *et al.*, "Optical coherence angiography," *Opt. Express* **14**, 7821–7840 (2006).
16. V. J. Srinivasan, J. Y. Jiang, M. A. Yaseen, *et al.*, "Rapid volumetric angiography of cortical microvasculature with optical coherence tomography," *Opt. Lett.* **35**, 43–45 (2010).
17. Y. K. Tao, A. M. Davis, and J. A. Izatt, "Single-pass volumetric bidirectional blood flow imaging spectral domain optical coherence tomography using a modified Hilbert transform," *Opt. Express* **16**, 12350–12361 (2008).
18. R. K. Wang, S. L. Jacques, Z. Ma, *et al.*, "Three dimensional optical angiography," *Opt. Express* **15**, 4083–4097 (2007).
19. A. Javed, A. Khanna, E. Palmer, *et al.*, "Optical coherence tomography angiography: A review of the current literature," *J. Int. Med. Res.* **51**, 03000605231187933 (2023).
20. B. Tan, R. Sim, J. Chua, *et al.*, "Approaches to quantify optical coherence tomography angiography metrics," *Ann. Transl. Med.* **8**, 1205 (2020).
21. P. Zabel, J. J. Kaluzny, M. Wilkosc-Debczynska, *et al.*, "Comparison of retinal microvasculature in patients with Alzheimer's disease and primary open-angle glaucoma by optical coherence tomography angiography," *Investig. Ophthalmol. & Vis. Sci.* **60**, 3447–3455 (2019).
22. W. Tan, X. Yao, T.-T. Le, *et al.*, "The application of optical coherence tomography angiography in systemic hypertension: A meta-analysis," *Front. Med.* **8**, 778330 (2021).
23. X. Yao, M. Ke, Y. Ho, *et al.*, "Comparison of retinal vessel diameter measurements from swept-source OCT angiography and adaptive optics ophthalmoscope," *The Br. J. Ophthalmol.* **105**, 426–431 (2021).
24. H. Lee, M. Lee, H. Chung, and H. C. Kim, "Quantification of retinal vessel tortuosity in diabetic retinopathy using optical coherence tomography angiography," *Retina* **38**, 976–985 (2018).
25. N. Mehta, K. Liu, A. Y. Alibhai, *et al.*, "Impact of binarization thresholding and brightness/contrast adjustment methodology on optical coherence tomography angiography image quantification," *Am. J. Ophthalmol.* **205**, 54–65 (2019).
26. A. Rabiolo, F. Gelormini, R. Sacconi, *et al.*, "Comparison of methods to quantify macular and peripapillary vessel density in optical coherence tomography angiography," *PLoS One* **13**, e0205773 (2018).
27. E. Borrelli, R. Sacconi, M. Parravano, *et al.*, "Optical coherence tomography angiography assessment of the diabetic macula: A comparison study among different algorithms," *Retina* **41**, 1799–1808 (2021).
28. I. G. Freedman, E. Li, L. Hui, *et al.*, "The impact of image processing algorithms on optical coherence tomography angiography metrics and study conclusions in diabetic retinopathy," *Transl. Vis. Sci. & Technol.* **11**, 7 (2022).
29. A. Arrigo, E. Aragona, A. Saladino, *et al.*, "The impact of different thresholds on optical coherence tomography angiography images binarization and quantitative metrics," *Sci. Reports* **11**, 14758 (2021).
30. J. Hong, B. Tan, N. D. Quang, *et al.*, "Intra-session repeatability of quantitative metrics using widefield optical coherence tomography angiography (OCTA) in elderly subjects," *Acta Ophthalmol.* **98**, e570–e578 (2020).
31. J. Hong, M. Ke, B. Tan, *et al.*, "Effect of vessel enhancement filters on the repeatability of measurements obtained from widefield swept-source optical coherence tomography angiography," *Sci. Reports* **10**, 22179 (2020).
32. P. Cimalla, J. Walther, M. Mittasch, and E. Koch, "Shear flow-induced optical inhomogeneity of blood assessed in vivo and in vitro by spectral domain optical coherence tomography in the 1.3 μm wavelength range," *J. Biomed. Opt.* **16**, 116020 (2011).
33. J. Zhu, M. T. Bernucci, C. W. Merkle, and V. J. Srinivasan, "Visibility of microvessels in optical coherence tomography angiography depends on angular orientation," *J. Biophotonics* **13**, e202000090 (2020).
34. M. T. Bernucci, C. W. Merkle, and V. J. Srinivasan, "Investigation of artifacts in retinal and choroidal OCT angiography with a contrast agent," *Biomed. Opt. Express* **9**, 1020–1040 (2018).
35. Y. Pan, J. You, N. D. Volkow, *et al.*, "Ultrasensitive detection of 3D cerebral microvascular network dynamics in vivo," *NeuroImage* **103**, 492–501 (2014).
36. C. W. Merkle, C. Leahy, and V. J. Srinivasan, "Dynamic contrast optical coherence tomography images transit time and quantifies microvascular plasma volume and flow in the retina and choriocapillaris," *Biomed. Opt. Express* **7**, 4289 (2016).
37. C. W. Merkle and V. J. Srinivasan, "Laminar microvascular transit time distribution in the mouse somatosensory cortex revealed by dynamic contrast optical coherence tomography," *NeuroImage* **125**, 350–362 (2016).
38. L. Zhou, M. S. Rzeszutarski, L. J. Singerman, and J. M. Chokreff, "The detection and quantification of retinopathy using digital angiograms," *IEEE Trans. on Med. Imaging* **13**, 619–626 (1994).
39. J. Lowell, A. Hunter, D. Steel, *et al.*, "Measurement of retinal vessel widths from fundus images based on 2-D modeling," *IEEE Trans. on Med. Imaging* **23**, 1196–1204 (2004).
40. C. Leahy, H. Radhakrishnan, G. Weiner, *et al.*, "Mapping the 3D connectivity of the rat inner retinal vascular network using OCT angiography," *Investig. Ophthalmol. & Vis. Sci.* **56**, 5785–5793 (2015).
41. G. Garhöfer, C. Zawinka, H. Resch, *et al.*, "Diffuse luminance flicker increases blood flow in major retinal arteries and veins," *Vis. Res.* **44**, 833–838 (2004).
42. M. Sharifzad, K. J. Witkowska, G. C. Aschinger, *et al.*, "Factors determining flicker-induced retinal vasodilation in healthy subjects," *Investig. Ophthalmol. & Vis. Sci.* **57**, 3306–3312 (2016).
43. G. Michelson, A. Patzelt, and J. Harazny, "Flickering light increases retinal blood flow," *Retina* **22**, 336–343 (2002).
44. G. C. Aschinger, L. Schmetterer, K. Fondi, *et al.*, "Effect of diffuse luminance flicker light stimulation on total retinal

- blood flow assessed with dual-beam bidirectional Doppler OCT," *Investig. Ophthalmol. & Vis. Sci.* **58**, 1167–1178 (2017).
45. T. E. Kornfield and E. A. Newman, "Regulation of blood flow in the retinal trilateral vascular network," *The J. Neurosci.* **34**, 11504 LP – 11513 (2014).
 46. Y.-Y. I. Shih, L. Wang, B. H. De La Garza, *et al.*, "Quantitative retinal and choroidal blood flow during light, dark adaptation and flicker light stimulation in rats using fluorescent microspheres," *Curr. Eye Res.* **38**, 292–298 (2013).
 47. T. Son, B. Wang, D. Thapa, *et al.*, "Optical coherence tomography angiography of stimulus evoked hemodynamic responses in individual retinal layers," *Biomed. Opt. Express* **7**, 3151–3162 (2016).
 48. S. Fialová, M. Augustin, M. Glösmann, *et al.*, "Polarization properties of single layers in the posterior eyes of mice and rats investigated using high resolution polarization sensitive optical coherence tomography," *Biomed. Opt. Express* **7**, 1479–1495 (2016).
 49. C. W. Merkle, M. Augustin, D. J. Harper, *et al.*, "High-resolution, depth-resolved vascular leakage measurements using contrast-enhanced, correlation-gated optical coherence tomography in mice," *Biomed. Opt. Express* **12**, 1774–1791 (2021).
 50. Y. Fu and K.-W. Yau, "Phototransduction in mouse rods and cones," *Pflügers Arch. J. Physiol.* **454**, 805–819 (2007).
 51. C. T. Sullender, L. M. Richards, F. He, *et al.*, "Dynamics of isoflurane-induced vasodilation and blood flow of cerebral vasculature revealed by multi-exposure speckle imaging," *J. Neurosci. Methods* **366**, 109434 (2022).
 52. S. Worm, G. Ladurner, L. May, *et al.*, "Retinal OCT parameters and physiological dynamics as an effect of anesthesia in C57BL/6J mice," *Investig. Ophthalmol. & Vis. Sci.* **67**, 50 (2026).
 53. M. Augustin, S. Fialová, R. Plasenzotti, *et al.*, "Multi-functional OCT image processing for rodent eyes," (OSA - The Optical Society, Center for Medical Physics and Biomedical Engineering, Medical University of Vienna, Wachringerguertel 18-20, 4L, Vienna, A-1090, Austria, 2016), p. 3.
 54. G. R. Untracht, M. S. Durkee, M. Zhao, *et al.*, "Towards standardising retinal OCT angiography image analysis with open-source toolbox OCTAVA," *Sci. Reports* **14** (2024).
 55. T. T. Hormel, J. Wang, S. T. Bailey, *et al.*, "Maximum value projection produces better en face OCT angiograms than mean value projection," *Biomed. Opt. Express* **9**, 6412 (2018).
 56. D. M. Sampson, A. M. Dubis, F. K. Chen, *et al.*, "Towards standardizing retinal optical coherence tomography angiography: a review," *Light. Sci. & Appl.* **11** (2022).
 57. G. Feng, A. Joseph, K. Dholakia, *et al.*, "High-resolution structural and functional retinal imaging in the awake behaving mouse," *Commun. Biol.* **6**, 572 (2023).
 58. D. M. Stein, G. Wollstein, H. Ishikawa, *et al.*, "Effect of corneal drying on optical coherence tomography," *Ophthalmology* **113**, 985–991 (2006).
 59. G. J. McLellan and C. A. Rasmussen, "Optical coherence tomography for the evaluation of retinal and optic nerve morphology in animal subjects: Practical considerations," *Vet. Ophthalmol.* **15**, 13–28 (2012).
 60. A. F. Frangi, W. J. Niessen, K. L. Vincken, and M. A. Viergever, "Multiscale vessel enhancement filtering," in *Medical Image Computing and Computer-Assisted Intervention — MICCAI'98*, vol. 1496 of *Lecture Notes in Computer Science* W. M. Wells, A. Colchester, and S. Delp, eds. (Springer, Berlin, Heidelberg, 1998), pp. 130–137.
 61. F. Formaz, C. E. Riva, and M. Geiser, "Diffuse luminance flicker increases retinal vessel diameter in humans," *Curr. Eye Res.* **16**, 1252–1257 (1997).
 62. K. Polak, L. Schmetterer, and C. E. Riva, "Influence of flicker frequency on flicker-induced changes of retinal vessel diameter," *Investig. Ophthalmol. & Vis. Sci.* **43**, 2721–2726 (2002).
 63. E. Nagel and W. Vilser, "Flicker observation light induces diameter response in retinal arterioles: A clinical methodological study," *Der Ophthalmol.* **102**, 305–309 (2005).
 64. M. Kallab, N. Hommer, B. Tan, *et al.*, "Plexus-specific effect of flicker-light stimulation on the retinal microvasculature assessed with optical coherence tomography angiography," *Am. J. Physiol. Circ. Physiol.* **320**, H23–H28 (2020).
 65. T. Son, G. Ma, and X. Yao, "Functional OCT reveals anisotropic changes of retinal flicker-evoked vasodilation," *Opt. Lett.* **49**, 2121–2124 (2024).
 66. H. Radhakrishnan and V. J. Srinivasan, "Multiparametric optical coherence tomography imaging of the inner retinal hemodynamic response to visual stimulation," *J. Biomed. Opt.* **18**, 086010 (2013).
 67. W. Albanna, K. Kotliar, J. N. Lüke, *et al.*, "Non-invasive evaluation of neurovascular coupling in the murine retina by dynamic retinal vessel analysis," *PLOS ONE* **13**, e0204689 (2018).
 68. F. Neumaier, K. Kotliar, R. H. L. Haeren, *et al.*, "Evaluation of flicker-induced retinal vessel response in Cav2.3-deficient mice using dynamic retinal vessel analysis," *Front. Neurol.* **12**, 659890 (2021).
 69. M. Rai, Y. Lakshmanan, K. Y. Choi, and H. H.-I. Chan, "Effects of flickering light stimulation on retinal blood flow and full-field electroretinogram in mice," *Documenta Ophthalmol.* **151**, 205–218 (2025).