

Universal manuscript template for Optica Publishing Group journals

YASH PATEL,¹ PRAJAKTA BELEKAR,¹ GEORG LADURNER,¹ SYBREN WORM,¹ LUCAS MAY,¹ MARIA VARAKA,¹ THERESA ROHRMOSER,¹ FERNANDO GARCÍA-RAMÍREZ,¹ RENÉ WERKMEISTER,¹ BERNHARD BAUMANN,^{1,2} AND CONRAD W. MERKLE^{1,*}

¹Center for Medical Physics and Biomedical Engineering, Medical University of Vienna, Währinger Gürtel 18-20/4L, 1090 Vienna, Austria

²Institute of Biomedical Physics, Medical University of Innsbruck, Müllerstraße 44, 6020 Innsbruck, Austria

*conrad.merkle@meduniwien.ac.at

Abstract: This template contains important information on submissions to Optica Publishing Group journals. We encourage authors to focus foremost on their content and not on formatting. The template is provided as a guide, but following the visual styling is optional. Each manuscript will be formatted in a consistent way during production. Authors also have the option to [submit articles](#) to the Optica Publishing Group preprint server, [Optica Open](#). You may find it helpful to use our optional [Paperpal manuscript readiness check](#) and [language polishing service](#). Note that copyright and licensing information should not be added to your journal or Optica Open manuscript.

1. Introduction

Optical coherence tomography angiography (OCTA) is an extension of OCT that visualizes vascular perfusion using the motion contrast from red blood cells (RBCs). Over the years, OCTA has demonstrated widespread acceptance and use for clinical evaluations of vascular features in relation to disease [1]. OCT angiograms can be used to evaluate characteristics on the scale of vascular layers by measuring vascular density or non-perfusion zones, or it can be used to examine individual vessels with measurements of diameter or tortuosity. Nevertheless, OCT is prone to certain artifacts which specifically affect visualization of the vasculature. These artifacts may lead to underestimations of measurements of vessel diameter and density. When viewed in cross-section, large vessels often demonstrate an hourglass appearance with stronger scattering in the center than at the sides (Figure 1A). This appearance is due to orientation effects of RBCs, which align with the blood flow, causing a reduced scattering cross-section at the sides of the vessels. To compensate for these artifacts, OCT contrast agents such as Intralipid have previously been used to increase intravascular intensity signals, particularly in regions where RBC scattering is diminished. Here we examine different methods of measuring vessel size from in vivo OCT scans of mouse retinal vasculature, and we use a contrast agent to obtain ground truth vessel diameter measurements for comparison. We finally propose ways of mitigating or correcting for underestimations of vessel diameter related to typical OCT artifacts.

Optical coherence tomography (OCT) is a non-invasive optical imaging method, which uses interferometry to measure the transit time of light that is backscattered from a sample. OCT has found widespread clinical use in ophthalmology with XXX number of systems and XXX OCT scans performed annually around the world. OCT angiography (OCTA) is a functional extension of OCT which highlights motion between repeated OCT scans to generate images of vascular networks. Typically, OCTA signals are first calculated from the OCT intensity and/or phase information, segmented in 3D manually, algorithmically, or using a deep learning approach, and finally compressed into a 2D image using a maximum intensity projection. These 2D vascular maps can then be binarised by using a global or adaptive threshold, and finally vascular metrics

such as vessel density and diameter can be computed. Of these processing stages, the thresholding step is perhaps the most important. There are dozens of different thresholding methods that have been proposed over the years and each of these may produce different results in the binary map, which would in turn influence the quantification of vascular metrics. In this work, we examine an alternate approach for extracting quantitative vessel diameter by fitting a model to the OCTA data before binarisation to avoid thresholding-based effects.

In addition to thresholding-based variability, we hypothesize that OCTA metrics are also affected by OCT artifacts caused by the non-uniform backscattering profiles of red blood cells (RBCs). Because RBCs have a distinct bi-concave disc shape, they more effectively backscatter light when aligned perpendicular to the beam, rather than parallel to it. It has already been observed that in large vessels, a so-called hourglass artifact appears when viewed cross-sectionally. This is caused by the alignment of the RBCs with the vessel walls while under laminar flow, where the RBCs are perpendicular to the beam in the center of the vessels and parallel to the beam at the edges of the vessels. This artifact artificially suppresses the OCT and OCTA signals at the edges of the vessels, making them appear more narrow when projected into an en-face view, which is the preferred way of displaying angiographic data. Here we use contrast enhanced OCT (CE-OCT) to investigate how significant these effects are and how they can be corrected.

2. Methods and Materials

2.1. Imaging system

All imaging was performed using a custom-built polarisation-sensitive spectral-domain OCT ophthalmoscope, described in detail previously [2]. The system was centred at 840 nm with a spectral bandwidth of 100 nm at full width half maximum (FWHM), providing an axial resolution of approximately 5.1 μm in air and 3.8 μm in tissue (assuming a refractive index of 1.35). A sensitivity of 96 dB was achieved at an A-scan rate of 80 kHz and a beam diameter of 0.5 mm at the pupil plane (power at cornea = 2.85 mW) [3]. This system was employed to acquire volumetric OCT data from mouse retinas before, during, and after an intravascular contrast agent administration, as well as before and after functional stimulation of the mouse retina using a custom 502 nm light-emitting diode (LED) ring flashing at 10 Hz (power at cornea = 26.6 μW). This wavelength was selected to target the peak absorption of mouse rhodopsin (~500 nm), enabling predominantly rod-driven photoreceptor stimulation [4].

2.2. Experimental protocol

Three mouse models were used in this study: VLDLR knockout mice (B6;129S7-*Vldlr*^{tm1Her}/J, The Jackson Laboratory, Bar Harbor, USA), C57BL/6 mice (C57BL/6, The Jackson Laboratory, Bar Harbor, USA), and SOD1 transgenic mice (strain details, The Jackson Laboratory, Bar Harbor, USA), with $n = \text{XX}$ mice per group. All imaging sessions were conducted under general anaesthesia using either isoflurane or a ketamine/xylazine combination. Isoflurane was induced at 4% in oxygen for 4 minutes and subsequently maintained at 2% in oxygen at a flow rate of 1.5–2 L/min. In cases where isoflurane alone did not provide adequate sedation, an intraperitoneal injection of ketamine/xylazine cocktail (100 mg/kg ketamine, 6 mg/kg xylazine; 10 mL/kg body weight) was administered. To maintain core body temperature of the mice throughout imaging, they were positioned on a heating pad. Mouse pupils were dilated with one drop of tropicamide (5 mg/mL, Agepha Pharma s.r.o., Senec, Slovakia) per eye, and artificial tear drops (Oculotect, ALCON Pharma GmbH, Freiburg im Breisgau, Germany) were applied at regular intervals to prevent the cornea from drying out.

Intralipid has been shown to be a highly scattering contrast agent that fills the vascular lumen uniformly, thereby compensating for the artefact caused by the diminished backscattering at vessel peripheries caused by RBC orientation effects [3,5,6]. For this study, Intralipid 20% (3 mL/kg

body weight) was administered as a bolus injection via the tail vein. Volumetric angiography data were acquired over a 1 mm × 1 mm field of view centred on the optic nerve head (ONH), with a scan density of 500 A-scans per B-scan and five repeated B-scans at each of 400 slow-axis positions (acquisition time approximately 16 seconds). For this study, pre- and post-injection datasets and pre- and post-stimulation datasets were acquired. **These volumetric data were used solely for qualitative visualisation of the contrast enhancement effect across the retinal vascular plexuses and were not used for quantitative diameter analysis.** Additionally, a dynamic-contrast OCT (DyC-OCT) protocol [7] was employed during the contrast injection or during stimulation, consisting of 2000 repeated B-scans over approximately 16 seconds (interscan time 8.2 ms) at the same cross-sectional location over a 1 mm lateral field of view. DyC-OCT scans allow us to track the same vessels over time to precisely measure the effects of the contrast agent or the stimulation with a high degree of spatial coregistration. **All quantitative vessel diameter measurements reported in this study were derived from the DyC-OCT data.**

Functional flicker stimulation was performed in a C57BL/6 mouse to assess whether the choice of diameter measurement method affects the detection of physiological vascular responses. For this experiment, ketamine/xylazine cocktail was chosen over isoflurane to minimise anaesthesia-induced vasodilation. A DyC-OCT acquisition was performed during which the retina was unstimulated for approximately 4 seconds (baseline), followed by 10 Hz flicker stimulation from the LED ring for approximately 7.5 seconds, and a post-stimulation period of approximately 4.5 seconds **(I wrote this to match the pre- and post-stim lines in the stimulation result plots) CM[This isn't the stimulation protocol that was used though. The lines just mark where we were averaging, but I believe in this data set there was already stimulation before the data started acquiring and it continued through the full acquisition. Do you remember how much this was?].**

All animal procedures were approved by the ethics committee of the Medical University of Vienna and the Austrian Federal Ministry of Education, Science, and Research (BMBWF/66.009/0272-V/3b/2019 and GZ 2024-0.802.524).

2.3. Post-processing and analysis

Cross-polarised OCT intensity data from the retinal pigment epithelium (RPE) were used for inter-frame alignment and B-scan flattening [3, 8]. Angiograms were derived by applying the complex subtraction OCT angiography (OCTA) algorithm [9, 10]. We evaluated multiple combinations of methods for quantifying vessel diameter from the OCTA data to determine which combinations of steps work best under different conditions. Specifically, we looked at two types of data projections (maximum intensity and mean intensity) to generate vessel profiles and several different quantifications of those profiles including a traditional binarisation, a Gaussian model approach, and a Generalised-Gaussian (GG) model approach. These methods represent commonly used methods in the literature for OCTA quantification [11–14]. The different methods were evaluated against the ground truth as described in this section. All data processing and analysis was performed in MATLAB (R2019b, MathWorks, Natick, MA, USA).

2.3.1. Ground truth calculation

Ground truth vessel diameters were obtained from contrast-enhanced DyC-OCT B-scan angiograms, where the uniformly scattering Intralipid fills the full vascular lumen regardless of RBC orientation. The width and the height of the vessel were manually measured from the vessel's cross-section. The ground truth diameter was defined as the minimum of the two measurements to account for off-axis cross-sectioning of the vessel, which produces a more elliptical appearance in the B-scan. For cases where the vessel is perpendicular to the B-scan, the height and width would therefore be in agreement. Because the lateral diameter is studied here, we defined a correction factor of *height/width* for cases where the width was larger than the height due to skew effects and was equal to 1 otherwise. By multiplying this correction factor against diameter

141 measured from the ROI projection, we ensured that the measured vessel diameter reflected the
 142 true size of the vessel.

143 2.3.2. Vessel projection calculation

144 For each vessel of interest, a rectangular region of interest (ROI) encompassing the vessel cross-
 145 section, along the lateral (x) and axial (z) dimensions, was selected from the DyC-OCT B-scan
 146 angiograms. En face intensity profiles were then generated by projecting the A-scans within this
 147 ROI along the axial (z) direction yielding a time-stack of 1-D lateral intensity profiles for each
 148 vessel. Two projection methods were compared: maximum intensity projection (MIP), which
 149 selects the highest intensity value along the depth axis of each ROI, and mean intensity projection
 150 (MeIP), which averages the values along each ROI. Each projection resulted in a time-stack of
 151 one-dimensional intensity profiles for each vessel, which were then used for subsequent diameter
 152 quantification. For the model-based approaches, the profiles were normalised to a range of 0–1
 153 to ensure a good fit.

154 2.3.3. Model-based vessel diameter calculation

155 The two parametric models were fitted (Fig. 1a) to the one-dimensional intensity profiles from
 156 the previous step, using nonlinear least-squares regression in MATLAB.

157 Standard Gaussian function:

$$f(x) = A \exp\left(-\frac{(x - \mu)^2}{2\sigma^2}\right) + C \quad (1)$$

158 where A is the peak amplitude, μ is the centre position, σ is the standard deviation, and C is a
 159 constant baseline offset.

160 Generalised-Gaussian (GG) function:

$$f(x) = A \exp\left(-\left(\frac{|x - \mu|}{\alpha}\right)^\beta\right) + C \quad (2)$$

161 where A is the peak amplitude, μ is the centre position, α is the scale parameter, β is the shape
 162 parameter, and C is a constant baseline offset. Given the imaging system configuration, all vessels
 163 are expected to produce at minimum a Gaussian intensity profile in the B-scan cross-section,
 164 and so profiles sharper than Gaussian profile ($\beta < 2$) are therefore not technically meaningful.
 165 The shape parameter β was hence constrained to $\beta \geq 2$ during fitting. With this constraint, the
 166 GG accommodated to profiles ranging from Gaussian ($\beta = 2$, where $\sigma = \alpha/\sqrt{2}$) to increasingly
 167 flat-topped ($\beta > 2$).

168 Two standard width metrics were calculated from the fitted models: the full width at half
 169 maximum (FWHM) and the $1/e^2$ width. For the standard Gaussian fit, the $1/e^2$ width was
 170 calculated as:

$$w_{1/e^2, G} = 4\sigma \quad (3)$$

171 The FWHM for a Gaussian is related to the $1/e^2$ width by a fixed proportionality:

$$\text{FWHM}_G = 2\sigma\sqrt{2\ln 2} = \frac{\sqrt{2\ln 2}}{2} w_{1/e^2, G} \quad (4)$$

172 and therefore only the $1/e^2$ width was calculated directly from the fit, with the FWHM obtainable
 173 via this known relationship.

174 For the GG fit, no such linear relationship exists between the FWHM and $1/e^2$ width, since
 175 both depend on the shape parameter β . Both metrics were therefore calculated independently as:

$$\text{FWHM}_{GG} = 2\alpha(\ln 2)^{1/\beta} \quad (5)$$

176

$$w_{1/e^2, GG} = 2\alpha \cdot 2^{1/\beta} \quad (6)$$




Fig. 1. (a) GG function for shape parameters $\beta = 1, 2, 3$, and 4 . At $\beta = 2$ the function reduces to a standard Gaussian; higher β values produce increasingly flat-topped profiles. (b) Example en face vessel intensity profile (black) with standard Gaussian (blue) and GG (red) fits.

2.3.4. Binary mask based vessel diameter calculation

For comparison with OCTA processing pipelines that rely on image binarisation, vessel diameters were also measured from binary masks generated by global thresholding. A 3D averaging filter with a kernel size of $50 \times 50 \times 10$ voxels (lateral $x \times$ axial $z \times$ temporal t) was applied to the DyC-OCT angiogram volumes to reduce noise and minimize the effects of outliers. To produce a binarised black and white (BW) mask, the maximum angiogram intensity within these smoothed volumes were calculated and five different threshold values were defined as fixed percentages of these maxima: **BW-1 = XX%, BW-2 = XX%, BW-3 = XX%, BW-4 = XX%, and BW-5 = XX%**. Each threshold was applied to the unsmoothed en face projection (MIP and MeIP) to generate a binary mask, and the vessel diameter was measured as the number of pixels above the threshold across the lateral vessel cross-section.

2.3.5. Statistical analysis

To evaluate the accuracy and precision of all diameter measurement approaches, the calculated diameter values were plotted against the ground truth for each vessel across three conditions: pre-contrast, peak contrast, and post-contrast. A linear regression was fitted to each condition, yielding the slope (m) and coefficient of determination (R^2). A slope of $m = 1$ indicated perfect agreement with the ground truth, and R^2 quantified the goodness of fit of the linear model. Additionally, the coefficient of variation (CoV) was computed for each method to assess measurement precision (repeatability). For the model-based methods, six combinations were evaluated: two projection methods (MIP and MeIP) by three width metrics ($w_{1/e^2, G}$, $FWHM_{GG}$, and $w_{1/e^2, GG}$). For the binarisation-based methods, diameter was measured at five threshold levels (BW-1 through BW-5) for both MIP and MeIP. For the functional flicker stimulation experiment, the Pearson correlation coefficient (ρ) between measured diameter values and angiogram intensity was computed for each method and vessel, to assess whether the diameter measurements were affected by changes in angiogram intensity. Measurements with $p < 0.05$ were considered statistically significant.

3. Results

3.1. Qualitative assessment of vessel underestimation

Figure 2 shows OCTA MIP enface images of the superficial vascular plexus (SVP), intermediate capillary plexus (ICP), and deep capillary plexus (DCP) before and after contrast injection. Across




Fig. 2. OCTA MIP en face of the SVP, ICP, and DCP before (left) and after (right) the administration of Intralipid 20%. Vessels appear visibly wider and brighter post-contrast across all layers, most prominently in the larger vessels.

all three layers, vessels appeared visibly wider and brighter post-contrast injection, consistent with previous literature showing that OCTA underestimates vessel diameter in the absence of a contrast agent [3]. This effect was most pronounced in the larger vessels, especially in the SVP, where the major retinal arteries and veins appeared noticeably thicker after contrast administration. In the ICP and DCP, the capillary networks also showed improved visibility, though the effect was less dramatic.

DyC OCTA projections helped visualise the temporal changes in vessel diameter pre- and post-contrast injection (Fig. 3). Pre-contrast injection, the vessel cross-sections appeared narrower than their true diameter (Fig. 3a), which could be due to lateral edge signal suppression by the RBC orientation artefact [15]. At peak contrast, the Intralipid filled the vascular lumen uniformly and showed the full extent of the vessel cross-section (Fig. 3b). However, the high scattering nature of the contrast agent produced light diffusion beyond the vessel wall, resulting in an apparent vessel diameter that slightly exceeded the true anatomical diameter. In the post-contrast phase, the signal stabilised as the initial bolus cleared, and the vessel boundaries remained more visible than before injection without the overestimation observed at peak contrast (Fig. 3c). The MIP and the corresponding binary mask over time (Figs. 3d–e) showed the observed temporal changes in vessel diameter across the three phases. The mean OCTA intensity (Fig. 3f) showed a sharp increase post-contrast injection, followed by a decrease to a new steady state that remained elevated above the pre-contrast baseline. These observations confirm that the contrast agent does significantly enhance vessel visibility and diameter, and that the peak- and post-contrast images show improved vessel delineation compared to pre-contrast.




Fig. 3. DyC-OCT B-scan angiograms at (a) pre-contrast, (b) peak contrast, and (c) post-contrast timepoints. (d) MIP and (e) binarisation of the vessel cross-sections over time, with coloured boxes indicating the pre-contrast (blue), peak-contrast (green), and post-contrast (red) windows. (f) Mean OCTA intensity over time, showing the temporal intensity dynamics. All scale bars are 100 μm .

3.2. Accuracy and precision of diameter measurements

The accuracy (m , R^2) and precision (CoV) of all diameter measurement methods are summarised in Table 1.

All model-based methods underestimated vessel diameter in pre-contrast data, with slopes ranging from 0.42 to 0.73 (Table 1). Contrast enhancement improved accuracy across all metrics, though the degree of improvement varied. The FWHM_{GG} metric remained well below unity even post-contrast. The $w_{1/e^2, \text{G}}$ with MIP slightly overestimated post-contrast ($m = 1.12\text{--}1.15$), while MeIP yielded values closer to unity ($m = 1.04$). The $w_{1/e^2, \text{GG}}$ with MeIP provided the best overall agreement, achieving a slope closest to unity with $m = 1.001$ ($R^2 = 0.97$) post-contrast, which was the closest to the identity line of all tested model-based methods. Across all metrics, MeIP mostly outperformed MIP in both accuracy and linearity (Table 1, Fig. 4).

For binary mask based methods, the slope decreased with increasing threshold across all contrast conditions and projections (Table 1). This degradation was steeper for MIP than for MeIP. MIP slopes dropped from near unity at BW-1 to well below it at BW-5, whereas MeIP slopes declined more gradually and remained closer to unity across the threshold range. MeIP slopes were consistently higher than MIP at every threshold and contrast condition, though at lower thresholds and at peak and post-contrast, MeIP tended to slightly overestimate vessel diameter. The R^2 values followed a similar pattern. For MIP, pre- and post-contrast R^2 decreased substantially with increasing threshold, whereas for MeIP, R^2 remained relatively stable. Overall, MeIP provided more reliable binarisation-based diameter measurements than MIP.

The precision of each method was assessed via the CoV (Table 1, Fig. 7). For MIP, the binary mask based CoV increased substantially with threshold, whereas the model-based CoV remained stable across all contrast conditions. For MeIP, the binary mask based CoV was notably more stable across thresholds than for MIP, and the model-based CoV values were lower overall, though they increased slightly at peak contrast before decreasing post-contrast. Across both methods, MeIP yielded lower CoV values than MIP pre-contrast and more stable values post-contrast. Among the model-based metrics, FWHM_{GG} consistently exhibited the weakest accuracy while




Fig. 4. Modelled versus ground truth vessel diameter using MIP (top row) and MeIP (bottom row). Columns correspond to (a, b) $w_{1/e^2, G}$, (c, d) FWHM_{GG} , and (e, f) $w_{1/e^2, GG}$. Linear regression lines and statistics are shown for pre-contrast (blue), peak contrast (green), and post-contrast (red) conditions. The black line indicates the identity line ($m = 1$).




Fig. 5. Vessel diameter measured from binarised masks versus ground truth for MIP at five threshold levels: (a) BW-1 through (e) BW-5. Linear regression lines and statistics are shown for pre-contrast (blue), peak contrast (green), and post-contrast (red) conditions.




Fig. 6. Vessel diameter measured from binarised masks versus ground truth for MeIP at five threshold levels: (a) BW-1 through (e) BW-5. Linear regression lines and statistics are shown for pre-contrast (blue), peak contrast (green), and post-contrast (red) conditions.

maintaining comparable precision. Overall, model-based methods consistently outperformed binarisation in both accuracy and precision, and MeIP yielded superior performance compared to MIP across all methods.

3.3. Functional flicker stimulation

The model-based diameter time series (Figs. 8a,b) showed increase in diameter post-stimulation for most of the vessels. The binary mask based diameter time series (BW-3; Fig. 8c) showed a mixed trend, where most vessels showed a decrease in diameter post-stimulation. Bar plots of the mean diameter pre- and post-stimulation (Figs. 8d-f) confirmed this pattern. For the Gaussian model, four of five vessels showed significant diameter increases ($p < 0.01$), and the GG model showed significant increases in three of five vessels (Table 2). In contrast, the binary mask based diameter showed significant decreases in four of five vessels.

The angiogram intensity time series (Fig. 8g) showed a visible decline in angiogram intensity during the duration of the DyC-OCT scan across all five vessels. The mean intensity comparison pre- and post-stimulation (Fig. 8h) confirmed that the decrease was significant for all vessels, ranging from -14.1% to -19.9% (Table 2). The binary mask based method showed uniformly positive per-vessel correlations, with an overall correlation of $\rho = 0.523$ ($p < 0.0005$). The Gaussian and GG models yielded overall correlations of $\rho = -0.101$ and $\rho = -0.098$, respectively. These results demonstrated that the model-based diameter calculation provided more accurate and repeatable measurements, and also recovered physiologically meaningful vascular responses that binary mask based methods misrepresented.




Fig. 7. Summary of accuracy and precision across all diameter measurement methods. (a, b) Coefficient of variation (CoV) and (c, d) coefficient of determination (R^2) for MIP (a, c) and MeIP (b, d), respectively. Bars are grouped by method (BW-1 through BW-5, G-e², GG-HM, GG-e²) and colour-coded by contrast condition: pre-contrast (blue), peak-contrast (green), and post-contrast (red).

Table 1. Summary of slope (m), coefficient of determination (R^2), and coefficient of variation (CoV) for all diameter measurement methods across pre-contrast, peak contrast, and post-contrast conditions. Methods are grouped as binary mask based and model-based. Bold values indicate slope agreement within 10% of unity ($0.9 \leq m \leq 1.1$), strong linearity ($R^2 \geq 0.90$), or high precision (CoV ≤ 0.10). (confirm values with MATLAB.)

	BW-1		BW-2		BW-3		BW-4		BW-5		$w_{1/e^2, G}$		FWHM _{GG}		$w_{1/e^2, GG}$	
	MIP	MeIP	MIP	MeIP	MIP	MeIP	MIP	MeIP	MIP	MeIP	MIP	MeIP	MIP	MeIP	MIP	MeIP
<i>Slope (m)</i>																
Pre	0.61	0.83	0.48	0.74	0.38	0.66	0.30	0.59	0.24	0.53	0.70	0.73	0.42	0.43	0.64	0.71
Peak	0.96	1.14	0.86	1.06	0.78	0.99	0.71	0.93	0.65	0.88	1.15	1.04	0.66	0.61	0.98	0.98
Post	0.92	1.11	0.82	1.03	0.73	0.97	0.63	0.91	0.54	0.86	1.12	1.04	0.66	0.61	1.02	1.00
<i>Coefficient of determination (R^2)</i>																
Pre	0.58	0.59	0.54	0.62	0.45	0.64	0.34	0.65	0.24	0.65	0.83	0.84	0.79	0.83	0.87	0.88
Peak	0.88	0.78	0.89	0.82	0.86	0.86	0.81	0.88	0.74	0.89	0.87	0.95	0.95	0.95	0.98	0.97
Post	0.76	0.77	0.69	0.79	0.57	0.80	0.45	0.80	0.34	0.78	0.89	0.97	0.95	0.98	0.93	0.97
<i>Coefficient of variation (CoV) (values extracted from bar plots, confirm with MATLAB)</i>																
Pre	0.20	0.20	0.24	0.20	0.29	0.19	0.37	0.19	0.47	0.20	0.26	0.14	0.26	0.15	0.25	0.14
Peak	0.17	0.16	0.20	0.15	0.23	0.17	0.27	0.18	0.32	0.18	0.22	0.16	0.23	0.17	0.21	0.15
Post	0.10	0.12	0.10	0.12	0.11	0.11	0.15	0.10	0.19	0.09	0.19	0.10	0.19	0.11	0.16	0.09




Fig. 8. Functional flicker stimulation results. (a–c) Vessel diameter time series pre- and post-stimulation for (a) Gaussian model, (b) Generalised Gaussian model, and (c) binarised angiogram, across five vessels. Vertical dashed lines indicate stimulation onset and offset. (d–f) Mean diameter per vessel before (dark bars) and after (light bars) stimulation for each method. Veins are shown in blue/cyan and arteries in dark/light red. Significance levels: ** $p < 0.01$, *** $p < 0.001$. (g) Angiogram intensity time series and (h) mean angiogram intensity per vessel pre- and post-stimulation.

Table 2. Flicker stimulation results. Δd : diameter change (%); ρ : Correlation between diameter and angiogram intensity; ΔI : intensity change (%). Statistically significant values are bold (* $p < 0.05$, ** $p < 0.01$, *** $p < 0.001$).

	Gaussian		GG		Binarised		Intensity
	Δd (%)	ρ	Δd (%)	ρ	Δd (%)	ρ	
Vessel 1	+9.2***	-0.07	+1.5	0.03	-12.5***	0.59***	-14.1***
Vessel 2	+9.6**	0.23*	+9.9***	0.07	-10.2***	0.80***	-18.2***
Vessel 3	+1.5	-0.28**	-0.6	-0.28**	-4.2**	0.27**	-14.3***
Vessel 4	+28.5***	-0.15	+28.4***	-0.14	+1.0	0.46***	-19.9***
Vessel 5	+4.9***	-0.24*	+3.1**	-0.18	-9.3***	0.52***	-16.3***
Overall ρ		-0.101		-0.098		0.523	

275 4. Discussion

276 4.1. *Model-based vs binary mask based diameter measurement*

277 The model-based and binary mask based diameter methods define vessel boundary in fundamen-
278 tally different ways. Binarisation of the en face maps discards the spatial intensity distribution
279 and reduces each pixel to a binary decision. The measured diameter directly depends on the
280 chosen threshold value and on angiogram intensity, both of which may vary across imaging
281 sessions, contrast conditions, and even between vessels within the same scan. Model-based
282 methods fit a parametric model function to the maximum or mean intensity profiles, and calculate
283 the diameter from the fitted parameters. As the fit captures the shape of the profile rather than
284 the absolute intensity, it is inherently less sensitive to changes in angiogram signal and does not
285 require a varying threshold.

286 This difference explains the trends observed in Fig. 7. Model-based methods maintained
287 strong linearity with the ground truth across all contrast conditions, whereas the binary mask
288 based slopes degraded progressively with increasing threshold. Pre-contrast, the hourglass
289 artefact caused by RBC orientation within vessel lumen suppresses edge signals, causing all
290 methods to underestimate diameter [15]. However, model-based methods still maintained a
291 linear relationship with the ground truth, producing systematic, predictable underestimation.
292 Binary mask methods on the other hand showed much weaker linearity, potentially because the
293 suppressed edge signals did not consistently exceed the binarisation threshold, depending on
294 vessel size and local intensity.

295 The generalised Gaussian (GG) model outperformed the standard Gaussian because the
296 intensity profiles of the vessels are usually not purely Gaussian (Fig. 1b). The standard Gaussian
297 cannot capture relatively flat-topped angiogram intensity profiles and tend to fit a narrower
298 curve, whereas the GG model accommodates through the shape parameter β , which ranges
299 from Gaussian ($\beta = 2$) to increasingly flat-topped profiles ($\beta > 2$). Among the width metrics,
300 FWHM_{GG} consistently showed the weakest accuracy. It could be because the half-maximum
301 determines the vessel boundary where the intensity has already dropped inside the true edge.
302 The $1/e^2$ width captures more of the transition region and so may have provided a more accurate
303 measure of the vessel diameter.

304 4.2. *Maximum vs mean intensity projection*

305 Across nearly all methods and conditions, MeIP outperformed MIP in both accuracy and
306 precision (Table 1). This finding is noteworthy because MIP has been shown to produce higher
307 signal-to-noise ratio (SNR) and contrast in en face OCTA images [12], and is the default in many
308 commercial OCTA systems and analysis pipelines [11, 13]. However, we observed that higher
309 SNR did not necessarily translate to more accurate diameter measurements. The MIP selects the
310 single highest intensity value along each A-scan within the vessel ROI, which makes it inherently
311 sensitive to outliers. MeIP averages all intensity values along each A-scan, which smooths out
312 intensity fluctuations and provides a more average profile. This averaging could be beneficial for
313 vessel diameter measurement, as the goal is to capture the typical vessel boundary rather than
314 extreme signal events.

315 The difference between MIP and MeIP was especially apparent in the binary mask based
316 results (Table 1). For MeIP, the slope declined more gradually and remained above 0.86 even at
317 BW-5 post-contrast. Whereas for MIP, the slope decreased steeply with increasing threshold,
318 dropping from 0.92 at BW-1 to 0.54 at BW-5 post-contrast. These findings suggest that MeIP
319 produces a more uniform intensity distribution across the vessel profile, making the binarisation
320 result less sensitive to the exact threshold value. The CoV data (Table 1) reinforced this as the
321 MIP-based binary mask CoV increased substantially with threshold, whereas the MeIP-based
322 CoV remained relatively stable. For model-based methods, MeIP also showed better performance.

323 The $w_{1/e^2, G}$ using MIP slightly overestimated vessel diameter post-contrast with $m = 1.12$, while
 324 MeIP reduced this to $m = 1.04$, a much closer agreement with the ground truth. These results
 325 suggest that while MIP may be preferred for visualisation purposes where contrast and SNR
 326 are important [12], MeIP should be the preferred projection method for quantitative diameter
 327 measurements.

328 4.3. Best choices under different conditions

329 Under the conditions where a contrast agent is available and accuracy is the primary goal, the
 330 $w_{1/e^2, GG}$ metric using MeIP provides the best results post-contrast. This combination achieved
 331 a slope of $m = 1.001$ with $R^2 = 0.97$ and a CoV of 0.09, yielding the best accuracy, linearity,
 332 and precision of all tested methods. If only pre-contrast data is available, the same metric
 333 still provides the best linearity ($R^2 = 0.88$) among all methods, though the slope of $m = 0.71$
 334 indicates that diameter could be underestimated by approximately 30%. In this case, a correction
 335 factor could be applied if absolute diameter values are needed, though relative comparisons
 336 between vessels or timepoints would remain valid without correction.

337 When only MIP data is available, which is the case for many commercial OCTA machines and
 338 existing datasets, the $w_{1/e^2, GG}$ still provides good post-contrast accuracy ($m = 1.02$, $R^2 = 0.93$).
 339 For binary mask based analysis of MIP data, only the lowest threshold (BW-1) post-contrast
 340 achieved a slope near unity ($m = 0.92$), and the slope degraded rapidly with increasing threshold
 341 values. This highlights that if binarisation must be used with MIP data, a lower threshold value is
 342 essential, though this may come at the cost of increased noise sensitivity.

343 If computation time is a limiting factor and model fitting is not practical, MeIP with a moderate
 344 threshold (e.g. BW-3) provides a reasonable alternative. At post-contrast, BW-3 using MeIP
 345 achieved $m = 0.97$ and $R^2 = 0.80$, which is acceptable for many applications. MeIP-based
 346 binary mask based analysis showed more stable CoV values across all thresholds compared to
 347 MIP, making the results less sensitive to the exact threshold choice. This stability makes MeIP
 348 the preferred projection for binary mask based OCTA.

349 4.4. Influence of angiogram intensity on measurements

350 The functional flicker stimulation experiment provided a direct demonstration of how changes
 351 in angiogram signal intensity can affect diameter measurements based on the method used.
 352 During the DyC-OCT acquisition, the angiogram intensity declined significantly across all five
 353 vessels (Fig. 8g,h), with decreases ranging from -14.1% to -19.9% (Table 2). This decline is
 354 expected during prolonged imaging sessions (any reference?). Simultaneously, the physiological
 355 response to flicker stimulation produces vasodilation, which has been well established in retinal
 356 physiology [16, 17].

357 The model-based methods correctly detected this expected vasodilation, with the Gaussian
 358 model showing significant diameter increases in four of five vessels (up to $+28.5\%$) and the
 359 GG model showing significant increases in three of five vessels (Table 2). The binary mask
 360 based method, however, reported significant diameter decreases in four of five vessels, suggesting
 361 vasoconstriction which is a paradoxical result. The reason behind this behaviour could be
 362 understood by observing how each method responds to declining signal intensity (Fig. 9). For
 363 the binary mask based method, the threshold is fixed and diameter is determined by the number
 364 of pixels that exceed this threshold. When the overall angiogram intensity drops, fewer pixels at
 365 the vessel edges exceed the threshold, and the vessel appears to shrink even if the physical vessel
 366 diameter has increased. This is confirmed by the strong positive correlation between measured
 367 diameter and angiogram intensity for the binary mask based method ($\rho = 0.523$, $p < 0.0005$). In
 368 contrast, the model-based methods showed weak overall correlations with intensity ($\rho = -0.101$
 369 for Gaussian, $\rho = -0.098$ for GG), indicating that the fitted diameter was effectively independent
 370 of the signal intensity.




Fig. 9. Measured vessel diameter as a function of angiogram signal intensity during functional flicker stimulation for (a) Gaussian model, (b) generalised Gaussian model, and (c) binary mask based method. Each colour represents one of five vessels. ρ denotes the Pearson correlation coefficient.

These findings have important implications for longitudinal studies and any experiment where angiogram intensity may vary between or within imaging sessions. In longitudinal mouse eye imaging, intensity losses can occur due to cataract formation, corneal drying, or changes in anaesthesia depth ([find references](#)). If a binary mask based method is used under such conditions, observed changes in vessel diameter may reflect intensity fluctuations rather than true vascular changes. In contrast, model-based methods may be less sensitive to these intensity variations, as they characterise the shape of the intensity profile rather than relying primarily on absolute intensity values to define vessel boundaries.

4.5. Literature Comparison

Hormel et al. [12] showed that MIP produces en face OCTA images with higher SNR and contrast compared to MeIP, and concluded that MIP should be preferred for OCTA visualisation. Our results are consistent with this for image quality purposes, but reveal a different picture when it comes to quantitative vessel diameter measurement. We found that MeIP consistently yielded more accurate and precise diameter values than MIP across both model-based and binary mask based approaches. This suggests that the projection method best suited for visualisation is not necessarily the best suited for quantification, and that the choice of projection should depend on the intended analysis.

The challenge of standardising OCTA image analysis has been highlighted by Sampson et al. [13] and Untracht et al. [11], both of whom emphasised the need for transparent, reproducible analysis pipelines. Our findings add to this discussion by showing that the choice of diameter measurement method introduces a source of variability that is at least as significant as the choice of binarisation threshold. The model-based approach presented here offers a path toward more standardised vessel diameter quantification, as it reduces the dependence on selected thresholds and angiogram intensity, and provides measurements that are more robust across imaging conditions.

The RBC orientation artefact and its effect on OCTA vessel appearance has been described by Bernucci et al. [15] and Cimalla et al. [6], who used Intralipid contrast to demonstrate that the hourglass pattern in vessel cross-sections is caused by the directional scattering properties of RBCs rather than by true variations in flow. Our work extends this observation by quantifying the degree to which this artefact affects diameter measurements and by proposing model-based fitting as a practical correction strategy. While contrast enhancement with Intralipid is effective for revealing the true vessel diameter, it requires a contrast injection and is not yet applicable in clinical settings. Using a calibrated model-based approach provides a way to partially recover the underestimated diameter from native OCTA data without the need for a contrast agent.

Recent work by Son et al. [18] using functional OCT to image flicker-evoked vasodilation

405 in the mouse retina showed that vasodilation is anisotropic, with larger changes in the axial
406 than the lateral direction. Our flicker stimulation results complement this by demonstrating that
407 the choice of diameter measurement method can determine whether the vascular response is
408 correctly detected at all. The fact that binary mask based methods reported vasoconstriction
409 where model-based methods correctly detected vasodilation underscores the importance of
410 measurement methodology when studying neurovascular coupling and other functional vascular
411 responses.

412 4.6. *Future Directions*

413 A future direction of this work is to move toward fully automated segmentation and calculation of
414 vessel diameters from en face angiogram maps. While ROIs were manually selected in this work,
415 it is also possible to automatically segment angiographic data using traditional skeletonisation
416 methods, which are commonly used for automated OCTA analysis. From a skeletonised image,
417 vessel angle can be computed for further extraction of vascular cross-sections perpendicular to
418 the vessel axis. These cross-sections could then be processed using the model-based methods
419 described here to enable fully automated analysis of 3D OCTA data sets, rather than the semi-
420 automated approach shown here for 2D OCTA data. This type of model-based approach is
421 particularly important for comparing 3D volumes acquired at different time points in the mouse
422 eye, as there are often intensity losses caused by drying of the eye, cataract formation, etc. As we
423 have shown in this work, even low levels of intensity loss occurring over a short time window
424 can significantly affect the measured vessel diameter when using a binarisation approach, but not
425 when using the models presented here.

426 Additionally, this work focused on the larger vessels of the superficial vascular plexus.
427 Extending the model-based approach to the intermediate and deep capillary plexuses, where
428 vessels are smaller and signal levels are lower, would be valuable for studying microvascular
429 changes in disease. For very small vessels approaching the lateral resolution limit of the imaging
430 system, the measured intensity profile is a convolution of the true vessel profile with the point
431 spread function (PSF) of the incident beam on the retina. Since the lateral resolution of OCT is
432 typically worse than the axial resolution — being determined by the beam optics rather than the
433 source bandwidth — this convolution becomes increasingly significant for smaller vessels. If the
434 PSF were known, a deconvolution approach could be used to recover the true vessel profile before
435 model fitting, potentially improving accuracy for capillary-scale vessels. However, characterising
436 the retinal PSF in vivo remains challenging, as it depends on the ocular optics of each individual
437 eye and may vary with retinal depth and eccentricity.

438 4.7. *Computation Time*

439 It is clear from the results that model-based approaches outperform traditional diameter measure-
440 ments based on a binarised image. Nevertheless, the improved linearity and repeatability come at
441 the cost of computation time. Fitting a model to the raw angiogram data is significantly more
442 computationally expensive and may not always be the most appropriate choice if results of the
443 processing are time sensitive. Nevertheless, these approaches are highly parallelisable given
444 the independent nature of the measurements, so it is expected that computation time could be
445 significantly accelerated with appropriate hardware.

446 4.8. *Thresholding*

447 The results presented here used a basic, global threshold for the binarisation of the angiogram
448 data. There are numerous other approaches for binarising the angiogram, which will most likely
449 each behave differently. For example, angiogram signals can be enhanced with vesselness filters
450 and dynamic thresholds can be used to binarise images with varying SNR. When applying a
451 vesselness filter, it is possible to enhance the shape of the angiogram and omit noise, but it is

also possible for such enhancement to overcompensate the signal, leading to larger than expected vessel appearance. Dynamic thresholding can also yield a more uniform angiogram appearance, but similarly risks over- or undercompensating vessels given that threshold level directly affects the measured diameter. Further analysis of these other methods would be required to determine their efficacy in returning accurate and precise diameter values.

References

1. A. Javed, A. Khanna, E. Palmer, *et al.*, "Optical coherence tomography angiography: A review of the current literature," *J. Int. Med. Res.* **51**, 03000605231187933 (2023). Doi: 10.1177/03000605231187933.
2. S. Fialová, M. Augustin, M. Glösmann, *et al.*, "Polarization properties of single layers in the posterior eyes of mice and rats investigated using high resolution polarization sensitive optical coherence tomography," *Biomed. Opt. Express* **7**, 1479–1495 (2016).
3. C. W. Merkle, M. Augustin, D. J. Harper, *et al.*, "High-resolution, depth-resolved vascular leakage measurements using contrast-enhanced, correlation-gated optical coherence tomography in mice," *Biomed. Opt. Express* **12**, 1774–1791 (2021).
4. Y. Fu and K.-W. Yau, "Phototransduction in mouse rods and cones," *Pflügers Arch. J. Physiol.* **454**, 805–819 (2007).
5. Y. Pan, J. You, N. D. Volkow, *et al.*, "Ultrasensitive detection of 3D cerebral microvascular network dynamics in vivo," *NeuroImage* **103**, 492–501 (2014).
6. P. Cimalla, J. Walther, M. Mittasch, and E. Koch, "Shear flow-induced optical inhomogeneity of blood assessed in vivo and in vitro by spectral domain optical coherence tomography in the 1.3 μm wavelength range," *J. Biomed. Opt.* **16**, 116020 (2011).
7. C. W. Merkle and V. J. Srinivasan, "Laminar microvascular transit time distribution in the mouse somatosensory cortex revealed by Dynamic Contrast Optical Coherence Tomography," *NeuroImage* **125**, 350–362 (2016).
8. M. Augustin, S. Fialová, R. Plasenzotti, *et al.*, "Multi-functional OCT image processing for rodent eyes," in *Optical Tomography and Spectroscopy*, (OSA - The Optical Society, Center for Medical Physics and Biomedical Engineering, Medical University of Vienna, Waehringer Guertel 18-20, 4L, Vienna, A-1090, Austria, 2016), p. 3. Export Date: 10 June 2020 Correspondence Address: Augustin, M.; Center for Medical Physics and Biomedical Engineering, Medical University of Vienna, Waehringer Guertel 18-20, 4L, Austria; email: marco.augustin@meduniwien.ac.at.
9. V. J. Srinivasan, J. Y. Jiang, M. A. Yaseen, *et al.*, "Rapid volumetric angiography of cortical microvasculature with optical coherence tomography," *Opt. Lett.* **35**, 43–45 (2010).
10. D. J. Harper, M. Augustin, A. Lichtenegger, *et al.*, "Retinal analysis of a mouse model of Alzheimer's disease with multicontrast optical coherence tomography," *Neurophotonics* **7**, 015006 (2020). Export Date: 4 June 2020 Correspondence Address: Harper, D.J.; Medical University of Vienna, Center for Medical Physics and Biomedical Engineering Austria; email: danielle.harper@meduniwien.ac.at.
11. G. R. Untracht, M. S. Durkee, M. Zhao, *et al.*, "Towards standardising retinal oct angiography image analysis with open-source toolbox octava," *Sci. Reports* **14** (2024).
12. T. T. Hormel, J. Wang, S. T. Bailey, *et al.*, "Maximum value projection produces better en face oct angiograms than mean value projection," *Biomed. Opt. Express* **9**, 6412 (2018).
13. D. M. Sampson, A. M. Dubis, F. K. Chen, *et al.*, "Towards standardizing retinal optical coherence tomography angiography: a review," *Light. Sci. & Appl.* **11** (2022).
14. C. S. Lee, A. J. Tying, Y. Wu, *et al.*, "Generating retinal flow maps from structural optical coherence tomography with artificial intelligence," *Sci. Reports* **9** (2019).
15. M. T. Bernucci, C. W. Merkle, and V. J. Srinivasan, "Investigation of artifacts in retinal and choroidal OCT angiography with a contrast agent," *Biomed. Opt. Express* **9**, 1020–1040 (2018).
16. G. Garhöfer, C. Zawinka, H. Resch, *et al.*, "Diffuse luminance flicker increases blood flow in major retinal arteries and veins," *Vis. Res.* **44**, 833–838 (2004).
17. M. Sharifzad, K. J. Witkowska, G. C. Aschinger, *et al.*, "Factors determining flicker-induced retinal vasodilation in healthy subjects," *Investig. Ophthalmol. Vis. Sci.* **57**, 3306–3312 (2016). Cited By :12 Export Date: 10 June 2020 CODEN: IOVSD Correspondence Address: Schmetterer, L.; Department of Clinical Pharmacology, Medical University of Vienna, Waehringer Gürtel 18-20, Austria; email: Leopold.schmetterer@meduniwien.ac.at.
18. T. Son, G. Ma, and X. Yao, "Functional OCT reveals anisotropic changes of retinal flicker-evoked vasodilation," *Opt. Lett.* **49**, 2121–2124 (2024).