



GOVERNMENT OF INDIA

DEPARTMENT OF CONSUMER AFFAIRS - LEGAL METROLOGY DIVISION

Verification Certificate under Standards of Weights & Measures (OIML R-76)

Certificate ID: LM-GOI-202609-D0FF

Date & Time: 06-Sep-2026 11:19

INSPECTION & CALIBRATION RECORD

Entity / Establishment Name	DSD
Inspecting Officer ID	LMO-101
Standard Working Mass Applied	1000.0 g
Instrument Reading Output	999.0 g
Calculated Error (E)	1.0 g
Max Permissible Error (MPE)	+/- 2.0 g
Verification Verdict	PASSED & DIGITALLY STAMPED



DIGITALLY VERIFIED AND SEALED

This equipment complies with statutory provisions of Legal Metrology Act, 2009. The non-automatic weighing instrument is certified for commercial trade usage. Scan the QR code to authenticate verification records on the national portal.

Authorised Legal Metrology Officer

Govt. of India, Inspection Directorate