

### Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	31-03-2025 13:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	31-03-2025 13:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	120 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Power
Department Name/विभाग का नाम	Na
Organisation Name/संगठन का नाम	Ntpc Limited
Office Name/कार्यालय का नाम	Ussc Cpg2 Sipat
क्रेता ईमेल/Buyer Email	bycon10.ntpc.solapur@gembuyer.in
Total Quantity/कुल मात्रा	112000
Item Category/मद केटेगरी	Sterile Hypodermic Syringe (V3) (Q2) , Scalp Vein Infusion Set (Q2)
MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट	No
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Certificate (Requested in ATC), OEM Authorization Certificate, Additional Doc 4 (Requested in ATC), Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Do you want to show documents uploaded by bidders to all bidders participated in bid?/	No
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid
Primary product category	Sterile Hypodermic Syringe (V3)
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days

Bid Details/बिड विवरण	
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation
Arbitration Clause	No
Mediation Clause	No

#### EMD Detail/ईएमडी विवरण

Required/आवश्यकता	No
-------------------	----

#### ePBG Detail/ईपीबीजी विवरण

Required/आवश्यकता	No
-------------------	----

#### MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	No
--	----

#### MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM\\_No.1\\_4\\_2021\\_PPD\\_dated\\_18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

#### Sterile Hypodermic Syringe (V3) ( 30000 pieces )

#### Technical Specifications/तकनीकी विशिष्टियाँ

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GENERAL	Product Description	Hypodermic Syringe For Manual Use
	Sterility	Sterile
	Usage	Single Use
PRODUCT INFORMATION	<b>Type of Syringe</b>	Hypodermic Syringe with Needle
	<b>Nominal Capacity of Syringe (ml)</b>	2 mL
	Tolerance, dead space, scale dimensions shall be as per Table 1 of IS 10258 Part 1 (Latest)	Yes
	Syringe type based on number of pieces	Three Piece Syringe
	Graduation Markings on the syringe	Yes
	Needle if provided conforming to Indian Standard (IS 10654 / ISO 7864)	Yes
	Needle cap or shield to protect the needle	Yes
	Sharp Protection Feature	Yes
	Syringe Tip	Leur Slip
MATERIAL	Material of Needle Tube (As per ISO 9626)	Stainless Steel
PACKING	Packaging & Marking as per IS 10258 Part 1 (Latest)	Yes
	Type of Packing	Individually packed in moisture proof packing
CERTIFICATIONS & REPORTS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid medical device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Manufacturing unit certification	ISO:13485 (Latest)
	Availability of Test Report for each supplied batch/product as per Medical Device Rules (MDR) 2017 as amended till date	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission and/or along with supplies as per buyer requirement	Yes
SHELF LIFE	Shelf life in months from the date of manufacture	36 Or higher (month)
	Minimum shelf life of the product at the time of delivery to the consignee	3/4 th of Total Shelf Life
ADVANCE SAMPLE	Agree to provide advance sample of the product for buyer's approval after award of contract and before commencement of supply in case of bidding only	Yes

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Jayaraju Pagi	505215,GSTIN: 36AAACN0255D1ZZ NTPC Stores Ramagundam Super Thermal Power Station PO JYOTINAGAR 505215 DISTRICT PEDDAPALLI	30000	30

**Sterile Hypodermic Syringe (V3) ( 30000 pieces )**

**Technical Specifications/तकनीकी विशिष्टियाँ**

\* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GENERAL	Product Description	Hypodermic Syringe For Manual Use
	Sterility	Sterile
	Usage	Single Use

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
PRODUCT INFORMATION	Type of Syringe	Hypodermic Syringe with Needle
	Nominal Capacity of Syringe (ml)	2 mL, 5 mL
	Tolerance, dead space, scale dimensions shall be as per Table 1 of IS 10258 Part 1 (Latest)	Yes
	Syringe type based on number of pieces	Three Piece Syringe
	Graduation Markings on the syringe	Yes
	Needle if provided conforming to Indian Standard (IS 10654 / ISO 7864)	Yes
	Needle cap or shield to protect the needle	Yes
	Sharp Protection Feature	Yes
	Syringe Tip	Leur Lock, Leur Slip
MATERIAL	Material of Needle Tube (As per ISO 9626)	Stainless Steel
PACKING	Packaging & Marking as per IS 10258 Part 1 (Latest)	Yes
	Type of Packing	Individually packed in moisture proof packing
CERTIFICATIONS & REPORTS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid medical device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Manufacturing unit certification	ISO:13485 (Latest)
	Availability of Test Report for each supplied batch/product as per Medical Device Rules (MDR) 2017 as amended till date	Yes
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission and/or along with supplies as per buyer requirement	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
SHELF LIFE	Shelf life in months from the date of manufacture	36 Or higher (month)
	Minimum shelf life of the product at the time of delivery to the consignee	3/4 th of Total Shelf Life
ADVANCE SAMPLE	Agree to provide advance sample of the product for buyer's approval after award of contract and before commencement of supply in case of bidding only	Yes

#### Additional Specification Parameters - Sterile Hypodermic Syringe (V3) ( 30000 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)
1.PRODUCT INFORMATION- Nominal Capacity of Syringe (ml)	Must be consider as 5ML only. instead of 2 mL, 5 mL.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

#### Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Jayaraju Pagi	505215,GSTIN: 36AAACN0255D1ZZ NTPC Stores Ramagundam Super Thermal Power Station PO JYOTINAGAR 505215 DISTRICT PEDDAPALLI	30000	30

#### Sterile Hypodermic Syringe (V3) ( 30000 pieces )

#### Technical Specifications/तकनीकी विशिष्टियाँ

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GENERAL	Product Description	Hypodermic Syringe For Manual Use

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Sterility	Sterile
	Usage	Single Use
CONFORMITY TO STANDARD	Conformity to Specifications	IS 10258 : Part 1 (Latest)
PRODUCT INFORMATION	<b>Type of Syringe</b>	Hypodermic Syringe with Needle
	<b>Nominal Capacity of Syringe (ml)</b>	3 mL
	Tolerance, dead space, scale dimensions shall be as per Table 1 of IS 10258 Part 1 (Latest)	Yes
	Syringe type based on number of pieces	Three Piece Syringe
	Graduation Markings on the syringe	Yes
	Needle if provided conforming to Indian Standard (IS 10654 / ISO 7864)	Yes
	Needle cap or shield to protect the needle	Yes
	Sharp Protection Feature	No
	Syringe Tip	Leur Slip
MATERIAL	Material of Needle Tube (As per ISO 9626)	Stainless Steel
PACKING	Packaging & Marking as per IS 10258 Part 1 (Latest)	Yes
	Type of Packing	Individually packed in moisture proof packing
CERTIFICATIONS & REPORTS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid medical device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Manufacturing unit certification	ISO:13485 (Latest)
	Availability of Test Report for each supplied batch/product as per Medical Device Rules (MDR) 2017 as amended till date	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission and/or along with supplies as per buyer requirement	Yes
SHELF LIFE	Shelf life in months from the date of manufacture	36 Or higher (month)
	Minimum shelf life of the product at the time of delivery to the consignee	3/4 th of Total Shelf Life
ADVANCE SAMPLE	Agree to provide advance sample of the product for buyer's approval after award of contract and before commencement of supply in case of bidding only	No, Yes

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Jayaraju Pagi	505215,GSTIN: 36AAACN0255D1ZZ NTPC Stores Ramagundam Super Thermal Power Station PO JYOTINAGAR 505215 DISTRICT PEDDAPALLI	30000	30

**Sterile Hypodermic Syringe (V3) ( 20000 pieces )**

**Technical Specifications/तकनीकी विशिष्टियाँ**

\* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GENERAL	Product Description	Hypodermic Syringe For Manual Use
	Sterility	Sterile
	Usage	Single Use



Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
PRODUCT INFORMATION	Type of Syringe	Hypodermic Syringe with Needle
	Nominal Capacity of Syringe (ml)	10 mL
	Tolerance, dead space, scale dimensions shall be as per Table 1 of IS 10258 Part 1 (Latest)	Yes
	Needle if provided conforming to Indian Standard (IS 10654 / ISO 7864)	Yes
	Needle cap or shield to protect the needle	Yes
	Sharp Protection Feature	No
	Syringe Tip	Leur Slip
MATERIAL	Material of Needle Tube (As per ISO 9626)	Stainless Steel
PACKING	Packaging & Marking as per IS 10258 Part 1 (Latest)	Yes
	Type of Packing	Individually packed in moisture proof packing
CERTIFICATIONS & REPORTS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid medical device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Manufacturing unit certification	ISO:13485 (Latest)
	Availability of Test Report for each supplied batch/product as per Medical Device Rules (MDR) 2017 as amended till date	Yes
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission and/or along with supplies as per buyer requirement	Yes
SHELF LIFE	Shelf life in months from the date of manufacture	36 Or higher <b>(month)</b>
	Minimum shelf life of the product at the time of delivery to the consignee	3/4 th of Total Shelf Life

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
ADVANCE SAMPLE	Agree to provide advance sample of the product for buyer's approval after award of contract and before commencement of supply in case of bidding only	No, Yes

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Jayaraju Pagi	505215,GSTIN: 36AAACN0255D1ZZ NTPC Stores Ramagundam Super Thermal Power Station PO JYOTINAGAR 505215 DISTRICT PEDDAPALLI	20000	30

### Scalp Vein Infusion Set ( 2000 pieces )

#### Technical Specifications/तकनीकी विशिष्टियाँ

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GENERAL	Product Description	Scalp Vein (Winged Needle) Infusion Set
	Sterility	Sterile
	Usage	Single-Use
STANDARD	Conformity to Standard	IS 16097 (Latest Revision)
PRODUCT SPECIFICATION	Lubricated Needle Tube	Yes
	Flexible kink resistant extension tube	Yes
	Separate protective needle sheath provided	Yes
	Color coded according to size of needle as per Table 1 of IS 16097	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Wing Material	Pigmented
	Needle Sheath Material	Unpigmented
	<b>Nominal outside diameter of needle tube</b>	1.2 mm
	Needle Length	19 mm
	Connection Type	Luer Slip
	Extension Tube Length (±5%)	200 mm
PACKAGING & MARKING	Labeling and packing shall be as per IS 16097 (Latest Revision)	Yes
CERTIFICATIONS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid medical device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Manufacturing unit certification	ISO:13485 (Latest)
	Availability of Test Report for each supplied batch/product as per Medical Device Rules (MDR) 2017 as amended till date	Yes
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission and/or along with supplies as per buyer requirement	Yes
SHELF LIFE	Shelf life in months from the date of manufacture	36 Or higher <b>(month)</b>
	Minimum shelf life of the product at the time of delivery to the consignee	3/4 th of Total Shelf Life
ADVANCE SAMPLE	Agree to provide advance sample of the product for buyer's approval before commencement of supply in case of bidding	Yes

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Jayaraju Pagi	505215,GSTIN: 36AAACN0255D1ZZ NTPC Stores Ramagundam Super Thermal Power Station PO JYOTINAGAR 505215 DISTRICT PEDDAPALLI	2000	30

**Special terms and conditions-Version:1 effective from 21-06-2024 for category Sterile Hypodermic Syringe (V3)**

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on the self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.

**Special terms and conditions-Version:1 effective from 15-03-2024 for category Scalp Vein Infusion Set**

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on the submission of valid Drug License and self declaration of product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of drug license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any

reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.

5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.

## **Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें**

### **1. Generic**

Bidder shall submit the following documents along with their bid for Vendor Code Creation:

- a. Copy of PAN Card.
- b. Copy of GSTIN.
- c. Copy of Cancelled Cheque.
- d. Copy of EFT Mandate duly certified by Bank.

### **2. Generic**

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

### **3. Scope of Supply**

Scope of supply (Bid price to include all cost components) : Only supply of Goods

### **4. Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

1. Bidder shall submit the following declarations :- "The sellers who participate in current Tender are deemed to have fully accepted all Terms & Conditions, Technical Specifications, Data Sheets, Drawings, Quality Plans, Buyer Added Bid Specific Terms and Conditions given in Bidding Documents, read in conjunction with Amendment(s)/ Clarification(s)/ Addenda/ Errata (if any) issued by the Buyer prior to opening of Techno-Commercial Bids, and the same are supposed to have been taken into consideration while making their Techno-Commercial Bid & Price Bid. Further it is construed that no deviation has been taken by the seller in this regard. In case any documents/ details/ catalogues are found to have been submitted by the bidder along with its bid, which is in contravention to buyer's bidding documents, those will not be taken into cognizance and will stand automatically withdrawn."

2. AS PER CLAUSE NO-8 DUTIES & TAXES OF GEM GTC, BIDDERS SHALL QUOTE ALL INCLUSIVE PRICES (i.e. INCLUSIVE OF GST, P&F, FREIGHT, INSURANCE ETC.).

**All post award communication/ grievance will be addressed to consignee at email id:- [store\\_inc\\_harge-ramagundam@ntpc.co.in](mailto:store_inc_harge-ramagundam@ntpc.co.in) In case of non-resolution of the issue, matter can be escalated along with previous correspondence(s) to email id :- [vendor\\_grievance\\_cpg2@ntpc.co.in](mailto:vendor_grievance_cpg2@ntpc.co.in)**

### **Consignee details :-**

-

Consignee details :-

Contact person :- Jayaraju Pagi

Email:- pjayaraju@ntpc.co.in

Contact person Mob no.- 9440918715

.

### **5. Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

## **Disclaimer/अस्वीकरण**

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

**All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The**

**Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.**

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।

**---Thank You/धन्यवाद---**