

**Bid Document/ बिड दस्तावेज़**

Bid Details/बिड विवरण	
<b>Bid End Date/Time/बिड बंद होने की तारीख/समय</b>	24-03-2025 11:00:00
<b>Bid Opening Date/Time/बिड खुलने की तारीख/समय</b>	24-03-2025 11:30:00
<b>Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)</b>	120 (Days)
<b>Ministry/State Name/मंत्रालय/राज्य का नाम</b>	Ministry Of Power
<b>Department Name/विभाग का नाम</b>	Na
<b>Organisation Name/संगठन का नाम</b>	Ntpc Limited
<b>Office Name/कार्यालय का नाम</b>	Ussc Cpg2 Sipat
<b>क्रेता ईमेल/Buyer Email</b>	bycon10.ntpc.solapur@gembuyer.in
<b>Total Quantity/कुल मात्रा</b>	1
<b>Item Category/मद केटेगरी</b>	Cardiac TMT Machine (V2) (Q2)
<b>MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट</b>	No
<b>Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट</b>	No
<b>Document required from seller/विक्रेता से मांगे गए दस्तावेज़</b>	Certificate (Requested in ATC), OEM Authorization Certificate, Additional Doc 4 (Requested in ATC), Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
<b>Do you want to show documents uploaded by bidders to all bidders participated in bid?/</b>	No
<b>Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया</b>	No
<b>Type of Bid/बिड का प्रकार</b>	Two Packet Bid
<b>Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय</b>	2 Days

Bid Details/बिड विवरण	
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation
Arbitration Clause	No
Mediation Clause	No

#### EMD Detail/ईएमडी विवरण

Required/आवश्यकता	No
-------------------	----

#### ePBG Detail/ईपीबीजी विवरण

Required/आवश्यकता	No
-------------------	----

#### MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	No
--	----

#### MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 100% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM\\_No.1\\_4\\_2021\\_PPD\\_dated\\_18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

2. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

**Cardiac TMT Machine (V2) ( 1 pieces )****Technical Specifications/तकनीकी विशिष्टियाँ**

\* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GENERAL FEATURES	Product Description	Cardiac TMT Machine with Monitor and Printer
	Product components	Heavy-duty medical treadmill with online UPS for the whole system, Computer system, printer, Voltage stabilizer and software along with trolley system to mount computer & printer
	Facility, whereby data should be able to be acquired and digitized near the patient, providing better ECG signal	Yes
TREADMILL SPECIFICATIONS	Treadmill Type	Four tier heavy duty noise less medical treadmill with coasters having locking arrangement
	Compatible online UPS for the whole system with minimum 3 hrs back up time	Yes
	Stop/start button on treadmill for emergency stop	Yes
	Emergency stop button (ESB) must have the ability to be located in the location of choice by the end user	Yes
	Treadmill should have zero start	Yes
	Automatic operation from the main unit as per test protocol	Yes
	Fold down hand rails on treadmill for allowing easy and convenient assemble and installation	Yes
	Length of Treadmill Floor (mm)	2001 to 2100mm
	Width of Treadmill Floor (mm)	801 to 900mm
	Length of walking area on treadmill (mm)	1500 to 1600mm
	Width of walking area on treadmill (mm)	500 to 600mm
	Maximum Patient Load Capacity of Treadmill (Kg)	200 Kg or more Or higher
	Speed range of Treadmill	0.1 to 14 km/Hr or higher range
	Inclination (continuously variable)	0 to 25 or more
	Treadmill motor	AC Motor

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Power requirements	220 - 240 VAC @ 50-60 Hz
	Double sided polish on walking surface of treadmill for prolonged product life span	Yes
	Remote keyboard operation to start, stop and increase or decrease elevation and speed of treadmill	Yes
	Self-aligning anti skid conveyor belt	Yes
	Enhanced tolerance to electrical changes for greater patient safety	Yes
	Self-calibrating system for speed	Yes
	Inbuilt digital control issues, precise speed and grade control in the system	Yes
	Delivers high torque even at low speeds which ensures no lag when the foot strikes the belt	Yes
	Provide smooth power and accuracy even at slow speeds	Yes
	Normal, slow deceleration and emergency stop modes on treadmill	Yes
	Compatible Voltage Stabilizer provided (Minimum 3KVA)	Yes
COMPUTER SYSTEM SPECIFICATIONS	<b>Touch Screen Monitor</b>	No
	Size of Monitor (inch)	24 Or higher ( <b>inch</b> )
	RAM size	4 GB, 6 GB, 8 GB Or higher
	Hard disk space	500 GB, 1 TB Or higher
	Processor	Intel core i5
	Operating System	Windows 10 or Higher
	CD ROM Drive	Yes
	Number of serial port	2
	Number of USB Ports	4
	Alphanumeric multimedia PC keyboard with systems dedicated keypad	Yes
	Standard PC mouse	Yes

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Built in uninterrupted power supply for minimum of 3 minutes	Yes
	QRS triggered output	Yes
	Parameters Displayed by the system	12 Lead raw ECG, 12 lead ST level & Slops, Enlarged QRS complex with maximum ST level, Exercise time, Target and Max Heart Rate with % of target achievement, Heart Rate & METS trends, NIBP Trends, Speed of treadmill, Protocol Name, Protocol stage, Patient info
	Built in thermal printer with compatible additional laser black and white printer to print report on A4 Size thermal and A4 size ordinary bond papers respectively	Yes
	Minimum Printing Speed of laser printer	20 pages / minute
	Minimum Paper tray capacity of laser printer	250 Sheets
	Print Resolution	1200 x 600 dpi
	Duplex Printing	No
	Connectivity / interface of Laser printer	USB
	Supported paper sizes	A4
TRANSMISSION OF DATA	Transmission of data available through	USB, PDF
	Advanced signal processing algorithms to provide exceptional accuracy in beat detection and noise rejection resulting in superior identification of events	Yes
	Unique source consistency filter, reducing noise and baseline artifact while preserving a diagnostic quality ECG signal	Yes
	Provision to eliminate artifacts due to respiration, muscle rub and AC interference, baseline wandering without compromising and distortion in ST segment changes	Yes
PROTOCOLS	Test Protocols include Bruce, Modified Bruce, Balke, Ellestad, Naughton protocols	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Facility to add user defined protocols	Yes
SYSTEM SOFTWARE FEATURES	Licensed copy of windows operating system	Yes
	Provision to record resting ECG & Exercise ECG	Yes
	Resting ECG software with detailed automatic interpretation statement for all 12 or 16 lead resting ECG should be present	Yes
	ECG data management software for storage and review	Yes
	Facility to review, edit and add ECG from full discloser storage post exam	Yes
	50mm sweep speed selection for ECG display in the system	Yes
	Enlarged QRS complexes with superimposition technique	Yes
	Automatic BP Measurement System integrated with the stress test system	Yes
	Provision of real time average complexes	Yes
	Provision of real time ST Analysis in bar graphs	Yes
	Provision of Re analysis of final summary report	Yes
	System should be able to present ST levels for 12 average complexes, slope for 12 average complexes and ST Profile	Yes
	Provision of online ST measurement adjustment	Yes
	Provision of reanalyzing the possibility of complete exercise test with ST/HR and NIBP trend graphically	Yes
	SCD predictive tools like T-wave alternans, ST /HR hysteresis analysis	Yes
	The system should have a Borg scale, symptom and point criteria table	Yes

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Full disclosure review/playback/scroll back and addition of past ECG events	Yes
	ST level and slope data for lead and worst case average beat update should be available continuously during the test	Yes
	Beat by beat review mode to allow full disclosure review during and post exam	Yes
	QRS signal averaging software, QT dispersion software, XYZ vector-cardiography software provided as standard part of the unit	Yes
	Facility to do reanalysis of scored ECG by changing line measurement point title ST and J	Yes
	The system should present comprehensive final report on minute by minute record of ST segment trend	Yes
	Facility of automatic arrhythmia detection	Yes
	Auto calibration with digital micro process control in the system and should also be manually operable	Yes
	Facility of pacemaker rhythm detection	Yes
	System should be able to view comprehensive ST segment and morphology analysis in an anatomically intelligent format	Yes
	ST profile, ST elevation, ST depression, heart rate, Mets, NIBP, ST index should be able to be clearly analyzed	Yes
	The unit should be able to quickly identify anomalies with the dynamic zoom ST display	Yes
	The system should be able to view ST maps, patented visual anatomical representations of ST deviations in-frontal planes	Yes
	Monitoring of ST changes with auto comparison of current and reference beats	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	The system should be able to display the traces incrementally	Yes
	Provision of updating ECG and ST segment changes and compare them on a pre-exercise and during exercise basis on a high resolution display	Yes
	Compact digital acquisition module that virtually eliminates non-cardiac electrical noises	Yes
	Provision of alarm on HR, BP & ST level	Yes
	Facility to mark the ECG strip to enter comments at any stage	Yes
DATA ACQUISITION AND PROCESSING	Data Acquisition	12 Leads ECG Simultaneously and Display 3x4, 4x6, 6x12, 12 lead data indicate significant changes
	A/D	12 bits or more
	Sampling Rate	300/Sec/channel or more
	Input impedance should be more than 90M Ohms	Yes
	CMRR Should be more than 100 dB	Yes
	Frequency response	0.05Hz to 100Hz
	Patient leakage Should be less than 100?A	Yes
	The system should have baseline correction	Yes
	Digital filters	50Hz, 20Hz, 35Hz
REPORTING	Facility to get system generated auto report	Yes
	Analysis should work on total of atleast 5 protocols	Yes
	No startup delay in the system ie instant ECG Reports at the push of a button should be available	Yes
	Final report with configurable lead groups	Yes



Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Patient report include protocol, Indications, medications, target heart rate reasons for end, symptoms, diagnosis, notes and conclusions with space for personnel and physician electronic signature	Yes
	Examination summary include heart rate/BP/Workload trends page, ST level trends/ST slope trends, Average QRS by stage or by minute, Mets, Maximum predicted heart rate, THR formula selection	Yes
	Ability to select final report segments including patient information, exam summary, rate, BP, work trends, ST level trends, ST slope trends, average QRS and events should be provided in the system	Yes
	Events should include 12 lead ECG analysis for arrhythmias and user added ECGs	Yes
	Provision of customizable system generated summary report to determine the contents of report with data storage facility	Yes
ACCESSORIES	Standard Accessories, consumables provided	2 sets of 12 lead ECG patient cables, 100 Electrodes, 10 rim A4 Reporting papers
	Detachable/replaceable cables and leads	Yes
	Minimum warranty(months) on accessories i e lead set	6, 12 Or higher <b>(year)</b>
	Appropriate sized cuffs of small, medium and large should be provided with the system	Yes
	Trolley system to mount computer & printer	Yes
STORAGE OF RECORDS	Storage of atleast 100 patients records	Yes
	Full disclosure archive to external drive	Yes
	Events/12 lead ECGs/Full disclosure	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
CERTIFICATION & REPORTS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid drug license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Manufacturing unit certification	ISO:13485 (Latest)
	Availability of Test Report for product as per Medical Device Rules (MDR) 2017 as amended till date	Yes
	Compliance to Safety Standards	IEC-60601-2-25 or equivalent BIS Standard
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission and/or along with supplies as per buyer requirement	Yes
WARRANTY	<b>Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)</b>	5 Or higher (year)
MISCELLANEOUS PARAMETERS	OEM/Reseller shall ensure uninterrupted availability of all spares for 10 years	Yes
	Availability of toll free facility for technical support maintained by OEM or authorized agencies	Yes
	User/Technical/Maintenance manuals to be supplied in English in hard and soft copy	Yes
	Details of equipments and procedures required for local calibration and routine maintenance to be supplied and advanced maintenance task documentation also to be furnished	Yes
	List of important spares and accessories, with their part numbers to be supplied to the buyer at the time of supplying the equipment	Yes

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Installation and Demonstration of equipment and training to be provided after completing supplies before acceptance	Yes
	Principal Manufacturer must have direct Presence/approved service center In India	Yes

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Jayaraju Pagi	505215,GSTIN: 36AAACN0255D1ZZ NTPC Stores Ramagundam Super Thermal Power Station PO JYOTINAGAR 505215 DISTRICT PEDDAPALLI	1	30

**Special terms and conditions-Version:1 effective from 06-10-2023 for category Cardiac TMT Machine (V2)**

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on the submission of valid Drug License and self declaration of product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of drug license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance

including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

## **Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें**

### **1. Generic**

Bidder shall submit the following documents along with their bid for Vendor Code Creation:

- a. Copy of PAN Card.
- b. Copy of GSTIN.
- c. Copy of Cancelled Cheque.

d. Copy of EFT Mandate duly certified by Bank.

## 2. **Generic**

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

## 3. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods

## 4. **Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

1. Bidder shall submit the following declarations :- “The sellers who participate in current Tender are deemed to have fully accepted all Terms & Conditions, Technical Specifications, Data Sheets, Drawings, Quality Plans, Buyer Added Bid Specific Terms and Conditions given in Bidding Documents, read in conjunction with Amendment(s)/ Clarification(s)/ Addenda/ Errata (if any) issued by the Buyer prior to opening of Techno-Commercial Bids, and the same are supposed to have been taken into consideration while making their Techno-Commercial Bid & Price Bid. Further it is construed that no deviation has been taken by the seller in this regard. In case any documents/ details/ catalogues are found to have been submitted by the bidder along with its bid, which is in contravention to buyer’s bidding documents, those will not be taken into cognizance and will stand automatically withdrawn.”

2. AS PER CLAUSE NO-8 DUTIES & TAXES OF GEM GTC, BIDDERS SHALL QUOTE ALL INCLUSIVE PRICES (i.e. INCLUSIVE OF GST, P&F, FREIGHT, INSURANCE ETC.).

***All post award communication/ grievance will be addressed to consignee at email id:- [store inc harge- telangana@ntpc.co.in](mailto:store_inc_harge-telangana@ntpc.co.in) In case of non-resolution of the issue, matter can be escalated along with previous correspondence(s) to email id :- [vendor\\_grievance\\_cpg2@ntpc.co.in](mailto:vendor_grievance_cpg2@ntpc.co.in)***

### **Consignee details :-**

-

Consignee details :-

Contact person:- Jayaraju Pagi

Email:- [pjayaraju@ntpc.co.in](mailto:pjayaraju@ntpc.co.in)

Contact person Mob no.- 9440918715

.

## 5. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

## Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

**All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.**

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।

---Thank You/धन्यवाद---

