

### Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	12-05-2025 20:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	12-05-2025 20:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Home Affairs
Department Name/विभाग का नाम	Central Armed Police Forces
Organisation Name/संगठन का नाम	Border Security Force (bsf)
Office Name/कार्यालय का नाम	Force Head Quarter Bsf New Delhi
क्रेता ईमेल/Buyer Email	comdtproc@bsf.nic.in
Total Quantity/कुल मात्रा	5
Item Category/मद केटेगरी	Laparoscope Surgery Set
GeMARPTS में खोजी गई स्ट्रिंग / Searched Strings used in GeMARPTS	Laparoscope Surgery Set
GeMARPTS में खोजा गया परिणाम / Searched Result generated in GeMARPTS	Veterinary Periodontal Instrument or Surgery Basic Kit (V2), Power Generator - DG Set (up to 900 KVA), Veterinary general surgery kit, Laparoscopic Surgery System (V2), Sofa Set (Steel Tube), High Frequency Radio Surgery Unit, Video Assisted Thoracoscopic Surgery Equipment, Television (TV) (V2), Teeth set, Desks and Bench/Chair set for Classroom/Training Area
अधिसूचना के लिए चयनित प्रासंगिक श्रेणियाँ / Relevant Categories selected for notification	<ul style="list-style-type: none"> <li>High Definition Laparoscopic Surgery System</li> </ul>
Minimum Average Annual Turnover of the bidder (For 3 Years)/बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का)	147 Lakh (s)
OEM Average Turnover (Last 3 Years)/मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)	294 Lakh (s)
Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष	3 Year (s)

Bid Details/बिड विवरण	
<b>MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है</b>	Yes
<b>Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है</b>	Yes
<b>Document required from seller/विक्रेता से मांगे गए दस्तावेज़</b>	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
<b>Do you want to show documents uploaded by bidders to all bidders participated in bid?/</b>	No
<b>Past Performance/विगत प्रदर्शन</b>	50 %
<b>Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया</b>	Yes
<b>RA Qualification Rule</b>	H1-Highest Priced Bid Elimination
<b>Comprehensive Maintenance Charges Required</b>	Yes
<b>Type of Bid/बिड का प्रकार</b>	Two Packet Bid
<b>Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय</b>	7 Days
<b>Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)</b>	No
<b>Evaluation Method/मूल्यांकन पद्धति</b>	Total value wise evaluation
<b>Financial Document Required/वित्तीय दस्तावेज़ की आवश्यकता है।</b>	Yes
<b>Arbitration Clause</b>	No
<b>Mediation Clause</b>	No

#### EMD Detail/ईएमडी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
EMD Amount/ईएमडी राशि	882030

#### ePBG Detail/ईपीबीजी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	5.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	28

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

#### Beneficiary/लाभार्थी :

DDO

Room No.704, Proc Cell, Prov Dte, 7th Floor, 10th Block, FHQ BSF, CGO Complex, Lodhi Road, New Delhi-110003  
(Ddo Fhq Bsf)

#### MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	No
--	----

#### Details of the Competent Authority for MII

Name of Competent Authority	Shri Sher Bahadur
Designation of Competent Authority	Under Secretary (Procurement Policy)
Office / Department / Division of Competent Authority	Ministry of Finance, Department of Expenditure, Procurement Policy Division
CA Approval Number	No.F4/1/2023-PPD(pt.)
Competent Authority Approval Date	28-06-2024
Brief Description of the Approval Granted by Competent Authority	Relaxation under Rule 161(iv) of General Financial Rules 2017 for issuance of Global Tender Enquiry (GTE) for procurement of Medical Devices - reg

Competent Authority Approval for not opting Make In India Preference : [View Document](#)

#### MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility

for exemption must be uploaded for evaluation by the buyer.

2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.

3. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

4. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

5. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

6. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

7. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

8. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- i. If number of technically qualified bidders are only 2 or 3.
- ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

#### Pre Bid Detail(s)

Pre-Bid Date and Time	Pre-Bid Venue

24-04-2025 11:00:00

BWWA Hall, Ground Floor, FHQ BSF, 10th Block, CGO Complex, Lodhi Road, New Delhi-110003

**Laparoscope Surgery Set ( 5 pieces )****Technical Specifications/तकनीकी विशिष्टियाँ**

Buyer Specification

Document/क्रेता विशिष्टि दस्तावेज़

[Download](#)**Comprehensive Maintenance**

Warranty of required product

2 Year

Comprehensive Maintenance Duration (Post Warranty)

5 Year

\*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Joydeep Chattopadhyay	110066, FHQ BSF HOSPITAL -I , EAST BLOCK-9 , SECTOR -I, R K PURAM , NEWDELHI- 110066	5	60

**Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें****1. Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be  $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$ , subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

**2. Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

### 3. **Service & Support**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

### 4. **Generic**

Data Sheet of the product(s) offered in the bid, are to be uploaded along with the bid documents. Buyers can match and verify the Data Sheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, the bid is liable for rejection.

### 5. **Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

### 6. **Certificates**

ISO 9001: The bidder or the OEM of the offered products must have ISO 9001 certification.

### 7. **Generic**

**Bidder financial standing:** The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.

### 8. **Warranty**

Warranty period of the supplied products shall be 2 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

### 9. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of

M/s xxxx

A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.

### 10. **Warranty**

Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 15 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

#### 11. Forms of EMD and PBG

Bidders can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C

M/s xxx

. The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date/ Bid Opening date

#### 12. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

#### 13. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

### **Additional Terms & Condition for Procurement of Laparoscope Surgery Set, Qty- 5 Nos as per latest QRs/TDs.**

#### 1. **Earnest Money Deposit (EMD):**

- (a) As per General Terms & Conditions, EMD is required to be submitted by the bidder. However, in case, the bidder is MSE / Startup (Manufacturers of Tendered item) or any exempted entity, they will submit Bid Securing Declaration on company/firm letter head along with Technical Bid in the proforma as per **Appendix-'A'**.
- (b) In order to get the benefit of MSE/Startup prospective bidders are required to register themselves under Udyam Aadhaar registration with **5-digit NIC category Code of 26600 & 32509**. Firms registered under this code will only be exempted from submitting the EMD. Firms with 5-digit NIC Category Code of 26600 & 32509 will be required to submit Bid Securing Declaration Certificate in the prescribed proforma as per **Appendix- 'A'**.
- (c) Traders / Authorized dealer / re-sellers are not exempted for EMD submission being not the OEM of the product.
- (d) In case Bid Security Declaration (BSD)/Scanned copy of EMD is not submitted, Bid shall be rejected / disqualified during initial scrutiny.
- (e) EMD will be submitted in favour of DDO, FHQ, BSF, New Delhi payable as under:-  
  
Bank- State Bank of India  
Branch- CGO Complex Branch  
IFSC – SBIN0007837
- (f) **In case EMD is submitted in the shape of e-Bank Guarantee, the copy of the same should also be submitted along with**

## Technical Bid.

- (g) **Bidders participating as Joint Venture firms are not exempted from submission of EMD** as per Question No.-47 of FAQ issued by GOI, Office of Development Commissioner Micro, Small & Medium Enterprises vide their OM dated 25/03/2022.
2. **Liquidated Damages** :- If the bidder fails to deliver any or all of the Goods/Services within the original/re-fixed delivery period(s) specified in the contract, the Buyer will be entitled to deduct/recover the Liquidated Damages for the delay, unless covered under Force Majeure conditions aforesaid, @ 0.5% of the contract value of delayed quantity per week or part of the week of delayed period as pre-estimated damages not exceed 10% of the contract value of delayed quantity without any controversy/dispute of any sort whatsoever.
  3. **Arbitration** :- As per General Terms & conditions of GeM.
  4. **Risk Purchase Clause**:- In the event of failure of seller to deliver or dispatch the stores within the stipulated dates/period of the supply order /AT , or the event of breach of any of the terms and conditions of the AT, the buyer will have the right to purchase the subject store elsewhere at the risk and cost of defaulting seller after giving a notice. The cost as per Risk Purchase exercise may be recovered from the bills pending with the buyer even against any other supplies outside this contract or even from the pending bills with any other Govt. Department/ Ministry. In the event of contract being cancelled for any breach committed and the purchase effecting repurchase of the subject store at the risk and cost of seller, the buyer is not bound to accept the lower offer of Benami or allied sister concern of the seller.
  5. **Right to Rejection: Laparoscope Surgery Set** will be accepted after inspection and only if they are found up to the standard specification. The decision of the BSF shall be final as to the quality of the stores and shall be binding upon the seller and in case of any of the stores supplied not being found as per specification shall be liable to be rejected or replaced and any expenses or losses caused to the seller should be borne by the seller and ensured by the seller that stores supplied should be best in quality and free from all defects. The acceptance of stores will be made only when the stores are inspected and found upto the standard specification and free from all defects. The rejected stores must be removed by the sellers from the consignee's premises within 15 days from the date of the intimation about rejection at the risk and cost of the seller. The in-charge stores concerned will take reasonable view of such materials but in no case shall be responsible for any loss, shortage, damage that may occur to it while it is in the premises of the consignee.
  6. All Queries related to the Tender will be submitted on GeM Portal only.



7. **FORCE MAJEURE CLAUSE (FMC):** A Force Majeure (FM) means extraordinary events or circumstances beyond human control such as an event described as an act of God (like a **natural calamity**) or events such as a war, strike, riots, crimes (but not including negligence or wrong-doing, predictable/seasonal rain and any other events specifically excluded in the clause). An FM clause in the contract frees both parties from contractual liability or obligation when prevented by such events from fulfilling their obligations under the contract. An FM clause does not excuse a party's non-performance entirely, but only suspends it for the duration of the FM. The firm has to give notice of FM as soon as it occurs and it cannot be claimed ex-post facto. There may be a FM situation affecting the purchase organisation only. In such a situation, the purchase organisation is to communicate with the supplier along similar lines as above for further necessary action. If the performance in whole or in part or any obligation under this contract is prevented or delayed by any reason of FM for a period exceeding 90 (Ninety) days, either party may at its option terminate the contract without any financial repercussion on either side.
8. The minimum average annual financial turnover of the bidder during the last three years, ending on 31<sup>st</sup> March of the previous financial year i.e. FY 2022-23, 2023-24 & 2024-25 should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
9. Bidder participating as authorized dealer will not be given exemption in annual turnover & experience criteria. Bidders participating as authorised dealer will upload authorisation letter/certificate issued by OEM in favour of them.
10. **Participation by Principal/OEM or Indian Agents:-** E-Bids will be evaluated on the basis of CVC guidelines issued vide Circular No. 03/01/12 (12.02.2006 CTE/ SPL(1)-2/161730 dated 13.01.2012.
- a) In a tender, either the Indian agent on behalf of the Principal/OEM or Principal/OEM itself can bid but both cannot bid simultaneously for the same item/product in the same tender.
- b) If an agent submits bid on behalf of the Principal/OEM, the same agent shall not submit a bid on behalf of another Principal/OEM in the same tender for the same item / product.
11. Bidder will upload a self-certificate stating that his firm has **not been blacklisted** by any government / PSU agencies in recent past and also not under liquidation / court receivership / similar proceeding.

12. **Land Border:-** As per Public Procurement OM No.6/18/2019-PPD dated 23 July 2020, any bidder from a country which shares a land border with India will be eligible to bid any procurement whether of goods, service (including consultancy services and non - consultancy service) or works (including turnkey projects) only if the bidder is registered with the Competent Authority specified in Annex-I'. Bidder is required to upload a certificate in this regard as per **Appendix-'C'**
13. QRs/Specification and Trial Directives will be as per **Appendix-'1(H)XX to Appendix-'1(H) XXXII' which is uploaded in Buyer Specification documents at Technical Specifications.**
14. **Joint Receipt Inspection (JRI):** Joint Receipt Inspection (JRI) will be carried out at the place of consignee location as per QRs/Specification and Trial Directives before final acceptance by the BSF Board of Officers and representative of the seller. If any **Laparoscope Surgery Set** is rejected in final inspection as not conforming to Tender Specifications, the same must be replaced within scheduled Delivery Period before final acceptance. The Consignee's right of rejection in this regard will be final and absolute.
15. Bidder should provide Certifications as per QRs/Specification and Trial Directives. It will be ensured that Certifications should be issued prior the bid opening date.
16. **Freight Charge :-** Freight charges will be borne by the supplier. The stores to be dispatched on free delivery at consignee locations.
17. **Training :-** Firm should install the equipment at the consignee location and provide demonstration / training regarding the equipment's at user location.
18. The company should have a local Service Centre.
19. The company should have proven track record in Govt. Sector.
20. Penalty clause and liquidated damages clause as per GeM terms and conditions shall be incorporated.
21. Copies of relevant inspection notes shall be forwarded to the indenter in addition to all concerned / consignee for monitoring the supply position.
22. The items be made available to the indenter within Two months from the date of issue of AT (Acceptance of Tender).
23. **Insurance :-** Insurance of stores will be governed as per clause 50104 (xii) (d) of chapter 5 of Manual for procurement of goods 2017 of MOF.
24. The stores supplied against the order should cover 2 Years warranty from the date of supply of unit, and 5 years CMC after 2 years warranty period should be quoted by vendor.
25. Firm should arrange to provide demonstration of the instrument during Technical Evaluation and JRI would be done at the consignee location as per

r QRs/Specification and Trial Directives.

26. The stores will be supplied by seller to following consignee location free of cost.

S/ N	NAME OF CONSIGNEE	QTY	POSTAL ADDRESS	CONSIGNEE HQ
1.	Dr. Prasun Maiti, DIG (Med)/MS	01	CAPFs COMPOSITE HOSPITAL BSF, KADAMTALA, SILIGURI, DIST-DARJEELING, WEST BENGAL, PIN-734011	COMPOSITE HOSPITAL KADAMTALA
2.	Dr. R K Kotnala, DIG(Medical)/MS	01	CAPFs COMPOSITE HOSPITAL BSF TEKANPUR, PO-TEKANPUR, DISTT-GWALIOR, MADHYA PRADESH, PIN-475005	COMPOSITE HOSPITAL TEKANPUR
3.	Dr. S P Yadav, DIG (Medical)/MS	01	BSF COMPOSITE HOSPITAL AGARTALA, POST OFFICE-SALBAGAN, DIST-AGARTALA, TRIPURA - 799012	COMPOSITE HOSPITAL AGARTALA
4.	Dr. Sailesh Pandit, DIG(Medical)	01	BSF COMPOSITE HOSPITAL, POST OFFICE - AN BLOCK, SECTOR-V, SALT LAKE KOLKATA, WEST BENGAL-700091	COMPOSITE HOSPITAL KOLKATA
5.	Dr. V RAJA, DIG(Medical)	01	BSF COMPOSITE HOSPITAL, POST OFFICE-BSF CAMPUS, MANDOR ROAD, JODHPUR, RAJASTHAN -342026	COMPOSITE HOSPITAL JODHPUR

27. **Conflict of Interest :-** The firm shall submit the declaration as per appendix "D" along with the bid documents failing which their bid shall be rejected.
28. User Manual should be provided with each **Laparoscope Surgery Set.**
29. All the equipment/ components of Laparoscopic Surgery Set should have standard, safety and training related approval/certification as mentioned below:-
- i) Should be USFDA or European CE approved.
  - ii) Manufacturer should have IOS certification, where Indian standards are

e not available.

30. As per **Srl. No.'A' of QRs / Specifications** - (a) should be USFDA or European CE approved. (b) Manufacture should have ISO certification, where Indian standards are not available.

**Clarification - Laparoscope Surgery Set** should be USFDA/CE/ C DSCO approved. As per guidelines of MoHFW, Dept of Health & Family Welfare (Drugs Regulation Section). OM dated 20 Feb' 2018, Para 2, sub para (I) no. Wherever the Indian standards are available, these would be sufficient and the indenting organization shall not insist on any specification or standards like USFDA or CE Certification etc.

31. For clause No.A - High Definition 4k Digital Camera - HD **(7)** of QRs / Specification and Trial Directives - It should have Pure digital signal with High definition video of 1920 x 1080p (minimum) native resolution and progressive scan technology both on camera head and console).

**Clarification** - As per approved QRs specifications, minimum has been mentioned. Anything above / advance, it is acceptable. Therefore, pure digital signal with 4k video of 3840 x 2160 native resolution and progressive scan technology both on camera head and console is acceptable.

Justification: 4k resolution (3840 x 2160), as per technology up gradation. 4k 3D ICG is latest technology for better results and better patient outcome.

On search of GeM Portal, it is found that total 105 Firms are available which have 3840 x 2160 (4k), while only 08 firms are available which have 1920 x 1080 (Full HD).

32. For clause No. Technical Specifications : **(1)** of QRs / Specification and Trial Directives - Image sensor (Image System : 1/3" Progressive Scan CCD).

**Clarification** - Image System 1/3" Progressive scan, 3 Chip CCD/C MOS.

Justification: CMOS (Complementary Metal Oxide Semi Conductor) are advance than CCD (Charge Coupled Device)

-On Search of GeM Portal, it is revealed that, CMOS/CCD shown.

33. For clause No. Technical Specifications : **(2)** of QRs / Specification and Trial Directives - Pixels (Pixels 1920 x 1080p pixels per chip (min).

**Clarification** - As per approved QRs specifications, minimum has been mentioned. Anything above / advance, it is acceptable. Pixels 3840 x 2160 p pixels per chip (min) are acceptable.

Justification: On search of GeM Portal, it is found that total 105 Firms are available which have 3840 x 2160 (4k), while only 08 firms are available which have 1920 x 1080 (Full HD).

34. For clause No. Technical Specifications : **(6)** of QRs / Specification and Trial Directives - Video Outputs (Video output : S-video signal).

**Clarification** - The bid was floated on GeM Portal on the basis of a

approved QRs. As such in addition to our requirement. If any Vendor provides any other latest / higher / advanced version. It will be permitted.

35. For clause No. High Resolution Monitor Qty 2 : **(1)** of QRs / Specification and Trial Directives – HDTV (HDTV display in original 16:9 HDTV format.)

**Clarification** - As per approved QRs specifications, minimum has been mentioned. Anything above / advance and for wider participation, it is acceptable, 4k medical grade with 30 inch screen or more is acceptable. However, Ultra HDTV display in original 16:9 HDTV format may also be acceptable as it is better / advance and anything above / advance of approved Specifications.

Justification: On search of GeM Portal, it is found that total 105 Firms are available which have 3840 x 2160 (4k), while only 08 firms are available which have 1920 x 1080 (Full HD).

36. For clause No. High Resolution Monitor Qty 2 : **(2)** of QRs / Specification and Trial Directives – 1080P (1080 p/50 & 1080 p/60 displays possible).

**Clarification** - 3840 x 2160 P can be preferred as it is better / advance. Anything above / advance of approved Specifications and for wider participation, it is acceptable.

Justification: On search of GeM Portal, it is found that total 105 Firms are available which have 3840 x 2160 (4k), while only 08 firms are available which have 1920 x 1080 (Full HD).

37. For clause No. High Resolution Monitor Qty 2 : **(5)** of QRs / Specification and Trial Directives – Max Resolution (Max resolution of 1920 x 1080).

**Clarification** - Resolution of 3840 x 2160 P can be preferred as it is better / advance. Anything above / advance of approved Specifications and for wider participation, it is acceptable.

Justification: On search of GeM Portal, it is found that total 105 Firms are available which have 3840 x 2160 (4k), while only 08 firms are available which have 1920 x 1080 (Full HD).

38. For clause No. High Resolution Monitor Qty 2 : **(8)** of QRs / Specification and Trial Directives – HD Medical Grade (HD medical grade monitor not less than 27 inches compatible with 4k digital camera).

**Clarification** - As per approved QRs specifications, minimum has been mentioned. Anything above / advance and for wider participation, it is acceptable. 4k medical grade with 30 inch screen or more is acceptable. However, UHD medical grade monitor not less than 32 inch compatible with 4k digital camera and has 12G-SDI for true 4k Image is also acceptable.

39. For clause No. C. Light Source, Qty 1 : QRs / Specification and Trial Directives – LED light Sources.

**Clarification** - 4k and ICG compatible Light Source. Since it is better and advance version.

40. For clause No. C. Light Source, Qty 1 : QRs / Specification and Trial Directives – Life of light bulb  $\geq$  30000 hrs.

**Clarification** - Lamp Life should be 30000 hours and more.

41. For clause No. D. Insufflator Qty 1 : QRs / Specification and Trial Directives – Flow rate 35-45 Liter / mins.

**Clarification** - Terms & Conditions of the Tender be adhered to.

42. For clause No. E. Laparoscope Qty 1 : QRs / Specification and Trial Directives.

**Clarification** - 4k and ICG compatible Laparoscope. Since it is better and advance version.

43. For clause No. F. HD Recorder : QRs / Specification and Trial Directives.

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement. If any Vendor provides any other latest / higher / advanced version. It will be permitted.

44. For clause No. F. HD Recorder : (3) of QRs / Specification and Trial Directives – 1080 P Input – (Should have video inputs minimum 2 nos. S-Video, 2 nos. Composite, 1 XGA (1024 X 768) and 1 High-Definition (1280 X 1024)).

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement, if any Vendor provides any other latest / higher / advanced version. It will be permitted.

45. For clause No. F. HD Recorder: (4) of QRs / Specification and Trial Directives – 1080 P Input – (Should have video inputs S-Video, 2 nos. Composite, DVI & XGA (1024 X 768, 1280 X 1024, and 1600 X 1200)).

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement. If any Vendor provides any other latest / higher / advanced version. It will be permitted.

46. For clause No. F. HD Recorder: (8) of QRs / Specification and Trial Directives – Touch Screen (Should have touch screen (min 10 inch) control panel interface).

**Clarification** - Terms & Conditions of the Tender be adhered to.

47. For clause No. F. HD Recorder: (9&10) of QRs / Specification and Trial Directives – Multi Session disc recording (Should support file formats for images: Bitmap (BMP), JPEG, JPEG2K, Tagged Image File Format (TIFF) Videos : MPEG-I, MPEG-2 and MPEG-4).

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement. If any Vendor provides any other latest / higher / advanced version. It will be permitted.

48. For clause No. F. HD Recorder: (12) of QRs / Specification and Trial Directives – Real time Broadcasting (Real time Broadcasting facilities over internet in HD Formats should be there).

**Clarification** - The provision of Real time Broadcasting was included in the bid on the basis of approved QRs and hence it may not be construed to be suiting one specific company and accordingly the said clause can not be deleted. Further such facilities are being provided not by specific c

company, rather also by other companies. As such Terms & Conditions of the Tender be adhered to.

**Justification:** On search at GeM Portal it is found that 94 Models which are made by 07 Companies are available which have live streaming facility.

49. For clause No. G. Instruments of QRs / Specification and Trial Directives - (Instruments & accessories for laparoscopy set).

**Clarification** - Terms & Conditions of the Tender be adhered to.

50. For clause No.A - High Definition 4k Digital Camera - HD **(4)** of QRs / Specification and Trial Directives - Universal camera head coupler.

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement. If any Vendor provides any other latest / higher / advanced version. It will be permitted.

51. For clause No.A - High Definition 4k Digital Camera - HD **(5)** of QRs / Specification and Trial Directives - Optical Zoom facility.

**Clarification** - For clause No.A - High Definition 4k Digital Camera - HD **(4)** of QRs / Specification and Trial Directives - Universal camera head coupler.

52. For clause No.A - High Definition 4k Digital Camera - HD **(12)** of QRs / Specification and Trial Directives - Video outputs: two DVI, one SVHS and one (optional) direct fibre optic output.

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement. If any Vendor provides any other latest / higher / advanced version. It will be permitted.

53. For clause No.A - High Definition 4k Digital Camera - HD **(14)** of QRs / Specification and Trial Directives - The system should be ready to use as soon as it is connected to the camera control unit.

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement. If any Vendor provides any other latest / higher / advanced version. It will be permitted.

54. For clause No.A - Technical Specification **(3)** of QRs / Specification and Trial Directives - Camera Head weight 150-250gm.

**Clarification** - Terms & Conditions of the Tender be adhered to.

55. For clause No.B - High Resolution Monitors Qty-2 **(6)** of QRs / Specification and Trial Directives - Desktop with pedestal.

**Clarification** - Terms & Conditions of the Tender be adhered to.

56. For clause No.B - High Resolution Monitors Qty-2 **(18)** of QRs / Specification and Trial Directives - Video Inputs 2\* DVI, 2\*3G SDI, 1\*S-Video, 1\*SO G, 1\*RGBNGA, 1\*RS232, 1\*RJ 45 interface.

**Clarification** - The bid was floated on GeM Portal on the basis of a

approved QRs. As such in addition to our requirement. If any Vendor provides any other latest / higher / advanced version. It will be permitted.

57. For clause No.B - High Resolution Monitors Qty-2 **(19)** of QRs / Specification and Trial Directives - Output 1\*DVI, 1\*3G SDI, 1\*S-Video.

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement. If any Vendor provides any other latest / higher / advanced version. It will be permitted.

58. For clause No.D - Laparoscope : Qty-01 each **(2)** of QRs / Specification and Trial Directives - Optimal Centre to edge resolution for enhanced picture quality.

**Clarification** - Optimal centre to edge 4k resolution for enhanced picture quality is acceptable.

Justification: Since it is better and advance technology.

59. For clause No. High Definition Digital Documentation System (Recorder System) Qty-01. **(3)** of QRs / Specification and Trial Directives - Should have video inputs minimum 2 nos S-Video, 2 nos Composite, 1 XGA (1024 X 768) and 1 High Definition (1280 X 1024).

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement. If any Vendor provides any other latest / higher / advanced version. It will be permitted.

60. For clause No. High Definition Digital Documentation System (Recorder System) Qty-01. **(7)** of QRs / Specification and Trial Directives - Should have Disc capacity of 250 GB/1Tera Byte.

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement. If any Vendor provides any other latest / higher / advanced version. It will be permitted.

61. For clause No.A High Definition 4k Digital camera - Qty-1. (1 to 15) of QRs / Specification and Trial Directives as under -

- 1.Digital matrix processing technology.
- 2.Progressive scanning system for camera.
- 3.Special filters for chrome endoscopy
- 4.Universal camera head coupler
- 5.Optical Zoom Facility
- 6.System capable for ICG fluoroscopy for better delineation of anatomy.
- 7.It should have Pure digital signal with High Definition video of 1920 x 1080p (minimum) native resolution and progressive scan technology both on camera head and console.
- 8.It should be compatible with aspect ratio of 4:3 and 16:9
- 9.The system should have Digital and Optical Zoom to enhance the quality of image size & cross specialty standardization of the camera system, regardless of the telescope used.
- 10.Digital Zoom, white balance control and two peripheral controls on camera head.
- 11.Integrated Gain /Shutter / Enhancement with automatic brightness control.



12.Video Outputs: two DVI, one SVHS and one (optional) direct fiber optic output.

13.The system should automatically optimize all settings.

14.The system should be ready to use as soon as it is connected to the camera control unit.

15.The system should be menu driven, thus allowing the surgeon to programme the camera head functions as per the surgical needs and requirements.

**Clarification** - As per approved QRs specifications, minimum has been mentioned. Anything above / advance, it is acceptable. Therefore, pure digital signal with 4k video of 3840 x 2160 native resolution and progressive scan technology both on camera head and console is acceptable.

Justification: 4k resolution (3840 x 2160), as per technology upgradation. 4k 3D ICG is latest technology for better results and better patient outcome.

-On Search of GeM Portal, it is found that total 105 Firms are available which have 3840 x 2160 (4k), while only 08 firms are available which have 1920 x 1080 (Full HD).

62. For clause No.A Technical Specifications : (1 to 8) of QRs / Specification and Trial Directives as under –

- 1.Image System : 1/3" Progressive Scan CCD
- 2.Pixels 1920 x 1080p pixels per chip (min)
- 3.Camera Head Weight 150-250 gm
- 4.AGC:Microprocessor controlled
- 5.Signal to noise ratio 65-75 dB
- 6.Video output : S-video signal
- 7.Digital Video Interface
- 8.Power Supply 100 – 240 VAC, 50/60 HZ

**Clarification** - As per approved QRs specifications, minimum has been mentioned. Anything above / advance, it is acceptable.

Therefore, 4k head should have 3840 x 2160 resolution with ICG also add 3D (3840 X 2160) 3D. Scope in 4k resolution with ICG Function is acceptable.

Justification: - 4k resolution (3840 x 2160), as per technology upgradation. 4k 3D ICG is latest technology for better results and better patient outcome.

- On Search of GeM Portal, it is found that total 105 Firms are available which have 3840 x 2160 (4K), while only 08 firms are available which have 1920 x 1080 (Full HD).

63. For clause No.B High Resolution Monitors: (1 to 19) of QRs / Specification and Trial Directives as under –

- 1.HDTV display in original 16:9 HDTV format.
- 2.1080p / 50 & 1080p /60 displays possible.

- 3.Drip water protected, dust proof housing.
- 4.LED crystal display.
- 5.Max resolution of 1920 x 1080
6. Desktop with pedestal.
- 7.Should have the facility of PIP mode.
- 8.HD medical grade monitor not less than 27 inch compatible with 4 k digital camera with screen diagonal "26".
- 9.Aspect Ratio 16:9 HD format.
- 10.Brightness 400cd/m2
- 11.Maximum viewing angle : 178 - degree vertical
- 12.Contrast ratio : 1000:1
- 13.Reaction time – 8 ms
- 14.Rated Power : 115 watts
- 15.Power supply 100-240 VAC 50 / 60 HZ
- 16.Screen dimensions : 673X418X88 mm
- 17.Video inputs 2 DVI 3 G SDI Video SOG 1 RGBNGA 1 inch RS 323 1 inch RJ interface.
- 18.Output 1 inch DVI 1- inch 3G SDI 1 inch S-video
- 19.Accessories External 24 VDC power supply, main cord, Pedtial.

**Clarification** - 4k 3D Monitor can be preferred as it is better / advance and anything above / advance of approved Specifications, it is acceptable.

Justification: Only 4k monitor able to provides clarity of the image as per 4k resolutions.

64. For clause No.C LED light Sources : Qty 1: (1 to 15) of QRs / Specification and Trial Directives as under –

- 1.Authomatic and manual light adjustment.
- 2.Color temperature 5500-6500 k
- 3.Life of light bulb > = 30000 hrs
- 4.Universal adapter for fibre optic cable
- 5.Efficient cooling technology in light sources.
- 6.Degree of protection IPx1.
7. 300 W
8. 220 volts
- 9.Light engine Red, green, & blue LEDs.
- 10.Increased patient safety & added protection in OR with safe light technology.
- 11.Intuitive simple user interface with LCD touch screen.
- 12.Standby mode
- 13.Universal jaw assembly to adapt any make of fibre optic cable.

14. More than 5000 hours bulb life.
15. Light intensity adjustment continuously adjustable from 0 to 100 % manually.

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement. If any Vendor provide any other latest / higher / advanced version. It will be permitted.

65. For clause No.D Laparoscope: Qty 1 each: (1 to 8) of QRs / Specification and Trial Directives as under –

1. Wide Angle Full Screen, Autoclavable, Forward – Oblique and lateral scope.
2. Optimal centre-to-edge resolution for enhanced picture quality.
3. Angle of view 30° and 0° of 10 diameter – 01 each.
4. Angle of view 30 degree of 5 mm diameter – 01 no.
5. Length 30 cm
6. Fibre optic light transmission incorporated.
7. Standard ocular window for coupling the camera head.
8. Scratch resistance sapphire coated tip lens.

**Clarification** - The bid was floated on GeM Portal on the basis of approved QRs. As such in addition to our requirement. If any Vendor provide any other latest / higher / advanced version. It will be permitted.

66. **Pre-Bid Meeting :-**

- i) All potential bidders should try to attend the Pre-Bid meeting scheduled on **24/04/2025 at 1100 Hrs**. In case any potential bidder is unable to attend the Pre-Bid meeting, he can send his queries on GeM Portal only before the Pre-Bid meeting date so that the queries can be discussed during the Pre-Bid meeting itself.
- ii) If any supplementary query arises after the Pre-Bid meeting, then the said bidder must send his **query on GeM Portal latest by 25/04/2025** so that the buyer is able to formulate reply/clarification of such query well in time before the Bid opening date.
- iii) **“Query Submission End Date”**:-The queries from bidders will be allowed only on GeM Portal and only till **25/04/2025, thereafter no query will be entertained.**
- iv) All queries will be entertained through GeM Portal window only. Other offline or online enquiries through phone or any other medium will not be entertained.

67. **Check list as per Annexure -A is to be submitted.**

**CHECK LIST FOR BIDDERS**

S. No	Requirements to be checked before submission of the Bid.	Complied (Please indicate YES or NO)	Indicate Page No.
1.	Whether the bidder is either:		
	(i) Original Equipment Manufacturer		
	(ii) Authorized Dealer/Re-seller etc.		
	(iii) If authorised Dealer/Re-seller then authorisation certificate issued by OEM be submitted.		
2.	Minimum Average Turnover of the Bidder (Last 3 years i.e 2022-23, 2023-24 & 2024-25)		
3.	Minimum Average Turnover of the OEM (Last 3 years i.e 2022-23, 2023-24 & 2024-25)		
4.	Bidder financial standing: The bidder should not be under liquidation, court receivership or similar proceedings should not be bankrupt. Bidder to upload undertaking to this effect with bid.		
5.	Bidder / OEM must have Dedicated / toll free Telephone Number for Service Support.		
6.	Earnest Money Deposit (EMD) has been enclosed. However, in case, the bidder is MSE /Startup ((Manufacturers of Tendered item) , they will submit Bid Security Declaration in format <b>(Appendix -‘A’)</b> .		

7.	QRs / Specifications and Trial Directives as per <b>Appendix-'F'</b> which is uploaded in Buyer specification documents at page No.5 of Custom Bid of Technical Specifications.		
8.	Warranty terms as per bid accepted as mentioned at Srl No.14 at P/No.7 of Custom- Bid.		
9.	Data Sheet of the products(s) offered in the bid as mentioned at Srl No.5 at P/No.6 of Custom - Bid.		
10.	Escalation Matrix For Service Support as mentioned at Srl No.6 at P/No.6 of Custom - Bid.		
11.	ISO 9001-Certificate as mentioned at Srl. No.15 at P/No.7 of Custom - Bid.		
12.	Acceptance of Option Clause as mentioned at Srl No.01 at P/No.5 of Custom - Bid.		
13.	Bidder Financial Standing as mentioned at Srl. No.2 at P/No.5 of custom bid.		
14.	Bidder will upload a self-certificate stating that his firm has not been blacklisted as mentioned at Srl No.12 at P/No.9 of ATC.		
15.	Freight charges will be borne by the supplier as mentioned at Srl.17 at P/No.10 of ATC .		
16.	Acceptance of consignee location as mentioned at Srl No.27 of ATC at Page No10.		
17.	Training terms as mentioned at Srl No.18 of ATC at Page No10.		
18.	Payment & Delivery terms as per bid accepted.		

19.	Acceptance of all Additional Terms & Conditions of bid by bidders.		
20.	Land Border Certificate as per <b>Appendix-‘C’</b> .		
21.	US certification required to upload if IS standards not available		
22.	Conflict of interest as per <b>Appendix-‘D’</b> .		
23.	USFDA or European CE certification required to upload		
24.	IOS certification required, where Indian standards are not available		
25.	User/Technical/Maintenance Manual should be provided in English / Hindi with each <b>Laparoscopy Surgery Set</b> .		
26.	Name and Complete Address for correspondence with E-mail & contract Number.		
27.	Warranty/Guarantee as per Bid		

**Note :- All column must be filled & paging be done in chronological order in respect of the above check list.**

**Appendix - “A”**

**Open Tender Enquiry No.....**  
**.....**

**Dated ...**

### **BID SECURING DECLARATION CERTIFICATE**

**( On Firm’s letter head )**

To

Director General, Border Security Force  
Procurement Cell, Provisioning Directorate  
10<sup>th</sup> Block, CGO Complex, Lodhi Road,  
New Delhi – 110 003

Sir,

Tender/GeM.Bid No. \_\_\_\_\_

We, **M/S** \_\_\_\_\_ under take that if we withdraw or modify our bid during the period of validity, or we are awarded the contract and we fail to sign the contract, or to submit a performance security before the deadline defined in bids documents, we will be suspended for the period of **02 years** from being eligible to submit bids for contract with procuring entity i.e. **Border Security Force (BSF), Ministry of Home Affairs.**

Yours Faithf

ully

(Signature of the Authorized Signatory of Bidder, with Official seal)

#### **Appendix-‘C’**

##### **(To be in the Company letter head)**

Certificate regarding compliance of Ministry of Finance, Department of Expenditure, Public Procurement Division Order No-F No-6/18/2019-PPD date 23/07/2020 .

I Certify that ..... (Name of the Firm) is not covered by the definition of the bidder as mentioned in Annexure -III of MoF order No. 6/18//2019-PPD dated 23/07/2020 and is thus not required to be registered with the Competent Authority for this purpose.

OR

I certify that ..... (Name of the Firm) comes within the ambit of clauses mentioned in Annexure -III of MoF order No. 6/18//2019-PPD dated 23/07/2020 and for this purpose, it is registered with the Competent Authority. Valid registration certificate No..... dated ... issued by ..... is enclosed.

(Strike off the portion not applicable)

Signature of tenderer :-----

Name of block letter :-----

Name of firm :-----

Full address :-----

-----

**Appendix-‘D’**

**DECLARATION IN RESPECT OF CONFLICT OF INTEREST**

01	Name of firm	Office address of firm, Phone No., Fax No & Email ID	Firms works Address as per Vendor Registration, Phone No Fax No Email ID	Type of Firm (Sole Proprietorship/ Partnership / Limited Liability Partnership/ Private Limited Company/ Public Limited Company)	CIN/LLPIN of firm	GST Registration No of the firm
	(1)	(2)	(3)	(4)	(5)	(6)



02	Name of all Owner(s) / Director(s) of firm	Father's / Husband's name of all Owners / Director(s)	DIN	Full residential Address	Aadhar Card of owner / Director(s) of firm be attached	Udyam Registration Certificate No., attach a copy
	(1)	(2)	(3)	(4)	(5)	(6)

03. WE HEREBY DECLARE AND CONFIRM THAT no bid has been submitted for or the quoted item by any other Private Limited or Public Limited Company or Limited Liability Partnership (LLP) or Partnership Firm or Proprietor Firm in which any relative of any Director or Partner or Proprietor of Bidder Firm is Director, Partner or Proprietor. The word "relative" is defined as under:-

The companies Act, 2013 Sec 2(77) "relative" & "MOF Manual for Procurement of Goods (Para 5.1.4) with reference to any person, means anyone who is related to another, if :-

they are member of a Hindu Undivided Family"

they are husband and wife " or

One person is related to the other in such manner as prescribed below:-

Mother including step-mother

Son including step-father

Son's wife

Daughter

Daughter' Husband

Brother including step-brother

Sister including step-sister

04. **WE FURTHER DECLARE THAT** we have carefully read and understood the clause relating to "Conflict of Interest" of Tender No.\_\_\_\_\_.

05. **We hereby certify that** our firm M/s \_\_\_\_\_ (firm name) \_\_\_\_\_ do not have any Conflict of Interest with other bidders for particular quoted item viz \_\_\_\_\_ (Item name) We hereby declare and confirm that the above information and particulars are true and correct.

For \_\_\_\_\_ (firm name)

Place :

Date :

Signature of Director / Partner / Proprietor / Authorized Signatory

Name : \_\_\_\_\_

Designation \_\_\_\_\_

(With official Stamp)

## **Disclaimer/अस्वीकरण**

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of

bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

## **Additional Clause For Comprehensive Maintenance Charges**

- 1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC.Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
- 2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.
- 3.GST shall be included in the CMC Charges quoted.
- 4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula.A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.
- 5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.
- 6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available.Under this option up to 10 years can be chosen for CMC charges beyond warranty period.
- 7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.
- 8.The CMC functionality shall be available in bid only and no direct RA shall be applicable.In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1. Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

CMC charges for 1st year after warranty period- Percentage to be indicated- A1

CMC charges for 2nd year after warranty period- Percentage to be indicated- A2

CMC charges for 3rd year after warranty period - Percentage to be indicated- A3

CMC charges for 4th year after warranty period - Percentage to be indicated- A4

CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

8.2. The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.

8.3. In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables

(i) Number of years for which CMC required.

(ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation=

$C + C * \{ (A1/100)/(1.10^n) + (A2/100)/(1.10^{n+1}) + (A3/100)/(1.10^{n+2}) + (A4/100)/(1.10^{n+3}) + (A5/100)/(1.10^{n+4}) \}$  and so on

C - Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1, A2, A3, A4 & A5 shall depend on how many years CMC selected. For 3 year CMC, only A1, A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4. CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5. The CMC charges shall be offered within range of 3 to 10% of cost of equipment.

9. Since CMC charges are to be paid only later for each year during CMC period, applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10. Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges. The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification. Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11. In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis. The equipment cost and CMC charges (year to year) shall be matched individually.

12. The CMC Contract shall be an offline contract to be handled by buyer. The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

13. CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

**All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.**

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action

in accordance with the laws./ जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा

जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

**---Thank You/धन्यवाद---**