

Test Report

Report To:
Company: Mister Shield Services Co., Ltd.
Address: 19/87-90 Soi Sukhumvit 13,
Klong Toei Nuea, Watthana
Bangkok 10110 ,THAILAND
Fax. :

Report No: 200000648790
Issued Date: Jul 10, 2020
Revision No: 00
Submission No: SP_20200700048
Page: 1 of 1

Sample type: Disinfectant
Sample condition: Room Temperature
Physical condition: Normal
Test date: Jul 07, 2020 - Jul 10, 2020

Sample collected date:
Sample collected from: -
Sample collected by: -
Sample received date: Jul 1, 2020

Sample ID	Sample description	*Dissolve Ratio	*Microorganism	*Diluent	*Contact time	*No. of initial cells(No)	*No. of survivors (Na)	*Reduction (R)
		-	-	-	-	cfu/mL	cfu/mL	-
001	HOCL	No dilution	<i>Pseudomonas aeruginosa</i> ATCC 15442	-	10 min	36000000	<10	>6.56

Reference Method

BS EN 1656:2009

Comment: Tested date : July 02, 2020, Test temperature : 22 C, Interfering substance : 0.03% Bovine albumin (Final concentration), Neutralizer : N-buffer, R=Log No - Log Na, According to BS EN 1656 : 2009, the disinfectant shall be deemed to have efficiency if value is not less than 5 under defined test conditions

Remarks: * Analysis is not accredited under ISO/IEC 17025

Approved by:

(Kanidtha Prapin)

Microbiology Department Manager

This electronic document has been generated and sent by Betagro Science Center (BSC) and is as valid as a printed copy provided by us.

This result report is valid for the received sample(s) only. It's not to be used for advertising purposes.
This report shall not be reproduced except in full and under approval of Betagro Science Center Co.,Ltd
136 Moo 9, Klong Nueng, Klong Luang, Pathumthani, 12120, Thailand Tel.0 2564 7932-40 Fax.0 2564 7941