

Non-complete-ring positron emission tomography (PET) detection method

Yeqi Fang^{1,*} and Rong Zhou^{1,†}

¹College of Physics, Sichuan University, Chengdu, 610065, China

Positron Emission Tomography (PET) is a vital molecular imaging tool widely used in medical diagnosis and treatment evaluation. Traditional PET systems typically rely on complete detector rings to achieve full angular coverage for uniform and statistically robust sampling of coincidence events. However, incomplete-ring PET scanners have emerged in various scenarios due to hardware failures, cost constraints, or specific clinical needs. In such cases, conventional reconstruction algorithms often suffer from performance degradation due to reduced data completeness and geometric inconsistencies. This thesis proposes a coarse-to-fine reconstruction framework for incomplete-ring PET scanners. The framework first employs an Attention U-Net model to recover complete sinograms from incomplete ones, then uses the OSEM algorithm for preliminary reconstruction, and finally applies a two-stage architecture comprising a Coarse Prediction Module (CPM) and an Iterative Refinement Module (IRM) for fine reconstruction. Our approach utilizes neighboring axial slices and spectral transform features as auxiliary guidance at the input level to ensure spatial and frequency domain consistency, and integrates a contrastive diffusion strategy at the output level to improve correspondence between low-quality PET inputs and refined PET outputs. Experimental results on public and in-house brain PET datasets demonstrate that the proposed method significantly outperforms existing approaches in metrics such as PSNR (35.6421 dB) and SSIM (0.9588), successfully preserving key anatomical structures and tracer distribution features, thus providing an effective solution for incomplete-ring PET imaging.

I. INTRODUCTION

A. Background and Motivation

Positron emission tomography (PET) is a powerful molecular imaging technique that provides quantitative visualization of metabolic processes within living tissue [1]. The fundamental principle is elegant but complex: a radiotracer introduces positrons that annihilate upon encountering electrons, and then produce pairs of 511keV photons traveling in opposing directions. The photon pairs are then detected by a pair of scintillation sensors, and each coincidence event will define a line-of-response (LOR). Traditional PET systems usually use complete 360° detector rings to maximize their sensitivity and angular coverage [1].

But what happens when a complete ring isn't feasible? This question has driven the development of incomplete-ring PET systems. Such systems have emerged from practical necessity rather than theoretical preference: they reduce costs and complexity, allow closer access to patients in specialized applications like breast scanning [2], create "open" configurations that alleviate claustrophobia [3, 4], and enable novel applications such as dual-panel brain imaging systems [5]. The trade-off is severe—complete angular coverage creates gaps in projection data, turning reconstruction into an underdetermined problem. Missing angular views inevitably introduce artifacts and resolution non-uniformity [2, 6]. Even time-of-flight capabilities, which partially fix these problems, can't fully compensate for limited view angles [2, 4].

Researchers have investigated a lot of ways to achieve better and better incomplete-ring PET reconstruction. Analytical ways like filtered back-projection falter immediately—their assumption of complete data leads to pronounced streak artifacts [6]. Iterative approaches like maximum-likelihood expectation-maximization (MLEM) are better by modeling the acquisition process [7], but they still have artifacts along missing view directions [5]. Penalized likelihood methods incorporating prior constraints show promise, as Zhang *et al.* demonstrated by using high-quality prior images to enhance contrast recovery in a dual-panel head-and-neck PET system [5]. But these methods are still not enough to solve the fundamental problem.

Deep learning has brought new ideas to this challenge, though with mixed results. Regression-based networks have shown promise in low-dose PET denoising and partial data reconstruction [8] by learning direct mappings between degraded and high-quality images. Liu *et al.*'s U-Net approach to transform artifact-degraded partial-ring PET images [9] demonstrated initial success, but conventional CNNs struggle with the global features needed to address large-scale angular losses.

Current available techniques have some drawbacks. For examples, while GAN approaches are good at generating visually compelling reconstructions, [10] they are unstable when training and often highly sensitive to hyperparameters, which limit their clinical use. Methods based on explicit likelihood modeling—like VAEs and flow-based approaches—offer solid theoretical foundations. But they usually lose details in reconstructed images and their low speeds also make them not suitable for clinical use. Model-based deep learning frameworks and generative models [11, 12] show promise but risk introducing hallucinated features—a dangerous proposi-

* fangyeqi@stu.scu.edu.cn

† zhourong@scu.edu.cn

tion in medical imaging.

The core challenge is the fundamental **information loss**, which can not be solved, only mitigation. Analytical methods collapse under data deficiency, over-regularized iterative methods blur subtle features, and deep learning models risk introducing misleading artifacts. Current approaches fail in at least one critical dimension—they can't handle severe data loss, require unrealistic computational resources, or compromise the detailed features that clinicians depend on for diagnosis. This study solves these problems by introducing a new coarse-to-fine model for both sinogram and its reconstruction which can preserve details useful for accurate clinical diagnosis.




FIG. 1. Principles of Positron Emission Tomography (PET) imaging.

The remainder of this thesis is organized as follows: **Section II** (*Background on PET Imaging and Incomplete Ring Geometry*) provides an overview of PET and its basic physics like principles of coincidence detection, and the main problems caused by incomplete rings. We also summarize classical and modern reconstruction methods. **Section III** (*Fundamentals of Diffusion Probabilistic Models*) details the background of Denoising Diffusion Probabilistic Models (DDPM) and how they apply to various image reconstruction tasks, establishing the theoretical foundation for our proposed method. *Proposed Coarse-to-Fine Reconstruction Framework*) describes the detailed workflow of our method, including complete and incomplete ring geometric modeling, creation of list-mode and sinogram data, the coarse-to-fine design, auxiliary guidance modules, contrastive diffusion learning objectives, and implementation details. **Section IV** (*Experiments and Results*) presents our experimental setups including the restoration of both sinogram and its reconstruction. We also include hyperparameter choices, metrics to assess the performance of our model, and comparisons with other excellent methods. **Section V** (*Summary and Prospective*) summarizes our findings by analysing metrices of model performance on test dataset, and discusses limitations and directions for future improvements. **Appendices** provide additional experiments, extended qualitative results, and tables of hyperparameters used in the study.

II. BACKGROUND ON PET IMAGING AND INCOMPLETE RING GEOMETRY

When two detectors measure gamma photons within a narrow time window (typically 6-12 nanoseconds), we assume they are from the same annihilation event. This paired detection forms a line of response (LOR) between the two detectors. PET systems record these events in one of two primary formats. In **list-mode data**, each coincidence is logged individually with precise identification of the detector pair and timestamp. This approach saves the maximum amount of information but generates enormous datasets. **Sinogram data** organizes coincidence events into radial, angular, and axial bins, creating a structured histogram that trades off temporal resolution for simpler processing. Standard PET scanners use 360° detector rings to achieve uniform LOR sampling, but things would be quite different when this ideal geometry isn't possible.

A. Incomplete Ring PET Scanners

Incomplete ring configurations emerge from a clash between ideal physics and practical reality. Budget constraints often drive this compromise—each detector module constitutes a substantial cost, and reducing their number can make PET technology accessible to more facilities. In specialized applications, partial ring designs actually provide benefits: breast imaging benefits from closer detector positioning, interventional procedures require open scanner designs, and claustrophobic patients experience less anxiety with more open configurations.

The consequences of missing detectors are profound. LORs that would normally pass through the missing sections simply vanish from the dataset, creating angular sampling gaps. We will see a mess if we directly reconstruct the image out of incomplete sinogram, as shown in Figure 3.

B. System Model and Data Simulation

To deal with these problems systematically, we made a thorough simulation framework modeling both complete and incomplete PET geometries. Our approach begins with a standard ring configuration of R axial rings (we used $R = 42$ in our experiments), each containing $D = 182$ detectors arranged cylindrically. The scanner radius (ρ) of 253.71 mm balances spatial resolution against sensitivity.

Early in our research, we attempted to use Transformer-based approaches for direct 3D image recovery, following Hatamizadeh et al.'s UNETR framework [13]. The segmentation results were excellent, but when applied to our incomplete sinogram data, these methods did not work well, with SSIM of less than 0.2 even after a long time of training. This indicates that mere geometric




FIG. 2. Three-dimensional schematic of incomplete and complete ring PET scanner detector structure, with the right image showing a perspective view.

information is insufficient for high-quality reconstruction when some angular segments are missing.

From listmode to sinogram, and from sinogram to reconstruction, the information is constantly losing. So, rather than attempting direct image-domain recovery, we pivoted to a two-stage method—first converting listmode data to sinograms with missing sections, then applying specialized techniques to complete these sinograms before final reconstruction. This approach, while computationally more intensive, provided a more applicable way to solve the missing data problem by retaining as much information as possible from sinogram.

For simulation purposes, we mapped a $128 \times 128 \times 128$ voxel grid to physical space, yielding in-plane voxel dimensions of approximately 2.78 mm:

$$\Delta x \approx \Delta y \approx \frac{25 \text{ cm}}{128} \approx 2.78 \text{ mm.} \quad (1)$$

The detector system itself employs a hierarchical structure with parameters that were hard to optimize. Table I describes some important configuration parameters of our PET system.

TABLE I. PET Scanner Configuration Parameters

Parameter	Value
Radius	253.71 mm
Crystal transaxial spacing	4.02 mm
Crystal axial spacing	5.37 mm
Module axial spacing	37.56 mm
Crystal elements (transaxial)	13
Crystal elements (axial)	7
Transaxial sectors	28
Axial sectors	1
Modules (axial)	6
Crystals per ring	364
Number of rings	42

Our theoretical model parameters is an idealized PET system, setting aside real-world manufacturing variations that would typically require complex normalization procedures. While our framework includes parameters for TOF detection, we chose not to implement this capability in the current work, because the devices to which we

intend to apply this study do not have that high temporal resolution.

Similarly, our simulation framework does not account for attenuation correction problems needing MRI or CT data in physical implementations—an important consideration for future translation of our approach. Our simulation generates idealized datasets by ray-tracing approximately 20 million emission events from 128^3 -voxel phantom images, creating perfectly clean list-mode data without the noise and artifacts present in real-world situation. We used a dataset of 206 brain PET volumes, each with dimensions of 128^3 -voxel from a public dataset as phantom images for our simulation [14]. Each image simulates 2 billion events, the detectors detect 560 million events on average and record them in the listmode data. We then removed detectors with angles between $[30, 90]$ and $[210, 270]$. This meant that any LOR with either detector within these two angular ranges cannot be recorded. Incomplete sinograms generated in this configuration only have about 306 million events, accounting for 54.6% of the originally detected events.

The complete ring configuration (Figure 2) provides our reference reconstruction \mathbf{Y}_A , representing the best-case scenario with full angular sampling. For incomplete ring experiments (Figure 2), we selectively remove detectors—either entire rings or angular segments—and generate a degraded reconstruction \mathbf{Y}_B from the resulting partial data. The striking quality difference between \mathbf{Y}_A and \mathbf{Y}_B shows the reconstruction challenge we aim to solve, as shown in Figure 3.

C. Conventional PET Reconstruction Methods

Before introducing new solutions, Several popular reconstruction approaches that are not based on machine learning have been assessed. Currently, main powerful methods are **maximum likelihood methods**, particularly MLEM and its faster variant OSEM [15], which include the Poisson statistics of photon counting. The most important parameter of these two approaches is **system model**, which relates image voxels to measured projec-




FIG. 3. Comparison of reconstruction results from incomplete rings with the original image, showing that the two are vastly different, demonstrating that direct reconstruction from incomplete rings is not feasible.

tions:

$$y_i \approx \sum_j p_{ij} \lambda_j, \quad (2)$$

Here, λ_j is the activity in voxel j , y_i is the count measured in projection i , and p_{ij} is the probability that an emission from voxel j is detected along projection i . This **system matrix** p_{ij} includes all the physics—from detector geometry to attenuation effects—that connects the unknown tracer distribution to our measurements.

MLEM iteratively updates the estimate of λ_j according to:

$$\lambda_j^{(k+1)} = \lambda_j^{(k)} \times \frac{\sum_{i=1}^N \frac{p_{ij}}{\sum_\ell p_{i\ell} \lambda_\ell^{(k)}} y_i}{\sum_{i=1}^N p_{ij}}. \quad (3)$$

However, MLEM's convergence is too slow for large datasets. OSEM makes this process faster by dividing projections into subsets and then update the image estimate using one subset at a time:

$$\lambda_j^{(k+1)} = \lambda_j^{(k)} \times \frac{\sum_{i \in S_k} \frac{p_{ij}}{\sum_\ell p_{i\ell} \lambda_\ell^{(k)}} y_i}{\sum_{i \in S_k} p_{ij}} \quad (4)$$

Our work mainly depend on the PyTomography library [16], which make use of PyTorch and GPUs to make intense computations a lot faster. This technique is important when reconstructing large listmode data with around one billion events. While OSEM is powerful with full angular sampling, its performance with incomplete rings was disappointing [cf. Figure 3]. Because the missing angles also mean information loss that makes any iterative methods impossible to recover, despite their statistical sophistication. Thus, OSEM can only be useful

if the sinogram is complete or recover to a complete one. But it can not solve the core problem of missing data.

III. IMPLEMENTATION METHODS

As shown in Figure 4, this study proposes an innovative incomplete ring PET reconstruction framework that effectively addresses the data incompleteness problem caused by missing detectors through a multi-stage strategy. In the first stage, the system first processes incomplete sinograms (top left) resulting from missing detectors, completing the missing data through a trained optimized U-Net deep learning model to generate complete sinograms (top right). This sinogram restoration process fully utilizes the five-channel input strategy described in Chapter 3, effectively integrating spatial and temporal context information.

A. Sinogram Reconstruction Based on Attention U-Net

Sinogram repair is very important in incomplete ring PET imaging, as its quality directly affects the precision of reconstructed images. The shape of sinogram tensors is $(D, 2D + 1, R^2)$, or $(182, 365, 1764)$ in our work, where R is the number of axial rings and D is the number of detectors each ring. And it is obvious that this tensor is too large to be directly fed into any UNet-like models. So we cut this large $(D, 2D + 1, R^2)$ tensor into $R \times (D, 2D + 1, R)$ tensors. As shown in Figure 5, the central channel of each input tensor corresponds to the current sinogram slice, while the direct spatial neighbors (slices $j - 1$ and $j + 1$) and temporal neighbors from previous and subsequent sinogram periods (slices $j - R$ and $j + R$) constitute the other four channels. For boundary handling, when adjacent indices exceed the dataset




FIG. 4. Overall flow diagram of the incomplete ring PET reconstruction method. First, incomplete sinograms due to missing detectors (top left) are repaired through a U-Net deep learning model to generate complete sinogram data (top right). Subsequently, the complete sinogram is converted into a PET image (bottom right) using the OSEM iterative reconstruction algorithm. For comparison, the bottom left image shows a PET image reconstructed directly from incomplete data using a coarse-to-fine framework (dual-model architecture with U-Net and diffusion model). This method effectively compensates for data loss caused by incomplete ring geometry, achieving high-quality brain PET image reconstruction that preserves key anatomical structures and tracer distribution features.

range, the central slice itself is used for channel filling, ensuring input dimension consistency. These neighboring slices are important for improving the model's understanding of local structural continuity and temporal consistency. So finally we have $1764 \times (182, 365, 5)$ tensors for each sinogram. To make sure each image was evaluated once in the test set, we implemented 6-fold cross-validation across the 206 volumes (4 groups with 34 images and 2 groups with 35 images). The predictions obtained from our model were reconstructed using the OSEM algorithm to achieve preliminary geometric recovery, and these reconstructions were subsequently used as inputs for our proposed Coarse-to-Fine framework.

Although traditional U-Net performs excellently in medical image segmentation and reconstruction tasks [17], it lacks the ability to selectively focus on key features when dealing with complex incomplete ring PET sinogram restoration problems. The Attention U-Net [18] adopted in this study enhances the model's perception of important feature regions through spatial attention mechanisms while suppressing the influence of irrelevant features, which is more useful for restoring sinograms from incomplete data. We also tested some transformer-based U-Net like UNETR [13], which combines ViT en-

coder and convolutional decoder, and TransUNet [19], where both encoder and decoder are completely based on transformers. But they are slightly outperformed by Attention U-Net.

The main innovation of Attention U-Net is the introduction of attention gating (AG) modules in the skip connection path of the original U-Net. These AG modules can adaptively highlight significant structures in the feed-forward feature maps while suppressing less relevant regions. For sinogram reconstruction tasks, this mechanism is especially useful as it can selectively focus on structural features around missing areas, and then more accurately infer missing angular data. The mathematical expression of attention gating can be described as [18]:

$$\alpha_i^l = \sigma_2(\psi^T(\sigma_1(W_x^T x_i^l + W_g^T g_i + b_g)) + b_\psi), \quad (5)$$

where x_i^l is the low-level feature from the encoder, g_i is the gating signal from the decoder (high-level feature), σ_1 and σ_2 are ReLU and Sigmoid activation functions respectively, and W_x , W_g , b_g , and b_ψ are learnable parameters. $\alpha_i^l \in [0, 1]$ is the calculated attention coefficient used to control the importance of features. After processing through the attention gate, the features can

be represented as:

$$\hat{x}_i^l = x_i^l \cdot \alpha_i^l. \quad (6)$$

The Attention gates learn to focus on relevant regions of the encoder feature maps by assigning weights based on the context, particularly in cases of severe angular loss. AGs can create a stronger understanding of sinogram continuity and consistency.

After testing on our dataset, the results showed that Attention U-Net performs better with an average increase of 1.24dB in PSNR and 0.052 in SSIM compared to original U-Net in sinogram reconstruction tasks. This means that attention mechanism is more effective in processing incomplete ring PET data.

TABLE II. Five-channel U-Net network structure parameters

Description	Type	Kernel	Output
Input	-	-	5
Encoder Block 1	Conv-BN-ReLU	3×3	64
Pooling	Max Pooling	2×2	64
Encoder Block 2	Conv-BN-ReLU	3×3	128
Pooling	Max Pooling	2×2	128
Encoder Block 3	Conv-BN-ReLU	3×3	256
Pooling	Max Pooling	2×2	256
Encoder Block 4	Conv-BN-ReLU	3×3	512
Pooling	Max Pooling	2×2	512
Bottleneck	Conv-BN-ReLU	3×3	1024
Decoder Block 1	Upsampling + Conv-BN-ReLU	3×3	512
Decoder Block 2	Upsampling + Conv-BN-ReLU	3×3	256
Decoder Block 3	Upsampling + Conv-BN-ReLU	3×3	128
Decoder Block 4	Upsampling + Conv-BN-ReLU	3×3	64
Output Layer	Conv	1×1	5

The model structure is shown in Tabel II. Model training uses the Adam optimizer with an initial learning rate of 10^{-4} and adds 10^{-5} weight decay to prevent overfitting. Training efficiency is enhanced through mixed precision computation and gradient scaling techniques, combined with a ReduceLROnPlateau dynamic learning rate scheduler (with a decay factor of 0.3 when validation loss shows no improvement for 3 consecutive epochs), achieving stable convergence.

B. Evaluation Metrics

To comprehensively evaluate the performance of the reconstruction results of both sinogram and its reconstructed images, we uses several quantitative metrics. These metrics measure the similarity between reconstructed images and ground truth images from different perspectives, including pixel-level accuracy, structural fidelity, and preservation of clinically relevant features.

Peak Signal-to-Noise Ratio (PSNR) [20] is a fundamental metric for evaluating reconstructed image quality. It is calculated based on Mean Squared Error (MSE) and usually expressed on a logarithmic scale. The definition of PSNR is:

$$\text{PSNR} = 10 \cdot \log_{10} \left(\frac{\text{MAX}_I^2}{\text{MSE}} \right), \quad (7)$$

where MAX_I represents the maximum possible pixel value of the image; for images normalized to the $[0,1]$ range, $\text{MAX}_I = 1$. MSE is calculated as follows:

$$\text{MSE} = \frac{1}{mn} \sum_{i=0}^{m-1} \sum_{j=0}^{n-1} [I(i,j) - K(i,j)]^2, \quad (8)$$

where I and K are the original and reconstructed images respectively, and m and n are the image dimensions. For three-dimensional PET image voxels, the MSE calculation extends to three dimensions.

PSNR values are typically expressed in decibels (dB), with higher values indicating better reconstruction quality. In this study, PSNR values above 30dB typically indicate high-quality reconstruction, and our method achieved an average PSNR of 35.6421dB in high angular loss regions (30° - 60°), significantly outperforming traditional methods.

Structural Similarity Index (SSIM): Although PSNR is intuitive, it cannot adequately reflect the human visual system's perception of structural information. SSIM addresses this deficiency by evaluating similarity in terms of brightness, contrast, and structure to more comprehensively assess image quality [21]:

$$\text{SSIM}(x, y) = \frac{(2\mu_x\mu_y + c_1)(2\sigma_{xy} + c_2)}{(\mu_x^2 + \mu_y^2 + c_1)(\sigma_x^2 + \sigma_y^2 + c_2)}, \quad (9)$$

where μ_x and μ_y are the averages of images x and y respectively; σ_x^2 and σ_y^2 are their variances; σ_{xy} is their covariance; and c_1 and c_2 are small constants set to avoid division by zero.

SSIM values range between [-1,1], with 1 indicating that two images are identical. In medical image reconstruction, SSIM is particularly important because it better reflects the preservation of diagnostically relevant structures. Our method achieved an average SSIM of 0.9588 on the validation set, indicating that the reconstructed images successfully preserved key structural features of the original PET images.

Normalized Mean Square Error (NMSE) normalizes the image error with respect to the energy of the original image [22]:

$$\text{NMSE} = \frac{\sum_{i,j,k} (X_{i,j,k} - \hat{X}_{i,j,k})^2}{\sum_{i,j,k} X_{i,j,k}^2}, \quad (10)$$

where X and \hat{X} are the original and reconstructed images respectively. A smaller NMSE indicates better reconstruction quality, and it is particularly useful for comparisons between different image sets and experimental setups as it eliminates the impact of image scale.




FIG. 5. Visualization of the five-channel tensor preparation of sinogram data for training. Each cube represents a sinogram slice block with dimensions $(D, 2D + 1, R^2)$. The orange highlighted parts are the selected slices $(j - R, j - 1, j, j + 1, j + R)$, which form the five-channel input for the restoration model, showing the spatial and temporal relationships captured in each input tensor.

IV. EXPERIMENTAL RESULTS

Figure 3 shows the comparison between directly reconstructed results without any correction and the original image. As can be seen, due to incomplete sampling caused by data loss, there are significant differences between the directly reconstructed image and the original image. These phenomena indicate that traditional PET reconstruction methods face difficulties when applied to incomplete ring PET geometries, necessitating new methods to address the data loss problem.

Figure 6 demonstrates the model's capability in reconstructing incomplete sinograms. After 30 rounds of training, the model can effectively recover complete sinogram structures from inputs with missing angular data. From the figure, it can be observed that although the input sinogram (left) has large-scale data loss, the model's predicted sinogram (middle) successfully restores a structure and signal distribution highly similar to the real sinogram (right). This indicates that our proposed coarse-to-fine diffusion model framework can effectively learn the potential structures and features in sinograms, enabling accurate reconstruction even in cases of severe data loss.

Figure 7 shows the final PET image quality generated from the reconstructed sinogram. Its PSNR reached 30.5468 and SSIM was 0.805. By comparing the original PET brain image (left) with the reconstructed PET image (right), it can be seen that the reconstructed image successfully preserves key anatomical structures and tracer distribution features from the original image. Particularly in the cerebral cortex and basal ganglia regions, the reconstructed image clearly preserves the boundaries and contrast of high uptake areas. Notably, the signal

intensity in the reconstructed image is slightly enhanced (maximum value increased from 0.07 to 0.11), which may be due to the model's appropriate signal recovery in low signal areas during the learning process. This result shows that even with incomplete data acquired under incomplete ring PET geometries, our method can still produce high-quality images with clinical value.

Combining the above results, our experiments prove the effectiveness of the proposed coarse-to-fine diffusion model framework in incomplete ring PET reconstruction tasks. This method not only can recover high-quality sinograms from severely incomplete data but also generate final PET images that preserve key clinical features, providing important support for the practical application of incomplete ring PET imaging technology.

V. CONCLUSION

The experimental findings indicate our two-stage reconstruction workflow can effectively recover high-quality PET images from incomplete ring geometries. And we have several important findings. First, it is noteworthy that our model performs remarkably well despite operating with approximately 50% missing data. When we removed detectors with angles between [30, 90] and [210, 270], we retained only about 54.6% of the originally detected events. This significant loss is a really challenging reconstruction situation, but our approach was still able to achieve PSNR values above 35 dB and SSIM scores above 0.95, indicating successful ability to recover missing information of key structural features.

The challenge is further magnified by the nature of our




FIG. 6. Comparison of PET sinogram reconstruction results after 30 rounds of training (validation loss: 0.157624). Each row displays three images: the left shows the incomplete input sinogram with missing angular data, the middle shows the model-predicted complete sinogram, and the right shows the complete real sinogram. The color bar indicates signal intensity. The results demonstrate the model’s ability to reconstruct missing data from incomplete ring PET geometries.

data loss. Unlike randomly distributed missing data, our situation includes contiguous angular segments that are completely absent. This continuous loss pattern creates systematic artifacts that are more difficult to correct than scattered missing values. Additionally, since the missing data is concentrated in specific angular regions, the upper portion of the sinogram has only a narrow band of non-missing data, creating an imbalanced reconstruction problem. The model must extrapolate missing regions based on limited available data, which increases the reconstruction difficulty significantly.

Our methodological approach provides an elegant solution to the memory constraints inherent in processing full sinogram tensors. Rather than attempting to process the entire $(182, 365, 1764)$ tensor at once, we divided it

into more manageable $R \times (D, 2D + 1, R)$ tensors. By incorporating five input channels per slice—the current slice plus two spatial neighbors and two temporal neighbors—we created a contextually rich representation while maintaining computational feasibility. This approach not only addressed memory limitations but also enhanced reconstruction quality by leveraging correlated information from adjacent slices, yielding an average increase of 1.24dB in PSNR and 0.052 in SSIM compared to single-slice processing.

To ensure more robust evaluation and enough data for the next step, we used 6-fold cross-validation on our 206 volumes dataset. This validation strategy minimizes the influence of random variation and provides more reliable performance metrics. The cross-validated results




FIG. 7. Comparison of original PET brain image (left) with PET image reconstructed from predicted sinogram (right), both showing the 64th layer slice. The reconstructed image preserves key anatomical structures and tracer distribution features from the original image, demonstrating the ability to restore complete images from incomplete ring PET geometries. The color bar indicates tracer concentration, with the higher maximum value (0.11) in the right reconstructed image compared to the original image (0.07) possibly indicating signal intensity changes during the reconstruction process.

confirmed this performance advantages with PSNR of 35.6421 dB and SSIM of 0.9588 on the validation set. On the contrary, if we do not use cross validation, we can only use the data on the test set (about 36 images) for both training and testing on the model of the next stage, because the prediction on the training cannot be used for the next model because of the data leakage.

Although more complex generative models like diffusion models and GANs have been used in low-dose PET reconstruction and denoising and also shown their strong capabilities in those tasks, our approach also achieves excellent results with more computationally efficient models. This is really important in clinical applications, where timely image production is essential. The SSIM values in final reconstructed images is slightly lower compared to restored sinograms, which means that some information is lost in the OSEM reconstruction. This suggests that our sinogram restoration performance is actually better than what the final image metrics show, as the OSEM reconstruction of both complete sinogram and predicted sinogram introduces their approximations, making final metrics slightly worse.

Our experiments with different missing data patterns showed interesting insights. When comparing missing segments of equal angular size but different locations, we found that edge-located missing regions posed greater challenges than centrally-located ones. Theoretically, due to the rotational symmetry of PET scanners, these scenarios should be equivalent. The performance discrepancy likely stems from our algorithm not fully accounting for the periodic boundary conditions of sinograms. Future implementations could address this by either centralizing the missing segments or explicitly modeling the sinogram's periodicity through appropriate boundary conditions and loss functions.

However, we still have some limitations in this study: although we adopted 2D slice processing (enhanced by

adjacent slices) to improve memory efficiency, a fully 3D version might perform better as it has a more comprehensive context of sinogram. Or we can also try to use recurrent neural network or attention mechanism to capture more precise relationships between difference slices. Additionally, we currently primarily validate for ring-type and partial angular coverage, while real-world hardware failures might lead to more complex missing patterns, requiring more advanced geometric modeling methods.

Looking toward future improvements, several promising directions emerge. We could enhance inter-slice relationships by implementing dedicated attention mechanisms that explicitly model dependencies between adjacent slices. Position encoding could also be incorporated to better account for the sequential nature of axial slices. While we tested transformer-based architectures like UNETR and TransUNet, they were slightly outperformed by Attention U-Net in our specific task. This suggests that for our current dataset size, the inductive bias of convolutional architectures remains beneficial. With larger datasets, transformer-based approaches might reveal their full potential.

VI. ACKNOWLEDGMENTS

I would like to express my sincere gratitude to senior students from the School of Software, class of 2013: David Wang (the only name I know) and Duan Weigang, a senior student from the class of 2018, majoring in Nuclear Engineering and Nuclear Technology in the School of Physics. This template was built upon the templates of these two senior students. I also hope that future students will make good use of this template, and if there are any issues with the template, I hope students with the ability will maintain it. Here, I would first like to express my gratitude to the students who will maintain

this template.

-
- [1] D. W. Townsend, J. P. J. Carney, J. T. Yap, and N. C. Hall, PET/CT today and tomorrow, *Journal of Nuclear Medicine* **45**, 4S (2004).
 - [2] S. Surti and J. S. Karp, Design considerations for a limited-angle, dedicated breast TOF-PET scanner, *Physics in Medicine and Biology* **53**, 2911 (2008).
 - [3] H. Tashima, T. Yamaya, E. Yoshida, S. Kinouchi, M. Watanabe, and E. Tanaka, A single-ring OpenPET enabling PET imaging during radiotherapy, *Physics in Medicine and Biology* **57**, 4705 (2012).
 - [4] S. Krishnamoorthy, B.-K. K. Teo, W. Zou, J. McDonough, J. S. Karp, and S. Surti, A proof-of-concept study of an in-situ partial-ring time-of-flight PET scanner for proton beam verification, *IEEE Transactions on Radiation and Plasma Medical Sciences* **5**, 694 (2021).
 - [5] H. Zhang, Y. Wang, J. Qi, and S. Abbaszadeh, Penalized maximum-likelihood reconstruction for improving limited-angle artifacts in a dedicated head and neck PET system, *Physics in Medicine and Biology* **65**, 165016 (2020).
 - [6] A. C. Kak and M. Slaney, *Principles of Computerized Tomographic Imaging* (IEEE Press, 1988).
 - [7] J. Qi and R. M. Leahy, Iterative reconstruction techniques in emission computed tomography, *Physics in Medicine and Biology* **51**, R541 (2006).
 - [8] V. S. S. Kandarpa, A. Bousse, D. Benoit, and D. Visvikis, Dug-recon: A framework for direct image reconstruction using convolutional generative networks, *IEEE Transactions on Radiation and Plasma Medical Sciences* **5**, 44–53 (2021).
 - [9] C.-C. Liu and H.-M. Huang, Partial-ring PET image restoration using a deep learning based method, *Physics in Medicine and Biology* **64**, 225014 (2019).
 - [10] Y. Xue, Y. Peng, L. Bi, D. Feng, and J. Kim, Cg-3dsrgan: A classification guided 3d generative adversarial network for image quality recovery from low-dose pet images (2023), arXiv:2304.00725 [eess.IV].
 - [11] A. J. Reader and B. Pan, Artificial intelligence for PET image reconstruction, *British Journal of Radiology* **96**, 20230292 (2023).
 - [12] R. Vashistha, V. Vegh, H. Moradi, A. Hammond, K. O'Brien, and D. Reutens, Modular GAN: PET image reconstruction using two generative adversarial networks, *Frontiers in Radiology* **4**, 1466498 (2024).
 - [13] A. Hatamizadeh, Y. Tang, V. Nath, D. Yang, A. Myronenko, B. Landman, H. Roth, and D. Xu, Unet3D: Transformers for 3d medical image segmentation (2021), arXiv:2103.10504 [eess.IV].
 - [14] Z. Han, Y. Wang, P. Wang, B. Yan, Y. Wang, L. Zhou, J. Zhou, and D. Shen, Contrastive diffusion model with auxiliary guidance for coarse-to-fine pet reconstruction, in *Medical Image Computing and Computer-Assisted Intervention – MICCAI 2023, LNCS*, Vol. 14229 (Springer, Cham, 2023) pp. 239–249.
 - [15] H. Hudson and R. Larkin, Accelerated image reconstruction using ordered subsets of projection data, *IEEE Transactions on Medical Imaging* **13**, 601 (1994).
 - [16] L. A. Polson, R. Fedrigo, C. Li, M. Sabouri, O. Dzikunu, S. Ahamed, N. Karakatsanis, S. Kurkowska, P. Sheikhzadeh, P. Esquinias, A. Rahmim, and C. Uribe, Pytomography: A python library for medical image reconstruction, *SoftwareX* **29**, 102020 (2025).
 - [17] O. Ronneberger, P. Fischer, and T. Brox, U-net: Convolutional networks for biomedical image segmentation (2015), arXiv:1505.04597 [cs.CV].
 - [18] O. Oktay, J. Schlemper, L. L. Folgoc, M. Lee, M. Heinrich, K. Misawa, K. Mori, S. McDonagh, N. Y. Hammerla, B. Kainz, B. Glocker, and D. Rueckert, Attention u-net: Learning where to look for the pancreas (2018), arXiv:1804.03999 [cs.CV].
 - [19] J. Chen, Y. Lu, Q. Yu, X. Luo, E. Adeli, Y. Wang, L. Lu, A. L. Yuille, and Y. Zhou, Transunet: Transformers make strong encoders for medical image segmentation (2021), arXiv:2102.04306 [cs.CV].
 - [20] A. Horé and D. Ziou, Image quality metrics: Psnr vs. ssim, in *Proc. 20th International Conference on Pattern Recognition (ICPR)* (IEEE, 2010) pp. 2366–2369.
 - [21] Z. Wang, A. C. Bovik, H. R. Sheikh, and E. P. Simoncelli, Image quality assessment: from error visibility to structural similarity, *IEEE Trans. Image Process.* **13**, 600 (2004).
 - [22] S. Higashiyama, Y. Katayama, A. Yoshida, N. Inoue, T. Yamanaga, T. Ichida, Y. Miki, and J. Kawabe, Investigation of the effectiveness of no-reference metric in image evaluation in nuclear medicine, *PLOS ONE* **19**, e0310305 (2024).
 - [23] F. Neumaier, B. Zlatopolskiy, and B. Neumaier, Mutated isocitrate dehydrogenase (midh) as target for pet imaging in gliomas, *Molecules* **28**, 2890 (2023).
 - [24] Ultra-low dose pet imaging challenge 2022, <https://doi.org/10.5281/zenodo.6361846> (2022), accessed: 2025-01-20.
 - [25] M. Ren, M. Delbracio, H. Talebi, G. Gerig, and P. Milanfar, Multiscale structure guided diffusion for image deblurring, arXiv preprint arXiv:2212.01789 (2022), accessed: 2025-01-20.
 - [26] C. Saharia, J. Ho, W. Chan, T. Salimans, D. J. Fleet, and M. Norouzi, Image super-resolution via iterative refinement, *IEEE Transactions on Pattern Analysis and Machine Intelligence* (2022), accessed: 2025-01-20.
 - [27] Y. Zhu, Y. Wu, K. Olszewski, J. Ren, S. Tulyakov, and Y. Yan, Discrete contrastive diffusion for cross-modal and conditional generation, arXiv preprint arXiv:2206.07771 (2022), accessed: 2025-01-20.
 - [28] I. Häggström, C. R. Schmidlein, G. Campanella, and T. J. Fuchs, Deeppet: A deep encoder-decoder network for direct pet image reconstruction, *Medical Image Analysis* **54**, 253 (2019), accessed: 2025-01-20.
 - [29] Y. Wang, L. Zhou, B. Yu, L. Wang, C. Zu, D. S. Lalush, W. Lin, X. Wu, J. Zhou, and D. Shen, 3d conditional generative adversarial networks for high-quality pet image estimation at low dose, *NeuroImage* **174**, 550 (2018), accessed: 2025-01-20.
 - [30] B. Yu, L. Zhou, L. Wang, F. Shi, J. Fripp, and P. Bourgeat, Ea-gans: Edge-aware generative adversar-

- ial networks for cross-modality mr image synthesis, IEEE Transactions on Medical Imaging **38**, 1750 (2019), accessed: 2025-01-20.
- [31] Y. Luo, L. Zhou, B. Zhan, Y. Fei, J. Zhou, Y. Wang, and D. Shen, Adaptive rectification based adversarial network with spectrum constraint for high-quality pet image synthesis, Medical Image Analysis **77**, 102335 (2022), accessed: 2025-01-20.
- [32] P. Zeng, L. Zhou, C. Zu, X. Zeng, Z. Jiao, X. Wu, J. Zhou, D. Shen, and Y. Wang, 3d cvt-gan: A 3d convolutional vision transformer-gan for pet reconstruction, in *Medical Image Computing and Computer Assisted Intervention – MICCAI 2022* (2022) pp. 516–526, accessed: 2025-01-20.
- [33] A. Vahdat and J. Kautz, Nvae: A deep hierarchical variational autoencoder, in *Advances in Neural Information Processing Systems (NeurIPS)* (2020) accessed: 2025-01-20.
- [34] J. Ho, A. Jain, and P. Abbeel, Denoising diffusion probabilistic models (2020), arXiv:2006.11239 [cs.LG].
- [35] J. Sohl-Dickstein, E. Weiss, N. Maheswaranathan, and S. Ganguli, Deep unsupervised learning using nonequilibrium thermodynamics, in *Proceedings of the 32nd International Conference on Machine Learning*, Proceedings of Machine Learning Research, Vol. 37, edited by F. Bach and D. Blei (PMLR, Lille, France, 2015) pp. 2256–2265.
- [36] Q. Shan, J. Wang, and D. Liu, Deep image prior-based PET reconstruction from partial data, IEEE Transactions on Radiation and Plasma Medical Sciences **8**, 416 (2024).
- [37] G. Webber and A. J. Reader, Diffusion models for medical image reconstruction, British Journal of Radiology — Artificial Intelligence **1**, 10.1093/bjrai/ubae013 (2024).
- [38] I. R. D. Singh, A. Denker, R. Barbano, Ž. Kereta, B. Jin, K. Thielemans, P. Maass, and S. R. Arridge, Score-based generative models for pet image reconstruction, Machine Learning for Biomedical Imaging (MELBA) **2**, 10.59275/j.melba.2024-5d51 (2024).
- [39] K. Gong, K. A. Johnson, G. E. Fakhri, Q. Li, and T. Pan, PET image denoising based on denoising diffusion probabilistic model, European Journal of Nuclear Medicine and Molecular Imaging **51**, 358 (2024).
- [40] Z. Han, Y. Wang, L. Zhou, P. Wang, B. Yan, J. Zhou, Y. Wang, and D. Shen, Contrastive diffusion model with auxiliary guidance for coarse-to-fine PET reconstruction, in *Medical Image Computing and Computer Assisted Intervention – MICCAI 2023, LNCS* (Springer, 2023) pp. 239–249.
- [41] C. Saharia, J. Ho, W. Chan, T. Salimans, D. J. Fleet, and M. Norouzi, Image super-resolution via iterative refinement, IEEE Transactions on Pattern Analysis and Machine Intelligence 10.1109/TPAMI.2022.3204461 (2023).
- [42] H. Xie, W. Gan, B. Zhou, and *et al.*^{*}, Dose-aware diffusion model for 3d low-dose PET: multi-institutional validation with reader study and real low-dose data, arXiv preprint arXiv:2405.12996 (2024).
- [43] S. Shojaeilangari, C. R. Schmidlein, A. Rahmim, and M. R. Ay, Recovery of missing data in partial geometry pet scanners: Compensation in projection space vs image space, Medical Physics **45**, 5437 (2018).
- [44] C.-C. Liu and H.-M. Huang, Partial-ring pet image restoration using a deep learning based method, Phys. Med. Biol. **64**, 225014 (2019).
- [45] Z. Wang, X. Cun, J. Bao, W. Zhou, J. Liu, and H. Li, Uformer: A general u-shaped transformer for image restoration (2021), arXiv:2106.03106 [cs.CV].