

Quarterly Notes

The device was calibrated at 25 degrees Celsius before every run.
Each measurement was repeated three times and averaged.
Results are reported with a tolerance of 0.5 percent.
The operator recorded the ambient humidity for each session.

Method

Samples were held at 25 C for 10 min and then cooled.
The cooling rate was fixed at 2 C per minute.
Each condition was repeated on two separate days.

Discussion

The second page continues the discussion of the method.

No further calibration was required.

The observed spread is consistent with the stated tolerance.