



Self-supervised cardiac ultrasound segmentation

A paper talk on label-free echocardiography chamber masks

Research Presentation Program

Department of Statistics & Data Science

2026-08-29



香港中文大學
The Chinese University of Hong Kong



The model object is a weak-label SSL segmentation pipeline

The source pipeline is manual-label-free for training. human segmentations in later figures are visual aids or evaluation references.



The image shows a small, rotated snippet of a table from a document. The table has multiple columns with headers and rows of data. The text is too small and rotated to be legible, but it appears to be a technical or scientific table.

Figure 1 page, Ferreira et al. 2025, CC BY.

3/7



A2C bootstraps the later chamber pipeline

Article


<https://doi.org/10.1038/s41467-025-6040-9>

Fig. 2 | Examples of segmentation and measurement improvement through successive steps of the A2C pipeline for the apical 2-chamber view. Visualization of segmentation for several examples for A2C (A–D). A shows a typical good quality A2C image, the initial weak label (red) and the pipeline's improvement (blue). B shows the pipeline's improvement (blue) and the pipeline's improvement (blue). C shows the pipeline's improvement (blue) and the pipeline's improvement (blue). D shows the pipeline's improvement (blue) and the pipeline's improvement (blue).

To demonstrate incremental impact of each step of the pipeline on predictive performance and generalizability, segmentations from intermediate steps of the pipeline were compared to clinical area measurements on images from the validation dataset (Fig. 3). The r^2 on chamber areas ranged from 0.06 to 0.22 when using initial weak labels, compared to 0.33–0.81 using the full pipeline. Bias and LOA similarly improved with successive training steps.

Pipeline-derived structure and function measurements are comparable to clinical echocardiogram measurements

Pipeline-derived measurements in the all-comers dataset were compared to clinical echocardiogram measurements (Fig. 6, Supplementary Table S1, Supplementary Fig. S3, S4).

Left ventricle. Pearson correlations (r) between the AI pipeline and clinical echocardiogram measurements for LV end-diastolic volume (LV EDV), LV end-systolic volume (LV ESV), and LV ejection fraction (LVEF) were 0.84, 0.93, and 0.81, respectively (Supplementary Table S2). r^2 for these measurements were 0.70, 0.86, and 0.65, respectively. For LVEF achieved through the self-supervised AI pipeline is better than those reported from supervised learning¹⁰ (Figs. 6A and S3, Supplementary Table S2). Bland-Altman bias $\pm$ LOA (two standard deviations) for LV EDV, LV ESV, and LVEF were 2.6 $\pm$ 13.6 mL, 5.5 $\pm$ 25.4 mL, and $-5.3 \pm$ 14.4%, respectively, similar to clinician variability studies^{11–13} and consistent with those achieved from supervised learning¹⁰ (Figs. 6A, S3, Supplementary Table S1, Supplementary Fig. S3). Correlations for LV mass were lower ($r = 0.52$, $r^2 = 0.27$), but are still useful to reported benchmarks (Fig. 6B).

Right ventricle. r^2 were 0.69 and 0.71 for RV EDV and RVESV, respectively, indicating a strong correlation¹⁴ between the AI pipeline and

and on images with a large LA mass (blue arrowhead) progresses into the LV (D). For each of A–D, complete normal segmentations are shown as a visual aid for the reader. Off-diagonal self-supervised segmentations, note that no manual segmentations were used in training the pipeline. Note two final clinical measurements were not made from a single image frame but rather of image frames in a given cardiac cycle, and often multiple views at once. A2C, apical 2-chamber; LV, left ventricle.

clinical measurements. r 's were 0.63 and 0.84 and bias $\pm$ LOA was -0.16 ± 5.4 mL², 1.6 ± 3.9 mL² (Supplementary Table S1, Figs. 6F, S5, Supplementary Fig. S3).

Aorta. The r^2 for left aortic volume was 0.84, showing a very strong correlation. Bias $\pm$ LOA was -0.13 ± 20 mL, consistent with clinical inter-observer bias and with supervised learning performance reported in the literature¹⁵ (Fig. 6C). Right aortic volume r^2 was 0.76 and bias $\pm$ LOA between AI and clinical measurements was -2.1 ± 21 mL (Fig. 6D).

Normal vs. abnormal. For further clinical context, the above measurements were indexed to body surface area where applicable and binned as normal or abnormal according to reference guidelines^{16–18}. Accuracy and Cohen's kappa values are presented in Fig. 6, Supplementary Table S1 and Supplementary Fig. S4. Accuracy ranged from 0.71 for LV mass to 0.97 for LVEF. Kappa values ranged from 0.56 to 0.79 showing a moderate to substantial agreement¹⁹ between AI pipeline and clinical echocardiogram measurements. LV mass and RVESV were the exception, where kappa values were only fair.

Taken together, measurements derived from self-supervised learning performed similarly to clinical inter- and intra-observer variability and to measurements derived from supervised deep learning.

Self-supervised segmentation and clinical echo correlate similarly to CMR gold standard

Compared to all-comers, the CMR subset is younger, more male, and has more LV dysfunction (Table 1). Also, due to the nature of different clinical imaging protocols, CMR studies did not include measurements for all chambers. Furthermore, methods for measurement differ

Nature Communications | (2025)16:4270

5

Figure A2C pipeline improves left and right chamber labels and the progression from weak labels to final predictions on real ultrasound frames.

The claim is scalable measurement, not label magic

Failure evidence

Noisy ultrasound boundaries and clinical frame selection can break a simple segmentation-quality story.




Observed limit

motivates

Manipulate sampling

- preserve view labels
- check chamber anatomy
- compare measurements

Comparator arms



Decision rule


go/no-go

Accept the pipeline only if real echo pixels and clinical measurement semantics remain tied to the article.

The paper-specific conclusion is that label-free training can scale segmentation, while validation remains clinical and anatomy-specific.



A2C, A4C, and SAX must not be conflated



All panels are source article pixels; the semantic risk is mixing view labels or treating clinical measurements as single-frame manual labels.