



# DECLARATION OF CONFORMITY

ACCORDING TO (EU) 2017/745 MEDICAL DEVICE REGULATION

## EU Representative

**SUNGO Europe B.V.**  
**Fascinatio Boulevard 522, Unit 1.7,**  
**2909VA Capelle aan den IJssel, The**  
**Netherlands**  
**SRN: NL-AR-000000247**

## Conformity Assessment

**Conformity Assessment Procedure**  
Annex II+III of Regulation (EU) 2017/745

### Applicable Standards

EN ISO14971:2019  
EN ISO15223-1:2021  
EN ISO10993-1:2020  
EN ISO10993-5:2009  
EN ISO10993-10:2010  
EN ISO 20417: 2021

### Remark

*The declaration of conformity is valid in connection with the release technical document CE/MDR-S0102-01.*

*All the supporting documentation is retained at the premises of the manufacturer.*

*The Declaration of Conformity is exclusively under the sole responsibility of the manufacturer.*

## Manufacturer

**Name: ANQING MEDCOLOR IMP. & EXP. CO.,LTD**  
**Address: Room 212,Building No.2,Science and**  
**Technology Park,Economic Development**  
**District,246000,Anqing City,Anhui Province,China**  
**SRN: CN-MF-000032447**

## Product Information

**Name : STERILIZATION POUCH/REEL**

**Model : MCSP/MCHP/MCHR/MCHG**

**EMDN : S0102**

**Basic UDI : 697599556SPR013N**

**Classification:** Class I, According to Rule 1, Annex VIII, Regulation (EU) 2017/745

## Declaration

We herewith declare that the above-mentioned products meet the requirements of Medical Device Regulation (EU) 2017/745 and the applicable standards above.

Signature:

Date: 2023. 2. 2

Position: GM

Place: Anqing/China

