

EQUIPMENT IDENTITY AND CUSTOMER DATA

Received Date : Calibration Date :

EQUIPMENT

1. Name :
2. Capacity : (....)
3. Resolution : (....)
4. Nilai *e* dan kelas : *e* = (.....) Kelas = (.....)
5. Type/Model :
6. Serial Number :
7. Merk/Manufacturing :

STANDARD

STANDARD

Weight Set E2 100 mf- 200g ☐
Weight Set F1 100 mg-500g ☐ 10 kg ☐
Weight F2 1kg ☐ 2 kg ☐ 5 kg ☐
Weight M2 20kg ☐
Calibration Metode : SIDIK-IK-CAL-0505
Calibration Location :

ENVIRONMENT COND.

First

End

Units

°C

%RH

THERMOHYGRO:

☐ TH-2

☐ TH-6

☐ TH-7

OWNER

1. Name :
2. Address :

CALIBRATION RESULT

1. SCALL OBSERVATION

Position	Standard Weight ()	The Reading ()
1	2	3
		$z_1 =$
Before		$m_1 =$
Adjustment		$m_2 =$
		$z_2 =$
		$z'_1 =$
After		$m'_1 =$
Adjustment		$m'_2 =$
		$z'_2 =$
Standar deviasi yang lalu (SD)		= ()

2. EFFECT OF TARE

Standard Weight ()	The Reading ()
	$m_1 =$
	$m_2 =$

Bentuk Pan : ...
Diameter Pan : ... ()

M = rata-rata ($m_1:m_2$)
z= Rata-rata ($z_1:z_2$)
C= $M_s - (M - z)$
Jika $C > 5 \times \text{resolusi}$, maka timbangan harus dilakukan adjustment
(Ref: Tabel Kinerja LOP Timbangan)

3. ACCURACY

Nominal ()	The Reading ()
Mref=	$z_1 =$
	$m_1 =$ $m_1' =$ $z_1' =$
Mref=	$z_2 =$
	$m_2 =$ $m_2' =$ $z_2' =$
Mref=	$z_3 =$
	$m_3 =$ $m_3' =$ $z_3' =$
Mref=	$z_4 =$
	$m_4 =$ $m_4' =$ $z_4' =$
Mref=	$z_5 =$
	$m_5 =$ $m_5' =$ $z_5' =$
Mref=	$z_6 =$
	$m_6 =$ $m_6' =$ $z_6' =$
Mref=	$z_7 =$
	$m_7 =$ $m_7' =$ $z_7' =$
Mref=	$z_8 =$
	$m_8 =$ $m_8' =$ $z_8' =$
Mref=	$z_9 =$
	$m_9 =$ $m_9' =$ $z_9' =$
Mref=	$z_{10} =$
	$m_{10} =$ $m_{10}' =$ $z_{10}' =$

4. REPEATABILITY

No.	Middle Capacity		Miximum Capacity	
	... (...)		... (...)	
	Zero (zi)	Reading (mi)	Zero (zi)	Reading (mi)
	1	2	1	2
1				
2				
3				
4				
5				
6				
7				
8				
9				
10				

5. LOADING INFLUENCE ON EACH POSITION

..... ()
Untuk keamanan loadcell timbangan, Gunakan beban uji 1/3 ~ 1/2 dari kapasitas maksimal alat, atau dapat disesuaikan dengan spek. manufaktur

Position	Center	Front	Back	Left	Right
Reading					

6. HYSTERISIS

Standar Weight ()	Reading 1 ()	Reading 2 ()
M =	$M_{(p1)} =$	$M_{(p3)} =$
M' =	$M+M' =$	$M+M' =$
	$M_{(q1)} =$	$M_{(q3)} =$
	Zero =	Zero =
	$M+M' =$	$M+M' =$
	$M_{(q2)} =$	$M_{(q4)} =$
	Zero =	Zero =
	$M_{(p2)} =$	$M_{(p4)} =$

7. DRIFT

Drift Massa Standar (d) = (...)

Catatan:

Calibrated By

Checked By

Name :
Sign :

Name :
Sign :

Page 1 of 1

Revise : 4

SIDIK-FM-CAL-0508.A